

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF092721\  
 Data File : BF125678.D  
 Acq On : 27 Sep 2021 18:30  
 Operator : JU/CG  
 Sample : M3938-07  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SUB-SLAB-08(20-22)

Quant Time: Sep 27 19:08:15 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF092421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 24 15:38:32 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.887  | 152  | 244743   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.169  | 136  | 1017080  | 20.000  | ng    | # 0.00   |
| 39) Acenaphthene-d10        | 9.922  | 164  | 535963   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.410 | 188  | 886330   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 14.039 | 240  | 471090   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.515 | 264  | 435041   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.510  | 112  | 1565312  | 105.374 | ng    | 0.00     |
| 7) Phenol-d6                | 6.510  | 99   | 1903642  | 95.320  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.445  | 82   | 1127017  | 77.290  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.710 | 330  | 516938   | 103.446 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.245  | 172  | 2101901  | 67.752  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.992 | 244  | 1865567  | 68.580  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF092721\  
Data File : BF125678.D  
Acq On : 27 Sep 2021 18:30  
Operator : JU/CG  
Sample : M3938-07  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SUB-SLAB-08(20-22)

Quant Time: Sep 27 19:08:15 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF092421.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 24 15:38:32 2021  
Response via : Initial Calibration

