

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092822\
 Data File : BF130447.D
 Acq On : 28 Sep 2022 14:30
 Operator : CG\JU
 Sample : N4835-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 529317-4

Quant Time: Sep 29 02:14:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.645	152	111337	20.000	ng	0.00
21) Naphthalene-d8	7.928	136	442705	20.000	ng	# 0.00
39) Acenaphthene-d10	9.675	164	231054	20.000	ng	0.00
64) Phenanthrene-d10	11.157	188	357355	20.000	ng	0.00
76) Chrysene-d12	13.786	240	191666	20.000	ng	0.00
86) Perylene-d12	15.174	264	210256	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	315899	49.212	ng	0.00
7) Phenol-d6	6.281	99	225396	28.063	ng	-0.01
23) Nitrobenzene-d5	7.210	82	703902	81.231	ng	0.00
42) 2,4,6-Tribromophenol	10.463	330	177590	80.215	ng	0.00
45) 2-Fluorobiphenyl	9.004	172	1340531	84.485	ng	0.00
79) Terphenyl-d14	12.745	244	1149788	99.371	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092822\
Data File : BF130447.D
Acq On : 28 Sep 2022 14:30
Operator : CG\JU
Sample : N4835-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
529317-4

Quant Time: Sep 29 02:14:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

