

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092822\
 Data File : BF130459.D
 Acq On : 28 Sep 2022 21:33
 Operator : CG\JU
 Sample : N4826-01 5X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-03-092622

Quant Time: Sep 29 02:18:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/29/2022
 Supervised By :Jagrut Upadhyay 09/29/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.646	152	86190	20.000	ng	0.00
21) Naphthalene-d8	7.928	136	292251	20.000	ng	0.00
39) Acenaphthene-d10	9.681	164	141453	20.000	ng	0.00
64) Phenanthrene-d10	11.157	188	258349	20.000	ng	0.00
76) Chrysene-d12	13.792	240	189709	20.000	ng	0.00
86) Perylene-d12	15.186	264	141235	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	96594	19.438	ng	0.00
7) Phenol-d6	6.287	99	111917	18.000	ng	0.00
23) Nitrobenzene-d5	7.210	82	60919	10.649	ng	0.00
42) 2,4,6-Tribromophenol	10.463	330	29190	21.536	ng	0.00
45) 2-Fluorobiphenyl	9.004	172	108345	11.154	ng	0.00
79) Terphenyl-d14	12.745	244	132315	11.553	ng	0.00
Target Compounds						Qvalue
75) Fluoranthene	12.369	202	42243	3.014	ng	98
78) Pyrene	12.592	202	50043	3.015	ng	98
88) Benzo(b)fluoranthene	14.792	252	21092m	2.288	ng	
90) Benzo(a)pyrene	15.127	252	14819	2.029	ng	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092822\
Data File : BF130459.D
Acq On : 28 Sep 2022 21:33
Operator : CG\JU
Sample : N4826-01 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-03-092622

Quant Time: Sep 29 02:18:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/29/2022
Supervised By : Jagrut Upadhyay 09/29/2022

