

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF093021\  
 Data File : BF125735.D  
 Acq On : 30 Sep 2021 16:05  
 Operator : JU/CG  
 Sample : M3971-02  
 Misc :  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 RO-COMP-1

Quant Time: Sep 30 17:16:35 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF092421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 24 15:38:32 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min)  |
|-----------------------------|--------|------|----------|---------|-------|-----------|
| -----                       |        |      |          |         |       |           |
| Internal Standards          |        |      |          |         |       |           |
| 1) 1,4-Dichlorobenzene-d4   | 6.887  | 152  | 203373   | 20.000  | ng    | 0.00      |
| 21) Naphthalene-d8          | 8.169  | 136  | 820597   | 20.000  | ng    | # 0.00    |
| 39) Acenaphthene-d10        | 9.922  | 164  | 438052   | 20.000  | ng    | 0.00      |
| 64) Phenanthrene-d10        | 11.404 | 188  | 779325   | 20.000  | ng    | 0.00      |
| 76) Chrysene-d12            | 14.039 | 240  | 571378   | 20.000  | ng    | 0.00      |
| 86) Perylene-d12            | 15.515 | 264  | 441343   | 20.000  | ng    | 0.00      |
| System Monitoring Compounds |        |      |          |         |       |           |
| 5) 2-Fluorophenol           | 5.516  | 112  | 1407495  | 114.024 | ng    | 0.01      |
| 7) Phenol-d6                | 6.510  | 99   | 1521956  | 91.710  | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 7.445  | 82   | 1109375  | 94.296  | ng    | 0.00      |
| 42) 2,4,6-Tribromophenol    | 10.710 | 330  | 553888   | 135.615 | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 9.245  | 172  | 2141922  | 88.392  | ng    | 0.00      |
| 79) Terphenyl-d14           | 12.992 | 244  | 2345648  | 71.093  | ng    | 0.00      |
| Target Compounds            |        |      |          |         |       |           |
| 77) Benzidine               | 12.733 | 184  | 431564   | 28.751  | ng    | Qvalue 96 |
| -----                       |        |      |          |         |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF093021\  
Data File : BF125735.D  
Acq On : 30 Sep 2021 16:05  
Operator : JU/CG  
Sample : M3971-02  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
RO-COMP-1

Quant Time: Sep 30 17:16:35 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF092421.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 24 15:38:32 2021  
Response via : Initial Calibration

