

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100119\
 Data File : BF117278.D
 Acq On : 1 Oct 2019 15:53
 Operator : JU
 Sample : K5019-02DL2 100X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 QNWP8-MW30D-GWDL2

Quant Time: Oct 01 16:47:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 13 16:23:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	28183	20.00	ng	-0.02
21) Naphthalene-d8	8.09	136	127414	20.00	ng	-0.02
39) Acenaphthene-d10	9.84	164	70743	20.00	ng	-0.02
64) Phenanthrene-d10	11.33	188	124320	20.00	ng	0.00
76) Chrysene-d12	13.97	240	106022	20.00	ng	0.00
87) Perylene-d12	15.42	264	107278	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	6.49	99	119	0.05	ng	0.03
23) Nitrobenzene-d5	7.39	82	140	0.06	ng	0.00
42) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
45) 2-Fluorobiphenyl	9.16	172	1425	0.31	ng	-0.01
79) Terphenyl-d14	12.91	244	2857	0.50	ng	0.00
Target Compounds						
31) Naphthalene	8.11	128	183429	29.934	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100119\
Data File : BF117278.D
Acq On : 1 Oct 2019 15:53
Operator : JU
Sample : K5019-02DL2 100X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
QNWP8-MW30D-GWDL2

Quant Time: Oct 01 16:47:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 13 16:23:17 2019
Response via : Initial Calibration

