

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100318\
 Data File : BF109557.D
 Acq On : 3 Oct 2018 20:02
 Operator : JU/SJ
 Sample : J5084-05
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-7(8-8.5)

Manual Integrations
 APPROVED

Sohil
 10/4/2018 11:44:30 AM

Quant Time: Oct 04 02:38:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	51623	20.00	ng	0.04
21) Naphthalene-d8	8.02	136	123273	20.00	ng	0.02
38) Acenaphthene-d10	9.74	164	89295	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	185991	20.00	ng	0.00
75) Chrysene-d12	13.84	240	186007	20.00	ng	0.00
86) Perylene-d12	15.24	264	131761	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	85840	27.39	ng	0.06
7) Phenol-d6	6.38	99	222461	55.62	ng	0.02
23) Nitrobenzene-d5	7.32	82	214726m	102.83	ng	0.04
41) 2,4,6-Tribromophenol	10.52	330	93567	106.29	ng	-0.01
44) 2-Fluorobiphenyl	9.06	172	409665	69.19	ng	-0.01
78) Terphenyl-d14	12.79	244	527185	69.56	ng	-0.01
Target Compounds						
31) Naphthalene	8.05	128	4266418	740.632	ng	96
37) 2-Methylnaphthalene	8.73	142	2860106	770.185	ng	# 94
45) 1,1'-Biphenyl	9.16	154	76116	10.462	ng	95
49) Dimethylphthalate	9.46	163	111017	17.093	ng	# 98
57) Fluorene	10.28	166	16709	2.970	ng	# 83
70) Phenanthrene	11.24	178	36061	3.883	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100318\
Data File : BF109557.D
Acq On : 3 Oct 2018 20:02
Operator : JU/SJ
Sample : J5084-05
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-7(8-8.5)

Manual Integrations
APPROVED

Sohil
10/4/2018 11:44:30 AM

Quant Time: Oct 04 02:38:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

