

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
 Data File : BF130516.D
 Acq On : 04 Oct 2022 14:09
 Operator : CG\JU
 Sample : N4914-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Oct 05 02:16:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.634	152	81060	20.000	ng	-0.01
21) Naphthalene-d8	7.916	136	307745	20.000	ng	-0.01
39) Acenaphthene-d10	9.663	164	154126	20.000	ng	-0.02
64) Phenanthrene-d10	11.145	188	227391	20.000	ng	-0.01
76) Chrysene-d12	13.774	240	150574	20.000	ng	-0.02
86) Perylene-d12	15.163	264	176972	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.251	112	663839	142.044	ng	0.00
7) Phenol-d6	6.287	99	755359	129.174	ng	0.00
23) Nitrobenzene-d5	7.204	82	561178	93.161	ng	-0.01
42) 2,4,6-Tribromophenol	10.457	330	224447	151.980	ng	-0.01
45) 2-Fluorobiphenyl	8.992	172	989980	93.534	ng	-0.02
79) Terphenyl-d14	12.733	244	895648	98.532	ng	-0.01
Target Compounds						Qvalue
3) Pyridine	3.010	79	80062	14.299	ng	94
10) Phenol	6.298	94	442002	64.499	ng	94
15) Benzyl Alcohol	6.898	79	45771	9.872	ng	# 1
17) 2-Methylphenol	6.898	107	92629	21.404	ng	98
18) Hexachloroethane	7.057	117	15747	6.688	ng	# 4
19) n-Nitroso-di-n-propyla...	7.204	70	75217	19.062	ng	# 75
20) 3+4-Methylphenols	7.057	107	281823	50.129	ng	# 54
25) Isophorone	7.204	82	560428	57.723	ng	# 66
27) 2,4-Dimethylphenol	7.587	122	77918	20.183	ng	98
28) bis(2-Chloroethoxy)met...	7.640	93	32481	5.941	ng	# 81
31) Naphthalene	7.945	128	3789587	239.021	ng	98
32) Benzoic acid	7.722	122	83183	27.862	ng	# 12
33) 4-Chloroaniline	7.945	127	520754	86.360	ng	# 37
37) 2-Methylnaphthalene	8.628	142	206772	20.455	ng	99
38) 1-Methylnaphthalene	8.722	142	158233	16.294	ng	99
46) 1,1'-Biphenyl	9.092	154	32931	2.861	ng	98
49) Acenaphthylene	9.522	152	155412	11.132	ng	99
51) 2,6-Dinitrotoluene	9.663	165	21364	9.595	ng	# 17
55) Dibenzofuran	9.869	168	51601	4.044	ng	96
57) 2,4-Dinitrotoluene	9.663	165	21364	7.508	ng	# 17
58) Fluorene	10.210	166	61152	5.861	ng	99
67) 4-Bromophenyl-phenylether	10.451	248	13936	5.556	ng	# 1
71) Phenanthrene	11.169	178	118089	9.250	ng	98
72) Anthracene	11.169	178	118089	9.585	ng	99
73) Carbazole	11.380	167	205127	18.449	ng	97
75) Fluoranthene	12.351	202	26685	2.163	ng	99
77) Benzidine	12.486	184	33727	7.716	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
Data File : BF130516.D
Acq On : 04 Oct 2022 14:09
Operator : CG\JU
Sample : N4914-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :


Quant Time: Oct 05 02:16:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.634 min Scan# 71
Delta R.T. -0.011 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:152 Resp: 81060
Ion Ratio Lower Upper
152 100
150 154.6 122.8 184.2
115 65.3 53.8 80.8



#3
Pyridine
Concen: 14.299 ng
RT: 3.010 min Scan# 157
Delta R.T. 0.071 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion: 79 Resp: 80062
Ion Ratio Lower Upper
79 100
52 72.4 62.4 93.6
51 38.0 32.2 48.4









#5
2-Fluorophenol
Concen: 142.044 ng
RT: 5.251 min Scan# 51
Delta R.T. 0.006 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:112 Resp: 663839
Ion Ratio Lower Upper
112 100
64 58.8 51.4 77.0
63 32.3 28.8 43.2



#7
Phenol-d6
Concen: 129.174 ng
RT: 6.287 min Scan# 714
Delta R.T. -0.006 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion: 99 Resp: 755359
Ion Ratio Lower Upper
99 100
42 19.0 17.0 25.6
71 37.9 32.3 48.5





#10

Phenol

Concen: 64.499 ng

RT: 6.298 min Scan# 71

Delta R.T. -0.006 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F

ClientSampleId :





Tgt Ion: 94 Resp: 442002

Ion Ratio Lower Upper

94 100

65 31.3 14.1 54.1

66 43.6 28.1 68.1



#15

Benzyl Alcohol

Concen: 9.872 ng

RT: 6.898 min Scan# 818

Delta R.T. 0.100 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09




Tgt Ion: 79 Resp: 45771


Ion Ratio Lower Upper

79 100

108 230.1 54.0 81.0#


77 105.3 51.3 76.9#








#17
2-Methylphenol
Concen: 21.404 ng
RT: 6.898 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:	107	Resp:	92629
Ion	Ratio	Lower	Upper
107	100		
108	115.1	89.4	134.0
77	52.7	40.6	60.8
79	50.0	40.2	60.2



#18
Hexachloroethane
Concen: 6.688 ng
RT: 7.057 min Scan# 845
Delta R.T. -0.100 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09


Tgt Ion:	117	Resp:	15747
Ion	Ratio	Lower	Upper
117	100		
119	0.0	77.8	116.6#
201	0.0	68.7	103.1#








#19
n-Nitroso-di-n-propylamine
Concen: 19.062 ng
RT: 7.204 min Scan# 819
Delta R.T. 0.135 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09


Instrument :
BNA_F
ClientSampleId :





Tgt Ion: 70 Resp: 75217
Ion Ratio Lower Upper
70 100
42 47.8 50.8 76.2#
101 0.0 8.2 12.4#
130 2.1 16.6 25.0#




#20
3+4-Methylphenols
Concen: 50.129 ng
RT: 7.057 min Scan# 845
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Tgt Ion: 107 Resp: 281823
Ion Ratio Lower Upper
107 100
108 94.0 74.8 114.8
77 37.9 125.8 165.8#
79 36.9 17.1 57.1








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 7.916 min Scan# 991
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:136	Resp: 307745
Ion Ratio	Lower Upper
136 100	
137 10.6	9.1 13.7
54 8.1	7.1 10.7
68 6.0	4.5 6.7




#23
Nitrobenzene-d5
Concen: 93.161 ng
RT: 7.204 min Scan# 870
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09



Tgt Ion: 82	Resp: 561178
Ion Ratio	Lower Upper
82 100	
128 43.2	32.6 49.0
54 49.8	43.8 65.6









#25
Isophorone
Concen: 57.723 ng
RT: 7.204 min Scan# 81
Delta R.T. -0.270 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 82 Resp: 560428
Ion Ratio Lower Upper
82 100
95 0.0 6.4 9.6#
138 0.0 13.8 20.6#



#27
2,4-Dimethylphenol
Concen: 20.183 ng
RT: 7.587 min Scan# 935
Delta R.T. -0.006 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion: 122 Resp: 77918
Ion Ratio Lower Upper
122 100
107 123.0 100.2 150.2
121 61.0 47.8 71.8









#28
bis(2-Chloroethoxy)methane
Concen: 5.941 ng
RT: 7.640 min Scan# 944
Delta R.T. -0.047 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 93 Resp: 32481
Ion Ratio Lower Upper
93 100
95 23.0 26.3 39.5#
123 3.4 9.4 14.2#



#31
Naphthalene
Concen: 239.021 ng
RT: 7.945 min Scan# 996
Delta R.T. -0.006 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion: 128 Resp: 3789587
Ion Ratio Lower Upper
128 100
129 12.3 9.1 13.7
127 14.0 10.6 16.0









#32
Benzoic acid
Concen: 27.862 ng
RT: 7.722 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:122 Resp: 83183
Ion Ratio Lower Upper
122 100
105 6.1 112.9 152.9#
77 39.2 82.8 122.8#



#33
4-Chloroaniline
Concen: 86.360 ng
RT: 7.945 min Scan# 996
Delta R.T. -0.059 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion:127 Resp: 520754
Ion Ratio Lower Upper
127 100
129 88.1 25.6 38.4#
65 4.7 27.0 40.6#
92 10.0 15.9 23.9#





#37

2-Methylnaphthalene

Concen: 20.455 ng

RT: 8.628 min Scan# 1112

Delta R.T. -0.012 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F





ClientSampleId :

Tgt Ion:142 Resp: 206772

Ion Ratio Lower Upper

142 100

141 88.1 70.0 105.0



#38

1-Methylnaphthalene

Concen: 16.294 ng

RT: 8.722 min Scan# 1128

Delta R.T. -0.018 min

Lab File: BF130516.D




Acq: 04 Oct 2022 14:09


Tgt Ion:142 Resp: 158233

Ion Ratio Lower Upper

142 100

141 90.9 73.8 110.8





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.663 min Scan# 11

Delta R.T. -0.018 min


Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:164 Resp: 154126

Ion Ratio Lower Upper

164 100

162 100.7 81.4 122.0

160 44.9 36.2 54.4



#42

2,4,6-Tribromophenol

Concen: 151.980 ng

RT: 10.457 min Scan# 1423

Delta R.T. -0.012 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09




Tgt Ion:330 Resp: 224447


Ion Ratio Lower Upper

330 100

332 104.8 83.4 125.2

141 35.9 34.0 51.0





#45

2-Fluorobiphenyl

Concen: 93.534 ng

RT: 8.992 min Scan# 1119

Delta R.T. -0.018 min


Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:172 Resp: 989980

Ion Ratio Lower Upper

172 100

171 35.1 28.3 42.5

170 23.3 18.7 28.1



#46

1,1'-Biphenyl

Concen: 2.861 ng

RT: 9.092 min Scan# 1191

Delta R.T. -0.012 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09




Tgt Ion:154 Resp: 32931


Ion Ratio Lower Upper

154 100

153 38.5 20.0 60.0

76 12.8 0.0 33.3





#49

Acenaphthylene

Concen: 11.132 ng

RT: 9.522 min Scan# 11

Delta R.T. -0.018 min


Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F

ClientSampleId :






Tgt Ion:152 Resp: 155412

Ion Ratio Lower Upper

152 100

151 19.6 16.2 24.2

153 13.4 10.6 15.8



#51

2,6-Dinitrotoluene

Concen: 9.595 ng

RT: 9.663 min Scan# 1288

Delta R.T. 0.188 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09




Tgt Ion:165 Resp: 21364


Ion Ratio Lower Upper

165 100

63 2.2 58.8 88.2#


89 1.9 54.5 81.7#








#55
Dibenzenofuran
Concen: 4.044 ng
RT: 9.869 min Scan# 11
Delta R.T. -0.018 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:168 Resp: 51601
Ion Ratio Lower Upper
168 100
139 40.7 35.0 52.4
169 13.8 10.9 16.3



#57
2,4-Dinitrotoluene
Concen: 7.508 ng
RT: 9.663 min Scan# 1288
Delta R.T. -0.212 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion:165 Resp: 21364
Ion Ratio Lower Upper
165 100
63 2.2 44.7 67.1#
89 1.9 69.0 103.4#
182 0.0 1.9 2.9#









#58
 Fluorene
 Concen: 5.861 ng
 RT: 10.210 min Scan# 1355
 Delta R.T. -0.018 min
 Lab File: BF130516.D
 Acq: 04 Oct 2022 14:09




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion:166 Resp: 61152
 Ion Ratio Lower Upper
 166 100
 165 99.2 78.8 118.2
 167 13.8 10.6 15.8



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.145 min Scan# 1540
 Delta R.T. -0.012 min
 Lab File: BF130516.D
 Acq: 04 Oct 2022 14:09


Tgt Ion:188 Resp: 227391
 Ion Ratio Lower Upper
 188 100
 94 8.5 7.4 11.2
 80 8.8 7.9 11.9








#67
4-Bromophenyl-phenylether
Concen: 5.556 ng
RT: 10.451 min Scan# 14
Delta R.T. -0.259 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:248 Resp: 13936
Ion Ratio Lower Upper
248 100
250 184.9 80.0 120.0#
141 587.3 68.5 102.7#




#71
Phenanthrene
Concen: 9.250 ng
RT: 11.169 min Scan# 1544
Delta R.T. -0.018 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09



Tgt Ion:178 Resp: 118089
Ion Ratio Lower Upper
178 100
176 18.8 15.4 23.0
179 14.6 12.4 18.6





#72

Anthracene

Concen: 9.585 ng

RT: 11.169 min Scan# 11

Delta R.T. -0.065 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:178 Resp: 118089

Ion Ratio Lower Upper

178 100

176 18.8 15.0 22.6

179 14.6 12.7 19.1



#73

Carbazole

Concen: 18.449 ng

RT: 11.380 min Scan# 1580

Delta R.T. -0.012 min

Lab File: BF130516.D

Acq: 04 Oct 2022 14:09




Tgt Ion:167 Resp: 205127


Ion Ratio Lower Upper

167 100

166 20.5 17.4 26.2


139 13.5 11.6 17.4








#75
Fluoranthene
Concen: 2.163 ng
RT: 12.351 min Scan# 11
Delta R.T. -0.018 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:202 Resp: 26685
Ion Ratio Lower Upper
202 100
101 11.1 0.0 31.8
203 17.6 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.774 min Scan# 1987
Delta R.T. -0.018 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Tgt Ion:240 Resp: 150574
Ion Ratio Lower Upper
240 100
120 11.2 8.7 13.1
236 25.4 20.9 31.3









#77
Benzidine
Concen: 7.716 ng
RT: 12.486 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:184 Resp: 33727
Ion Ratio Lower Upper
184 100
185 14.8 11.8 17.6
183 11.1 9.1 13.7



#79
Terphenyl-d14
Concen: 98.532 ng
RT: 12.733 min Scan# 1810
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09


Tgt Ion:244 Resp: 895648
Ion Ratio Lower Upper
244 100
212 7.1 5.9 8.9
122 10.3 8.5 12.7





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.163 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF130516.D
Acq: 04 Oct 2022 14:09

Instrument :
BNA_F
ClientSampleId :



Tgt Ion:264 Resp: 176972
Ion Ratio Lower Upper
264 100
260 22.9 19.3 28.9
265 21.3 17.0 25.4

