

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
 Data File : BF130521.D
 Acq On : 04 Oct 2022 16:43
 Operator : CG\JU
 Sample : N4909-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PE-4

Quant Time: Oct 05 02:19:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.634	152	77628	20.000	ng	-0.01
21) Naphthalene-d8	7.910	136	308270	20.000	ng	-0.02
39) Acenaphthene-d10	9.663	164	146066	20.000	ng	-0.02
64) Phenanthrene-d10	11.145	188	210939	20.000	ng	-0.01
76) Chrysene-d12	13.774	240	149302	20.000	ng	-0.02
86) Perylene-d12	15.163	264	183501	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.240	112	553750	123.726	ng	0.00
7) Phenol-d6	6.275	99	688918	123.020	ng	-0.02
23) Nitrobenzene-d5	7.198	82	441473	73.164	ng	-0.02
42) 2,4,6-Tribromophenol	10.451	330	143687	102.664	ng	-0.02
45) 2-Fluorobiphenyl	8.992	172	778035	77.565	ng	-0.02
79) Terphenyl-d14	12.727	244	592606	65.749	ng	-0.02
Target Compounds						
77) Benzidine	12.486	184	26978	6.225	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
Data File : BF130521.D
Acq On : 04 Oct 2022 16:43
Operator : CG\JU
Sample : N4909-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PE-4

Quant Time: Oct 05 02:19:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

