

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
 Data File : BF130525.D
 Acq On : 04 Oct 2022 18:47
 Operator : CG\JU
 Sample : N4917-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 265393-5

Quant Time: Oct 05 02:20:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.634	152	77847	20.000	ng	-0.01
21) Naphthalene-d8	7.910	136	292523	20.000	ng	-0.02
39) Acenaphthene-d10	9.663	164	138641	20.000	ng	-0.02
64) Phenanthrene-d10	11.145	188	200649	20.000	ng	-0.01
76) Chrysene-d12	13.774	240	152736	20.000	ng	-0.02
86) Perylene-d12	15.162	264	187143	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.234	112	304732	67.896	ng	-0.01
7) Phenol-d6	6.275	99	228655	40.716	ng	-0.02
23) Nitrobenzene-d5	7.198	82	542251	94.703	ng	-0.02
42) 2,4,6-Tribromophenol	10.451	330	186901	140.692	ng	-0.02
45) 2-Fluorobiphenyl	8.992	172	949432	99.722	ng	-0.02
79) Terphenyl-d14	12.727	244	790817	85.768	ng	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100422\
Data File : BF130525.D
Acq On : 04 Oct 2022 18:47
Operator : CG\JU
Sample : N4917-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
265393-5

Quant Time: Oct 05 02:20:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

