

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100423\
 Data File : BF135578.D
 Acq On : 04 Oct 2023 19:35
 Operator : CG\JU
 Sample : 04651-01MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP1-COMPMS

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Oct 04 23:26:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 04 17:06:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.739	152	163778	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	638838	20.000	ng	# 0.00
39) Acenaphthene-d10	9.774	164	319841	20.000	ng	0.00
64) Phenanthrene-d10	11.269	188	622099	20.000	ng	0.00
76) Chrysene-d12	13.927	240	306504	20.000	ng	0.00
86) Perylene-d12	15.386	264	317463	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1281902	127.303	ng	0.01
7) Phenol-d6	6.398	99	1602596	125.162	ng	0.01
23) Nitrobenzene-d5	7.310	82	1031836	87.752	ng	0.00
42) 2,4,6-Tribromophenol	10.574	330	379060	119.260	ng	0.00
45) 2-Fluorobiphenyl	9.098	172	1738988	82.030	ng	0.00
79) Terphenyl-d14	12.863	244	1764665	89.251	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.528	88	181041	41.012	ng	94
3) Pyridine	3.293	79	460189	38.734	ng	95
4) n-Nitrosodimethylamine	3.240	42	300794	47.710	ng	93
6) Aniline	6.410	93	543274	32.356	ng	# 26
8) 2-Chlorophenol	6.528	128	512312	47.659	ng	98
9) Benzaldehyde	6.298	77	103112	14.136	ng	99
10) Phenol	6.410	94	595265	42.397	ng	87
11) bis(2-Chloroethyl)ether	6.481	93	539310	51.208	ng	99
12) 1,3-Dichlorobenzene	6.675	146	483185	44.229	ng	99
13) 1,4-Dichlorobenzene	6.757	146	498483	44.849	ng	99
14) 1,2-Dichlorobenzene	6.904	146	463059	44.863	ng	98
15) Benzyl Alcohol	6.892	79	408693	44.481	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.016	45	897939	45.233	ng	73
17) 2-Methylphenol	7.010	107	398784	42.039	ng	# 89
18) Hexachloroethane	7.239	117	190691	46.433	ng	91
19) n-Nitroso-di-n-propyla...	7.157	70	325390	41.028	ng	94
20) 3+4-Methylphenols	7.163	107	487012	42.696	ng	# 72
22) Acetophenone	7.151	105	656156	44.925	ng	# 92
24) Nitrobenzene	7.328	77	536768	46.185	ng	100
25) Isophorone	7.563	82	994825	46.074	ng	98
26) 2-Nitrophenol	7.639	139	269866	48.380	ng	99
27) 2,4-Dimethylphenol	7.686	122	393294	41.947	ng	98
28) bis(2-Chloroethoxy)met...	7.775	93	617598	47.789	ng	99
29) 2,4-Dichlorophenol	7.886	162	407859	46.945	ng	99
30) 1,2,4-Trichlorobenzene	7.963	180	417187	45.941	ng	99
31) Naphthalene	8.039	128	1389484	44.765	ng	99
32) Benzoic acid	7.839	122	318845	45.161	ng	100
33) 4-Chloroaniline	8.098	127	275097	20.201	ng	99
34) Hexachlorobutadiene	8.151	225	235488	43.007	ng	98
35) Caprolactam	8.480	113	130727m	44.834	ng	
36) 4-Chloro-3-methylphenol	8.598	107	460360	47.676	ng	96
37) 2-Methylnaphthalene	8.733	142	876800	42.024	ng	97
38) 1-Methylnaphthalene	8.833	142	845539	43.285	ng	100
40) 1,2,4,5-Tetrachloroben...	8.898	216	418701	45.186	ng	98
41) Hexachlorocyclopentadiene	8.880	237	232741	59.126	ng	98
43) 2,4,6-Trichlorophenol	9.022	196	277306	45.708	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100423\
 Data File : BF135578.D
 Acq On : 04 Oct 2023 19:35
 Operator : CG\JU
 Sample : 04651-01MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP1-COMPMS

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Oct 04 23:26:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 04 17:06:20 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	9.069	196	287389	41.134	ng	98	10/05/2023
46) 1,1'-Biphenyl	9.198	154	1091202	44.395	ng	97	Supervised By :mohammad ahmed
47) 2-Chloronaphthalene	9.222	162	836495	46.905	ng	98	
48) 2-Nitroaniline	9.333	65	338966	48.925	ng	99	
49) Acenaphthylene	9.639	152	1329302	44.851	ng	99	
50) Dimethylphthalate	9.510	163	1007964	46.612	ng	99	
51) 2,6-Dinitrotoluene	9.575	165	224438	47.378	ng	98	10/06/2023
52) Acenaphthene	9.810	154	796412	45.487	ng	97	
53) 3-Nitroaniline	9.745	138	198881	35.765	ng	99	
54) 2,4-Dinitrophenol	9.869	184	223928	83.011	ng	97	
55) Dibenzofuran	9.986	168	1108406	41.485	ng	99	
56) 4-Nitrophenol	9.933	139	272204	77.022	ng	98	
57) 2,4-Dinitrotoluene	9.986	165	259139	43.695	ng	# 76	
58) Fluorene	10.327	166	889234	43.293	ng	99	
59) 2,3,4,6-Tetrachlorophenol	10.110	232	242836	45.183	ng	99	
60) Diethylphthalate	10.210	149	999709	46.065	ng	99	
61) 4-Chlorophenyl-phenyle...	10.322	204	415092	42.070	ng	94	
62) 4-Nitroaniline	10.374	138	243605	45.575	ng	94	
63) Azobenzene	10.480	77	1034959	48.727	ng	96	
65) 4,6-Dinitro-2-methylph...	10.404	198	151894	45.970	ng	87	
66) n-Nitrosodiphenylamine	10.445	169	858701	45.594	ng	97	
67) 4-Bromophenyl-phenylether	10.810	248	275212	44.481	ng	95	
68) Hexachlorobenzene	10.880	284	294516	46.724	ng	96	
69) Atrazine	10.980	200	256022	50.521	ng	99	
70) Pentachlorophenol	11.080	266	238687	63.290	ng	97	
71) Phenanthrene	11.298	178	1329318	45.180	ng	98	
72) Anthracene	11.345	178	1363641	45.089	ng	100	
73) Carbazole	11.510	167	1317465	47.854	ng	100	
74) Di-n-butylphthalate	11.833	149	1585167	48.863	ng	99	
75) Fluoranthene	12.492	202	1412384	46.069	ng	99	
77) Benzidine	12.616	184	508765	81.023	ng	99	
78) Pyrene	12.721	202	1446913	49.583	ng	100	
80) Butylbenzylphthalate	13.339	149	602513	58.643	ng	94	
81) Benzo(a)anthracene	13.915	228	967562	48.887	ng	99	
82) 3,3'-Dichlorobenzidine	13.880	252	249782	40.404	ng	99	
83) Chrysene	13.957	228	956305	48.599	ng	99	
84) Bis(2-ethylhexyl)phtha...	13.904	149	697757	66.180	ng	99	
85) Di-n-octyl phthalate	14.515	149	1043164	62.259	ng	100	
87) Indeno(1,2,3-cd)pyrene	16.874	276	1033050	49.630	ng	96	
88) Benzo(b)fluoranthene	14.962	252	846083	49.233	ng	# 96	
89) Benzo(k)fluoranthene	14.992	252	722803	42.578	ng	# 96	
90) Benzo(a)pyrene	15.327	252	747788	45.629	ng	99	
91) Dibenzo(a,h)anthracene	16.886	278	847501	49.762	ng	99	
92) Benzo(g,h,i)perylene	17.321	276	884774	49.743	ng	98	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.739 min Scan# 79
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:152 Resp: 163778
Ion Ratio Lower Upper
152 100
150 153.0 127.8 191.6
115 56.6 49.8 74.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#2
1,4-Dioxane
Concen: 41.012 ng
RT: 2.528 min Scan# 75
Delta R.T. 0.106 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 88 Resp: 181041
Ion Ratio Lower Upper
88 100
58 91.3 78.3 117.5
43 34.2 29.4 44.0





#3
Pyridine
Concen: 38.734 ng
RT: 3.293 min Scan# 205
Delta R.T. 0.129 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 79 Resp: 460189
Ion Ratio Lower Upper
79 100
52 88.0 74.1 111.1
51 41.9 35.8 53.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023



#4
n-Nitrosodimethylamine
Concen: 47.710 ng
RT: 3.240 min Scan# 196
Delta R.T. 0.106 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion: 42 Resp: 300794
Ion Ratio Lower Upper
42 100
74 117.3 87.7 131.5
44 7.4 5.6 8.4







#5
2-Fluorophenol
Concen: 127.303 ng
RT: 5.381 min Scan# 501
Delta R.T. 0.012 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:112 Resp: 128190
Ion Ratio Lower Upper
112 100
64 67.9 55.9 83.9
63 36.1 29.0 43.6



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#6
Aniline
Concen: 32.356 ng
RT: 6.410 min Scan# 735
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 93 Resp: 543274
Ion Ratio Lower Upper
93 100
66 89.4 35.4 53.0#
65 61.9 17.8 26.8#





#7
Phenol-d6
Concen: 125.162 ng
RT: 6.398 min Scan# 711
Delta R.T. 0.012 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 99 Resp: 1602590
Ion Ratio Lower Upper
99 100
42 22.1 17.9 26.9
71 32.1 27.0 40.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#8
2-Chlorophenol
Concen: 47.659 ng
RT: 6.528 min Scan# 755
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:128 Resp: 512312
Ion Ratio Lower Upper
128 100
130 31.6 12.7 52.7
64 53.9 35.9 75.9







#9
Benzaldehyde
Concen: 14.136 ng
RT: 6.298 min Scan# 716
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion: 77 Resp: 10311
Ion Ratio Lower Upper
77 100
105 99.7 81.4 121.4
106 96.0 76.9 116.9



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#10
Phenol
Concen: 42.397 ng
RT: 6.410 min Scan# 735
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 94 Resp: 595265
Ion Ratio Lower Upper
94 100
65 41.5 15.9 55.9
66 59.9 29.8 69.8





#11
bis(2-Chloroethyl)ether
Concen: 51.208 ng
RT: 6.481 min Scan# 74
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 93 Resp: 539310
Ion Ratio Lower Upper
93 100
63 87.9 67.3 107.3
95 32.3 12.7 52.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#12
1,3-Dichlorobenzene
Concen: 44.229 ng
RT: 6.675 min Scan# 780
Delta R.T. -0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:146 Resp: 483185
Ion Ratio Lower Upper
146 100
148 63.8 51.3 76.9
75 35.6 27.4 41.2





#13

1,4-Dichlorobenzene

Concen: 44.849 ng

RT: 6.757 min Scan# 798

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

SP1-COMPMS

Tgt Ion:146 Resp: 49848

Ion Ratio Lower Upper

146 100

148 63.7 51.3 76.9



111 42.2 35.1 52.7

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



10/05/2023

Supervised By :mohammad
ahmed

10/06/2023



#14

1,2-Dichlorobenzene

Concen: 44.863 ng

RT: 6.904 min Scan# 819

Delta R.T. -0.006 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35




Tgt Ion:146 Resp: 463059


Ion Ratio Lower Upper

146 100

148 63.9 49.7 74.5

111 46.2 37.4 56.2







#15
Benzyl Alcohol
Concen: 44.481 ng
RT: 6.892 min Scan# 817
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion: 79 Resp: 40869
Ion Ratio Lower Upper
79 100
108 79.4 61.2 91.8
77 61.6 50.5 75.7



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#16
2,2'-oxybis(1-Chloropropane)
Concen: 45.233 ng
RT: 7.016 min Scan# 838
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 45 Resp: 897939
Ion Ratio Lower Upper
45 100
77 20.5 0.0 30.6
79 17.9 0.0 28.3





#17
2-Methylphenol
Concen: 42.039 ng
RT: 7.010 min Scan# 810
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:107 Resp: 398784
Ion Ratio Lower Upper
107 100
108 119.9 92.0 138.0
77 62.3 36.7 55.1
79 55.0 36.3 54.5#


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#18
Hexachloroethane
Concen: 46.433 ng
RT: 7.239 min Scan# 876
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:117 Resp: 190691
Ion Ratio Lower Upper
117 100
119 96.7 74.7 112.1
201 87.7 81.8 122.8





#19
n-Nitroso-di-n-propylamine
Concen: 41.028 ng
RT: 7.157 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 70 Resp: 325390
Ion Ratio Lower Upper
70 100
42 76.2 55.4 83.0
101 9.6 7.7 11.5
130 21.1 17.0 25.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#20
3+4-Methylphenols
Concen: 42.696 ng
RT: 7.163 min Scan# 863
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:107 Resp: 487012
Ion Ratio Lower Upper
107 100
108 93.4 74.7 114.7
77 60.4 102.0 142.0#
79 28.6 11.7 51.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:136 Resp: 638838
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 10.6 7.7 11.5
68 7.2 4.6 6.8#


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#22
Acetophenone
Concen: 44.925 ng
RT: 7.151 min Scan# 861
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:105 Resp: 656156
Ion Ratio Lower Upper
105 100
71 5.8 5.8 8.6#
51 38.2 37.0 55.4
120 21.8 17.0 25.6





#23

Nitrobenzene-d5

Concen: 87.752 ng

RT: 7.310 min Scan# 890

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

SP1-COMPMS

Tgt Ion: 82 Resp: 1031830

Ion Ratio Lower Upper

82 100

128 45.4 35.8 53.6



54 63.2 50.6 76.0

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



10/05/2023

Supervised By :mohammad
ahmed

10/06/2023



#24

Nitrobenzene

Concen: 46.185 ng

RT: 7.328 min Scan# 891

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35




Tgt Ion: 77 Resp: 536768

Ion Ratio Lower Upper

77 100

123 47.0 37.6 56.4


65 14.9 12.2 18.4





#25
Isophorone
Concen: 46.074 ng
RT: 7.563 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 82 Resp: 99482
Ion Ratio Lower Upper
82 100
95 6.6 6.0 9.0
138 17.4 14.5 21.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#26
2-Nitrophenol
Concen: 48.380 ng
RT: 7.639 min Scan# 944
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:139 Resp: 269866
Ion Ratio Lower Upper
139 100
109 26.6 22.6 33.8
65 58.4 47.0 70.4





#27
2,4-Dimethylphenol
Concen: 41.947 ng
RT: 7.686 min Scan# 91
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:122 Resp: 393294
Ion Ratio Lower Upper
122 100
107 109.4 90.2 135.4
121 57.4 46.3 69.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#28
bis(2-Chloroethoxy)methane
Concen: 47.789 ng
RT: 7.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 93 Resp: 617598
Ion Ratio Lower Upper
93 100
95 32.9 26.2 39.2
123 11.4 8.8 13.2





#29
2,4-Dichlorophenol
Concen: 46.945 ng
RT: 7.886 min Scan# 986
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:162 Resp: 407855
Ion Ratio Lower Upper
162 100
164 63.2 44.1 84.1
98 39.2 20.3 60.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023




#30
1,2,4-Trichlorobenzene
Concen: 45.941 ng
RT: 7.963 min Scan# 999
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35




Tgt Ion:180 Resp: 417187
Ion Ratio Lower Upper
180 100
182 95.7 76.9 115.3
145 29.8 24.5 36.7





#31
Naphthalene
Concen: 44.765 ng
RT: 8.039 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:128 Resp: 1389484
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 13.3 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#32
Benzoic acid
Concen: 45.161 ng
RT: 7.839 min Scan# 978
Delta R.T. 0.018 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:122 Resp: 318845
Ion Ratio Lower Upper
122 100
105 119.6 99.9 139.9
77 89.5 70.0 110.0







#33
4-Chloroaniline
Concen: 20.201 ng
RT: 8.098 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:	127	Resp:	27509
Ion Ratio	Lower	Upper	
127	100		
129	32.2	25.8	38.8
65	33.3	26.6	39.8
92	18.5	15.4	23.2



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#34
Hexachlorobutadiene
Concen: 43.007 ng
RT: 8.151 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:	225	Resp:	235488
Ion Ratio	Lower	Upper	
225	100		
223	61.5	50.6	75.8
227	62.7	50.8	76.2





#35
Caprolactam
Concen: 44.834 ng m
RT: 8.480 min Scan# 1090
Delta R.T. 0.012 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:113 Resp: 13072
Ion Ratio Lower Upper
113 100
55 222.8 199.3 239.3
56 158.4 146.3 186.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#36
4-Chloro-3-methylphenol
Concen: 47.676 ng
RT: 8.598 min Scan# 1107
Delta R.T. 0.012 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:107 Resp: 460360
Ion Ratio Lower Upper
107 100
144 25.6 19.0 28.4
142 78.9 60.2 90.2





#37

2-Methylnaphthalene

Concen: 42.024 ng

RT: 8.733 min Scan# 1134

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

SP1-COMPMS

Tgt Ion:142 Resp: 876800

Ion Ratio Lower Upper



142 100

141 86.1 71.0 106.6

Manual Integrations



APPROVED

Reviewed By :Yogesh
Patel



10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#38

1-Methylnaphthalene

Concen: 43.285 ng

RT: 8.833 min Scan# 1147

Delta R.T. 0.006 min

Lab File: BF135578.D




Acq: 04 Oct 2023 19:35


Tgt Ion:142 Resp: 845539

Ion Ratio Lower Upper

142 100


141 90.6 72.4 108.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.774 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:164 Resp: 31984
Ion Ratio Lower Upper
164 100
162 101.1 81.0 121.4
160 47.0 37.0 55.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023



#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.186 ng
RT: 8.898 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35




Tgt Ion:216 Resp: 418701
Ion Ratio Lower Upper
216 100
214 78.2 63.7 95.5
179 20.2 16.7 25.1
108 19.9 14.9 22.3





#41
Hexachlorocyclopentadiene
Concen: 59.126 ng
RT: 8.880 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:237 Resp: 23274
Ion Ratio Lower Upper
237 100
235 63.2 41.8 81.8
272 13.3 0.0 33.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023




#42
2,4,6-Tribromophenol
Concen: 119.260 ng
RT: 10.574 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:330 Resp: 379060
Ion Ratio Lower Upper
330 100
332 95.1 77.3 115.9
141 33.0 25.8 38.8







#43
2,4,6-Trichlorophenol
Concen: 45.708 ng
RT: 9.022 min Scan# 1182
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:196 Resp: 277300
Ion Ratio Lower Upper
196 100
198 95.6 77.0 115.6
200 30.4 24.5 36.7



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#44
2,4,5-Trichlorophenol
Concen: 41.134 ng
RT: 9.069 min Scan# 1187
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:196 Resp: 287389
Ion Ratio Lower Upper
196 100
198 97.4 77.5 116.3
97 53.4 40.5 60.7
132 29.7 23.3 34.9





#45
2-Fluorobiphenyl
Concen: 82.030 ng
RT: 9.098 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:172 Resp: 173898
Ion Ratio Lower Upper
172 100
171 35.1 28.4 42.6
170 24.3 19.0 28.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023



#46
1,1'-Biphenyl
Concen: 44.395 ng
RT: 9.198 min Scan# 1209
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:154 Resp: 1091202
Ion Ratio Lower Upper
154 100
153 41.0 19.6 59.6
76 13.8 0.0 31.0






#47
2-Chloronaphthalene
Concen: 46.905 ng
RT: 9.222 min Scan# 1213
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :



SP1-COMPMS



Tgt Ion:162 Resp: 83649
Ion Ratio Lower Upper
162 100
127 42.2 32.6 48.8
164 33.1 26.2 39.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#48
2-Nitroaniline
Concen: 48.925 ng
RT: 9.333 min Scan# 1232
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion: 65 Resp: 338966
Ion Ratio Lower Upper
65 100
92 61.2 49.4 74.0
138 95.5 77.4 116.0





#49
Acenaphthylene
Concen: 44.851 ng
RT: 9.639 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:152 Resp: 132930
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6
153 13.5 10.4 15.6



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#50
Dimethylphthalate
Concen: 46.612 ng
RT: 9.510 min Scan# 1262
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:163 Resp: 1007964
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.4 8.1 12.1





#51
2,6-Dinitrotoluene
Concen: 47.378 ng
RT: 9.575 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:165 Resp: 224438
Ion Ratio Lower Upper
165 100
63 73.0 59.4 89.2
89 60.3 49.8 74.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#52
Acenaphthene
Concen: 45.487 ng
RT: 9.810 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:154 Resp: 796412
Ion Ratio Lower Upper
154 100
153 110.9 92.1 138.1
152 53.7 43.1 64.7





#53
3-Nitroaniline
Concen: 35.765 ng
RT: 9.745 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:138 Resp: 19888
Ion Ratio Lower Upper
138 100
108 10.4 8.5 12.7
92 118.3 95.4 143.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#54
2,4-Dinitrophenol
Concen: 83.011 ng
RT: 9.869 min Scan# 1323
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:184 Resp: 223928
Ion Ratio Lower Upper
184 100
63 67.4 55.1 82.7
154 57.0 48.4 72.6





#55
Dibenzenofuran
Concen: 41.485 ng
RT: 9.986 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:168 Resp: 1108400
Ion Ratio Lower Upper
168 100
139 36.8 29.0 43.6
169 13.5 10.4 15.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#56
4-Nitrophenol
Concen: 77.022 ng
RT: 9.933 min Scan# 1334
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:139 Resp: 272204
Ion Ratio Lower Upper
139 100
109 70.5 49.8 89.8
65 115.0 92.7 132.7





#57
2,4-Dinitrotoluene
Concen: 43.695 ng
RT: 9.986 min Scan# 1346
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:165 Resp: 259139
Ion Ratio Lower Upper
165 100
63 82.8 50.1 75.1
89 92.8 59.3 88.9
182 2.2 2.3 3.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#58
Fluorene
Concen: 43.293 ng
RT: 10.327 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:166 Resp: 889234
Ion Ratio Lower Upper
166 100
165 92.0 74.3 111.5
167 13.7 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 45.183 ng
RT: 10.110 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:232 Resp: 242830
Ion Ratio Lower Upper
232 100
131 43.9 35.4 53.2
130 2.4 1.8 2.8
166 30.1 24.3 36.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#60
Diethylphthalate
Concen: 46.065 ng
RT: 10.210 min Scan# 1381
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:149 Resp: 999709
Ion Ratio Lower Upper
149 100
177 21.6 17.8 26.6
150 12.4 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 42.070 ng
RT: 10.322 min Scan# 1400
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:204 Resp: 415092
Ion Ratio Lower Upper
204 100
206 32.9 26.7 40.1
141 56.6 50.6 75.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#62
4-Nitroaniline
Concen: 45.575 ng
RT: 10.374 min Scan# 1409
Delta R.T. 0.012 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:138 Resp: 243605
Ion Ratio Lower Upper
138 100
92 47.7 31.8 71.8
108 67.5 52.0 92.0





#63
Azobenzene
Concen: 48.727 ng
RT: 10.480 min Scan# 1432
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion: 77 Resp: 1034959
Ion Ratio Lower Upper
77 100
182 22.1 6.4 46.4
105 22.7 3.1 43.1
51 38.3 16.1 56.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.269 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:188 Resp: 622099
Ion Ratio Lower Upper
188 100
94 10.0 7.0 10.6
80 10.3 7.5 11.3







#65
4,6-Dinitro-2-methylphenol
Concen: 45.970 ng
RT: 10.404 min Scan# 1415
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:198 Resp: 151894
Ion Ratio Lower Upper
198 100
51 54.0 47.7 87.7
105 42.7 27.4 67.4



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#66
n-Nitrosodiphenylamine
Concen: 45.594 ng
RT: 10.445 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:169 Resp: 858701
Ion Ratio Lower Upper
169 100
168 68.3 52.8 79.2
167 37.5 29.0 43.6







#67
4-Bromophenyl-phenylether
Concen: 44.481 ng
RT: 10.810 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:248 Resp: 275212
Ion Ratio Lower Upper
248 100
250 97.4 78.6 117.8
141 75.6 54.2 81.2



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#68
Hexachlorobenzene
Concen: 46.724 ng
RT: 10.880 min Scan# 1495
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:284 Resp: 294516
Ion Ratio Lower Upper
284 100
142 30.9 22.2 33.4
249 28.8 22.5 33.7





#69
Atrazine
Concen: 50.521 ng
RT: 10.980 min Scan# 1512
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:200 Resp: 25602
Ion Ratio Lower Upper
200 100
173 28.3 7.4 47.4
215 48.9 28.6 68.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#70
Pentachlorophenol
Concen: 63.290 ng
RT: 11.080 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:266 Resp: 238687
Ion Ratio Lower Upper
266 100
268 61.9 52.1 78.1
264 62.4 50.8 76.2







#71
Phenanthrene
Concen: 45.180 ng
RT: 11.298 min Scan# 1566
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:178 Resp: 1329318
Ion Ratio Lower Upper
178 100
176 20.4 15.6 23.4
179 15.7 12.0 18.0



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#72
Anthracene
Concen: 45.089 ng
RT: 11.345 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:178 Resp: 1363641
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 16.0 12.6 19.0





#73
 Carbazole
 Concen: 47.854 ng
 RT: 11.510 min Scan# 1602
 Delta R.T. 0.000 min
 Lab File: BF135578.D
 Acq: 04 Oct 2023 19:35



Instrument :
 BNA_F
 ClientSampleId :
 SP1-COMPMS



Tgt Ion:167 Resp: 131746
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.0
 139 12.9 10.6 15.8


Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




10/05/2023
 Supervised By :mohammad
 ahmed



10/06/2023




#74
 Di-n-butylphthalate
 Concen: 48.863 ng
 RT: 11.833 min Scan# 1657
 Delta R.T. 0.000 min
 Lab File: BF135578.D
 Acq: 04 Oct 2023 19:35




Tgt Ion:149 Resp: 1585167
 Ion Ratio Lower Upper
 149 100
 150 9.7 7.4 11.0
 104 4.9 3.6 5.4





#75
Fluoranthene
Concen: 46.069 ng
RT: 12.492 min Scan# 1772
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:202 Resp: 1412384
Ion Ratio Lower Upper
202 100
101 11.4 0.0 31.3
203 18.3 0.0 37.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023




#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.927 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:240 Resp: 306504
Ion Ratio Lower Upper
240 100
120 11.7 9.4 14.2
236 25.8 19.8 29.8







#77
Benzidine
Concen: 81.023 ng
RT: 12.616 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS

Tgt Ion:184 Resp: 50876
Ion Ratio Lower Upper
184 100
185 14.4 11.7 17.5
183 10.0 8.2 12.4



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#78
Pyrene
Concen: 49.583 ng
RT: 12.721 min Scan# 1808
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35

Tgt Ion:202 Resp: 1446913
Ion Ratio Lower Upper
202 100
200 20.4 16.2 24.2
203 17.7 14.2 21.4





#79

Terphenyl-d14

Concen: 89.251 ng

RT: 12.863 min Scan# 1832

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

SP1-COMPMS

Tgt Ion:244 Resp: 176466

Ion Ratio Lower Upper

244 100

212 6.9 5.1 7.7



122 11.8 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



10/05/2023

Supervised By :mohammad
ahmed

10/06/2023



#80

Butylbenzylphthalate

Concen: 58.643 ng

RT: 13.339 min Scan# 1913

Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35




Tgt Ion:149 Resp: 602513


Ion Ratio Lower Upper

149 100

91 75.0 54.7 82.1


206 19.9 16.3 24.5





#81
Benzo(a)anthracene
Concen: 48.887 ng
RT: 13.915 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:228 Resp: 967562
Ion Ratio Lower Upper
228 100
226 27.6 21.7 32.5
229 19.8 15.8 23.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed



10/06/2023




#82
3,3'-Dichlorobenzidine
Concen: 40.404 ng
RT: 13.880 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:252 Resp: 249782
Ion Ratio Lower Upper
252 100
254 63.0 51.0 76.4
126 11.4 7.8 11.8





#83

Chrysene

Concen: 48.599 ng

RT: 13.957 min Scan# 2009

Delta R.T. 0.006 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35

Instrument :

BNA_F



ClientSampleId :

SP1-COMPMS

Tgt Ion:228 Resp: 956309
 Ion Ratio Lower Upper
 228 100
 226 30.8 24.3 36.5
 229 19.6 16.0 24.0



Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel



10/05/2023
 Supervised By :mohammad
 ahmed

10/06/2023



#84

Bis(2-ethylhexyl)phthalate

Concen: 66.180 ng




RT: 13.904 min Scan# 2009


Delta R.T. 0.000 min

Lab File: BF135578.D

Acq: 04 Oct 2023 19:35


Tgt Ion:149 Resp: 697757
 Ion Ratio Lower Upper
 149 100
 167 27.9 21.9 32.9
 279 3.7 3.3 4.9





#85
Di-n-octyl phthalate
Concen: 62.259 ng
RT: 14.515 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:149 Resp: 1043164
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 12.3 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.386 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:264 Resp: 317463
Ion Ratio Lower Upper
264 100
260 23.6 19.8 29.6
265 21.4 17.0 25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 49.630 ng
RT: 16.874 min Scan# 2189
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:276 Resp: 1033050
Ion Ratio Lower Upper
276 100
138 27.3 19.2 28.8
277 25.4 19.7 29.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#88
Benzo(b)fluoranthene
Concen: 49.233 ng
RT: 14.962 min Scan# 2189
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:252 Resp: 846083
Ion Ratio Lower Upper
252 100
253 22.5 17.2 25.8
125 11.6 7.1 10.7#





#89
Benzo(k)fluoranthene
Concen: 42.578 ng
RT: 14.992 min Scan# 2198
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:252 Resp: 722800
Ion Ratio Lower Upper
252 100
253 22.9 17.4 26.0
125 11.2 7.2 10.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






10/05/2023
Supervised By :mohammad
ahmed


10/06/2023



#90
Benzo(a)pyrene
Concen: 45.629 ng
RT: 15.327 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35


Tgt Ion:252 Resp: 747788
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 12.2 9.2 13.8





#91
Dibenzo(a,h)anthracene
Concen: 49.762 ng
RT: 16.886 min Scan# 2518
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Instrument :
BNA_F
ClientSampleId :
SP1-COMPMS



Tgt Ion:278 Resp: 84750
Ion Ratio Lower Upper
278 100
139 17.1 13.4 20.0
279 23.1 19.0 28.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




10/05/2023
Supervised By :mohammad
ahmed

10/06/2023



#92
Benzo(g,h,i)perylene
Concen: 49.743 ng
RT: 17.321 min Scan# 2590
Delta R.T. 0.006 min
Lab File: BF135578.D
Acq: 04 Oct 2023 19:35



Tgt Ion:276 Resp: 884774
Ion Ratio Lower Upper
276 100
277 23.6 18.8 28.2
138 22.5 16.9 25.3

