

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100923\
 Data File : BF135659.D
 Acq On : 09 Oct 2023 19:38
 Operator : CG\JU
 Sample : 04731-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DTP-1

Quant Time: Oct 10 02:27:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 04 17:06:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.739	152	129002	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	487652	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	243451	20.000	ng	0.00
64) Phenanthrene-d10	11.269	188	401067	20.000	ng	0.00
76) Chrysene-d12	13.933	240	254136	20.000	ng	0.00
86) Perylene-d12	15.404	264	271009	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	834084	105.161	ng	0.02
7) Phenol-d6	6.392	99	949039	94.100	ng	0.00
23) Nitrobenzene-d5	7.310	82	727565	81.058	ng	0.00
42) 2,4,6-Tribromophenol	10.574	330	262963	108.693	ng	0.00
45) 2-Fluorobiphenyl	9.098	172	1272783	78.877	ng	0.00
79) Terphenyl-d14	12.868	244	1060673	64.699	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100923\
Data File : BF135659.D
Acq On : 09 Oct 2023 19:38
Operator : CG\JU
Sample : 04731-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DTP-1

Quant Time: Oct 10 02:27:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 04 17:06:20 2023
Response via : Initial Calibration

