

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101123\
 Data File : BF135717.D
 Acq On : 11 Oct 2023 17:13
 Operator : CG\JU
 Sample : 04812-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-223

Quant Time: Oct 12 01:26:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 02:43:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.739	152	83119	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	321908	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	179378	20.000	ng	0.00
64) Phenanthrene-d10	11.269	188	320560	20.000	ng	0.00
76) Chrysene-d12	13.933	240	176502	20.000	ng	0.00
86) Perylene-d12	15.404	264	207717	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.387	112	684762	133.992	ng	0.02
7) Phenol-d6	6.392	99	789767	121.535	ng	0.00
23) Nitrobenzene-d5	7.310	82	576953	97.374	ng	0.00
42) 2,4,6-Tribromophenol	10.575	330	245044	137.466	ng	0.00
45) 2-Fluorobiphenyl	9.098	172	1094484	92.055	ng	0.00
79) Terphenyl-d14	12.868	244	1082061	95.036	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101123\
Data File : BF135717.D
Acq On : 11 Oct 2023 17:13
Operator : CG\JU
Sample : 04812-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-223

Quant Time: Oct 12 01:26:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 02:43:17 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.739 min Scan# 795
Delta R.T. 0.000 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13




Instrument :
BNA_F
ClientSampleId :
VNJ-223


Tgt Ion:152 Resp: 83119
Ion Ratio Lower Upper
152 100
150 162.2 127.8 191.6
115 62.2 49.8 74.8



#5
2-Fluorophenol
Concen: 133.992 ng
RT: 5.387 min Scan# 561
Delta R.T. 0.018 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13

Tgt Ion:112 Resp: 684762
Ion Ratio Lower Upper
112 100
64 68.9 55.9 83.9
63 35.7 29.0 43.6









#7
Phenol-d6
Concen: 121.535 ng
RT: 6.392 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13




Instrument :
BNA_F
ClientSampleId :
VNJ-223


Tgt Ion: 99 Resp: 789767
Ion Ratio Lower Upper
99 100
42 23.4 17.9 26.9
71 34.6 27.0 40.6



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13

Tgt Ion: 136 Resp: 321908
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 10.3 7.7 11.5
68 5.3 4.6 6.8





#23

Nitrobenzene-d5

Concen: 97.374 ng

RT: 7.310 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13

Instrument :

BNA_F

ClientSampleId :

VNJ-223





Tgt Ion: 82 Resp: 576953

Ion Ratio Lower Upper

82 100

128 45.9 35.8 53.6

54 63.4 50.6 76.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.775 min Scan# 1307

Delta R.T. -0.006 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13




Tgt Ion: 164 Resp: 179378


Ion Ratio Lower Upper

164 100

162 102.4 81.0 121.4


160 45.1 37.0 55.4








#42
2,4,6-Tribromophenol
Concen: 137.466 ng
RT: 10.575 min Scan# 1443
Delta R.T. -0.000 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13

Instrument :
BNA_F
ClientSampleId :
VNJ-223







Tgt Ion:330 Resp: 245044
Ion Ratio Lower Upper
330 100
332 94.3 77.3 115.9
141 32.3 25.8 38.8



#45
2-Fluorobiphenyl
Concen: 92.055 ng
RT: 9.098 min Scan# 1192
Delta R.T. -0.000 min
Lab File: BF135717.D
Acq: 11 Oct 2023 17:13

Tgt Ion:172 Resp: 1094484
Ion Ratio Lower Upper
172 100
171 36.2 28.4 42.6
170 24.4 19.0 28.4





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.269 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13

Instrument :

BNA_F

ClientSampleId :

VNJ-223





Tgt Ion:188 Resp: 320560

Ion Ratio Lower Upper

188 100

94 8.8 7.0 10.6

80 9.5 7.5 11.3



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.933 min Scan# 2014

Delta R.T. -0.006 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13




Tgt Ion:240 Resp: 176502


Ion Ratio Lower Upper

240 100

120 10.4 9.4 14.2

236 25.8 19.8 29.8





#79

Terphenyl-d14

Concen: 95.036 ng

RT: 12.868 min Scan# 1833

Delta R.T. -0.000 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13

Instrument :

BNA_F

ClientSampleId :

VNJ-223





Tgt Ion:244 Resp: 1082061

Ion Ratio Lower Upper

244 100

212 6.7 5.1 7.7

122 10.5 7.9 11.9



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.404 min Scan# 2264

Delta R.T. -0.000 min

Lab File: BF135717.D

Acq: 11 Oct 2023 17:13

Tgt Ion:264 Resp: 207717

Ion Ratio Lower Upper

264 100

260 23.1 19.8 29.6

265 21.0 17.0 25.4

