

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101221\
 Data File : BF125801.D
 Acq On : 12 Oct 2021 14:41
 Operator : JU/CG
 Sample : M4077-10RE
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CC-34RE

Quant Time: Oct 12 16:27:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 07 17:16:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	74172	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	304743	20.00	ng	# 0.00
39) Acenaphthene-d10	9.874	164	156072	20.00	ng	0.00
64) Phenanthrene-d10	11.357	188	232026	20.00	ng	0.00
76) Chrysene-d12	13.992	240	157780	20.00	ng	0.00
86) Perylene-d12	15.445	264	127444	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.439	112	337771	74.39	ng	-0.02
7) Phenol-d6	6.457	99	408529	69.89	ng	-0.02
23) Nitrobenzene-d5	7.398	82	233007	47.53	ng	-0.01
42) 2,4,6-Tribromophenol	10.657	330	100678	65.34	ng	-0.01
45) 2-Fluorobiphenyl	9.198	172	519757	57.44	ng	0.00
79) Terphenyl-d14	12.939	244	434794	41.98	ng	0.00
Target Compounds						Qvalue
74) Di-n-butylphthalate	11.921	149	795690	62.46	ng	98
80) Butylbenzylphthalate	13.415	149	128995	25.11	ng	97
84) Bis(2-ethylhexyl)phtha...	13.980	149	624699	111.74	ng	99
85) Di-n-octyl phthalate	14.592	149	138005	14.82	ng	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101221\
Data File : BF125801.D
Acq On : 12 Oct 2021 14:41
Operator : JU/CG
Sample : M4077-10RE
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CC-34RE

Quant Time: Oct 12 16:27:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 07 17:16:23 2021
Response via : Initial Calibration

