

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101223\
 Data File : BF135743.D
 Acq On : 12 Oct 2023 19:35
 Operator : CG\JU
 Sample : 04657-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 V-2(0-2IN)

Quant Time: Oct 13 00:58:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 11 02:43:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.740	152	69084	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	237513	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	112956	20.000	ng	0.00
64) Phenanthrene-d10	11.275	188	153373	20.000	ng	0.00
76) Chrysene-d12	13.939	240	108940	20.000	ng	0.00
86) Perylene-d12	15.415	264	94490	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	380694	89.627	ng	0.01
7) Phenol-d6	6.393	99	465796	86.243	ng	0.00
23) Nitrobenzene-d5	7.304	82	264662	60.540	ng	0.00
42) 2,4,6-Tribromophenol	10.575	330	69674	62.070	ng	0.00
45) 2-Fluorobiphenyl	9.092	172	419537	56.036	ng	0.00
79) Terphenyl-d14	12.869	244	276461	39.340	ng	0.00
Target Compounds						
						Qvalue
19) n-Nitroso-di-n-propyla...	7.304	70	35355	10.568	ng	# 82
25) Isophorone	7.304	82	264662	32.969	ng	# 65
31) Naphthalene	8.040	128	54337	4.709	ng	98
37) 2-Methylnaphthalene	8.734	142	18059	2.328	ng	95
49) Acenaphthylene	9.639	152	69287	6.619	ng	99
51) 2,6-Dinitrotoluene	9.775	165	14969	8.947	ng	# 17
52) Acenaphthene	9.804	154	73372	11.866	ng	99
55) Dibenzofuran	9.981	168	48894	5.182	ng	95
56) 4-Nitrophenol	9.981	139	18895	15.139	ng	# 5
58) Fluorene	10.328	166	71057	9.796	ng	98
67) 4-Bromophenyl-phenylether	10.575	248	4654	3.051	ng	# 1
71) Phenanthrene	11.304	178	938451	129.372	ng	99
72) Anthracene	11.351	178	219836	29.484	ng	99
73) Carbazole	11.510	167	149238	21.987	ng	99
75) Fluoranthene	12.510	202	1627198	215.281	ng	98
78) Pyrene	12.739	202	1486919	143.360	ng	99
81) Benzo(a)anthracene	13.933	228	797149	113.319	ng	96
83) Chrysene	13.969	228	639402	91.422	ng	98
84) Bis(2-ethylhexyl)phtha...	13.910	149	123101	32.850	ng	98
87) Indeno(1,2,3-cd)pyrene	16.915	276	354938	57.291	ng	95
88) Benzo(b)fluoranthene	14.992	252	1111788	217.355	ng	# 96
89) Benzo(k)fluoranthene	14.992	252	1111788	220.034	ng	# 97
90) Benzo(a)pyrene	15.363	252	559009	114.600	ng	99
91) Dibenzo(a,h)anthracene	17.086	278	61760	12.183	ng	93
92) Benzo(g,h,i)perylene	17.368	276	342505	64.696	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101223\
Data File : BF135743.D
Acq On : 12 Oct 2023 19:35
Operator : CG\JU
Sample : 04657-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)

Quant Time: Oct 13 00:58:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 02:43:17 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.740 min Scan# 795
 Delta R.T. 0.001 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35




Instrument :
 BNA_F
 ClientSampleId :
 V-2(0-2IN)


Tgt Ion:152 Resp: 69084
 Ion Ratio Lower Upper
 152 100
 150 155.0 127.8 191.6
 115 59.7 49.8 74.8



#5
 2-Fluorophenol
 Concen: 89.627 ng
 RT: 5.381 min Scan# 560
 Delta R.T. 0.012 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35

Tgt Ion:112 Resp: 380694
 Ion Ratio Lower Upper
 112 100
 64 62.8 55.9 83.9
 63 33.5 29.0 43.6









#7
Phenol-d6
Concen: 86.243 ng
RT: 6.393 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion: 99 Resp: 465796
Ion Ratio Lower Upper
99 100
42 22.7 17.9 26.9
71 34.3 27.0 40.6



#19
n-Nitroso-di-n-propylamine
Concen: 10.568 ng
RT: 7.304 min Scan# 887
Delta R.T. 0.147 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 70 Resp: 35355
Ion Ratio Lower Upper
70 100
42 60.7 55.4 83.0
101 0.0 7.7 11.5#
130 2.0 17.0 25.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:136 Resp: 237513
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 10.8 7.7 11.5
68 5.6 4.6 6.8



#23
Nitrobenzene-d5
Concen: 60.540 ng
RT: 7.304 min Scan# 887
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 82 Resp: 264662
Ion Ratio Lower Upper
82 100
128 42.8 35.8 53.6
54 65.9 50.6 76.0









#25
Isophorone
Concen: 32.969 ng
RT: 7.304 min Scan# 81
Delta R.T. -0.259 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion: 82 Resp: 264662
Ion Ratio Lower Upper
82 100
95 0.0 6.0 9.0#
138 0.0 14.5 21.7#



#31
Naphthalene
Concen: 4.709 ng
RT: 8.040 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 128 Resp: 54337
Ion Ratio Lower Upper
128 100
129 10.5 8.7 13.1
127 12.3 10.7 16.1









#37
2-Methylnaphthalene
Concen: 2.328 ng
RT: 8.734 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:142 Resp: 18059
Ion Ratio Lower Upper
142 100
141 84.5 71.0 106.6



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.775 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35


Tgt Ion:164 Resp: 112956
Ion Ratio Lower Upper
164 100
162 102.1 81.0 121.4
160 45.8 37.0 55.4








#42
2,4,6-Tribromophenol
Concen: 62.070 ng
RT: 10.575 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)







Tgt Ion:330 Resp: 69674
Ion Ratio Lower Upper
330 100
332 97.6 77.3 115.9
141 33.6 25.8 38.8



#45
2-Fluorobiphenyl
Concen: 56.036 ng
RT: 9.092 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion:172 Resp: 419537
Ion Ratio Lower Upper
172 100
171 35.8 28.4 42.6
170 24.0 19.0 28.4





#49

Acenaphthylene

Concen: 6.619 ng

RT: 9.639 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

V-2(0-2IN)





Tgt Ion:152 Resp: 69287

Ion Ratio Lower Upper

152 100

151 20.3 16.4 24.6

153 12.3 10.4 15.6



#51

2,6-Dinitrotoluene

Concen: 8.947 ng

RT: 9.775 min Scan# 1307

Delta R.T. 0.194 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35




Tgt Ion:165 Resp: 14969


Ion Ratio Lower Upper

165 100

63 1.7 59.4 89.2#

89 1.4 49.8 74.8#









#52
Acenaphthene
Concen: 11.866 ng
RT: 9.804 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:154 Resp: 73372
Ion Ratio Lower Upper
154 100
153 114.3 92.1 138.1
152 54.5 43.1 64.7



#55
Dibenzofuran
Concen: 5.182 ng
RT: 9.981 min Scan# 1342
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion:168 Resp: 48894
Ion Ratio Lower Upper
168 100
139 39.8 29.0 43.6
169 14.5 10.4 15.6









#56
4-Nitrophenol
Concen: 15.139 ng
RT: 9.981 min Scan# 11
Delta R.T. 0.041 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:139 Resp: 18895
Ion Ratio Lower Upper
139 100
109 2.2 49.8 89.8#
65 2.9 92.7 132.7#



#58
Fluorene
Concen: 9.796 ng
RT: 10.328 min Scan# 1401
Delta R.T. -0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion:166 Resp: 71057
Ion Ratio Lower Upper
166 100
165 95.3 74.3 111.5
167 13.3 10.9 16.3





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.275 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

V-2(0-2IN)





Tgt Ion:188 Resp: 153373

Ion Ratio Lower Upper

188 100

94 8.5 7.0 10.6

80 8.9 7.5 11.3



#67

4-Bromophenyl-phenylether

Concen: 3.051 ng

RT: 10.575 min Scan# 1443

Delta R.T. -0.241 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35




Tgt Ion:248 Resp: 4654


Ion Ratio Lower Upper

248 100

250 177.2 78.6 117.8#

141 412.4 54.2 81.2#









#71
Phenanthrene
Concen: 129.372 ng
RT: 11.304 min Scan# 1567
Delta R.T. 0.006 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:178 Resp: 938451
Ion Ratio Lower Upper
178 100
176 19.9 15.6 23.4
179 15.5 12.0 18.0



#72
Anthracene
Concen: 29.484 ng
RT: 11.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion:178 Resp: 219836
Ion Ratio Lower Upper
178 100
176 18.5 15.4 23.2
179 15.5 12.6 19.0









#73
 Carbazole
 Concen: 21.987 ng
 RT: 11.510 min Scan# 1602
 Delta R.T. -0.006 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35




Instrument :
 BNA_F
 ClientSampleId :
 V-2(0-2IN)


Tgt Ion:167 Resp: 149238
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.4 26.0
 139 13.3 10.6 15.8



#75
 Fluoranthene
 Concen: 215.281 ng
 RT: 12.510 min Scan# 1772
 Delta R.T. 0.012 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35

Tgt Ion:202 Resp: 1627198
 Ion Ratio Lower Upper
 202 100
 101 10.3 0.0 31.3
 203 18.1 0.0 37.4





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.939 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

V-2(0-2IN)





Tgt Ion:240 Resp: 108940

Ion Ratio Lower Upper

240 100

120 13.0 9.4 14.2

236 25.7 19.8 29.8



#78

Pyrene

Concen: 143.360 ng

RT: 12.739 min Scan# 1811

Delta R.T. 0.012 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35




Tgt Ion:202 Resp: 1486919


Ion Ratio Lower Upper

202 100

200 20.6 16.2 24.2

203 17.7 14.2 21.4









#79
 Terphenyl-d14
 Concen: 39.340 ng
 RT: 12.869 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35




Instrument :
 BNA_F
 ClientSampleId :
 V-2(0-2IN)


Tgt Ion:244 Resp: 276461
 Ion Ratio Lower Upper
 244 100
 212 6.5 5.1 7.7
 122 9.9 7.9 11.9



#81
 Benzo(a)anthracene
 Concen: 113.319 ng
 RT: 13.933 min Scan# 2014
 Delta R.T. 0.006 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35

Tgt Ion:228 Resp: 797149
 Ion Ratio Lower Upper
 228 100
 226 30.3 21.7 32.5
 229 20.6 15.8 23.8





#83

Chrysene

Concen: 91.422 ng

RT: 13.969 min Scan# 20

Delta R.T. 0.006 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

V-2(0-2IN)





Tgt Ion:228 Resp: 639402

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 21.8 16.0 24.0



#84

Bis(2-ethylhexyl)phthalate

Concen: 32.850 ng

RT: 13.910 min Scan# 2010

Delta R.T. 0.000 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35




Tgt Ion:149 Resp: 123101


Ion Ratio Lower Upper

149 100

167 26.4 21.9 32.9

279 4.2 3.3 4.9





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.415 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_F

ClientSampleId :

V-2(0-2IN)





Tgt Ion:264 Resp: 94490

Ion Ratio Lower Upper

264 100

260 24.7 19.8 29.6

265 24.5 17.0 25.4



#87

Indeno(1,2,3-cd)pyrene

Concen: 57.291 ng

RT: 16.915 min Scan# 2521

Delta R.T. 0.012 min

Lab File: BF135743.D

Acq: 12 Oct 2023 19:35




Tgt Ion:276 Resp: 354938


Ion Ratio Lower Upper

276 100

138 19.5 19.2 28.8

277 24.2 19.7 29.5









#88
Benzo(b)fluoranthene
Concen: 217.355 ng
RT: 14.992 min Scan# 2194
Delta R.T. 0.012 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:252 Resp: 1111788
Ion Ratio Lower Upper
252 100
253 22.9 17.2 25.8
125 11.1 7.1 10.7#



#89
Benzo(k)fluoranthene
Concen: 220.034 ng
RT: 14.992 min Scan# 2194
Delta R.T. -0.012 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35

Tgt Ion:252 Resp: 1111788
Ion Ratio Lower Upper
252 100
253 22.9 17.4 26.0
125 11.1 7.2 10.8#









#90
Benzo(a)pyrene
Concen: 114.600 ng
RT: 15.363 min Scan# 21
Delta R.T. 0.018 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_F
ClientSampleId :
V-2(0-2IN)


Tgt Ion:252 Resp: 559009
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 10.3 9.2 13.8



#91
Dibenzo(a,h)anthracene
Concen: 12.183 ng
RT: 17.086 min Scan# 2550
Delta R.T. 0.177 min
Lab File: BF135743.D
Acq: 12 Oct 2023 19:35


Tgt Ion:278 Resp: 61760
Ion Ratio Lower Upper
278 100
139 17.9 13.4 20.0
279 28.4 19.0 28.6





#92
 Benzo(g,h,i)perylene
 Concen: 64.696 ng
 RT: 17.368 min Scan# 21
 Delta R.T. 0.018 min
 Lab File: BF135743.D
 Acq: 12 Oct 2023 19:35

Instrument :
 BNA_F
 ClientSampleId :
 V-2(0-2IN)



Tgt Ion:276 Resp: 342505
 Ion Ratio Lower Upper
 276 100
 277 24.0 18.8 28.2
 138 20.1 16.9 25.3

