

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101618\
 Data File : BF109910.D
 Acq On : 16 Oct 2018 17:27
 Operator : JU/SJ
 Sample : J5516-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSSI-1011-S-DH-30

Manual Integrations
 APPROVED

Sohil
 10/17/2018 11:15:35 AM

Quant Time: Oct 17 10:46:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 14:58:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.98	152	224155	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	791663	20.00	ng	0.00
39) Acenaphthene-d10	10.02	164	303143	20.00	ng	0.00
64) Phenanthrene-d10	11.52	188	484928	20.00	ng	0.00
76) Chrysene-d12	14.16	240	543590	20.00	ng	0.00
87) Perylene-d12	15.69	264	560607	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	1191120	84.23	ng	0.00
7) Phenol-d6	6.60	99	1445152	82.30	ng	0.00
23) Nitrobenzene-d5	7.54	82	853358	72.61	ng	-0.01
42) 2,4,6-Tribromophenol	10.81	330	223872	85.30	ng	0.00
45) 2-Fluorobiphenyl	9.33	172	1396144	73.49	ng	0.00
79) Terphenyl-d14	13.10	244	1263315	48.80	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.29	128	131558	3.624	ng	99
37) 2-Methylnaphthalene	8.97	142	955572	40.650	ng	98
38) 1-Methylnaphthalene	9.07	142	532793	23.390	ng	98
46) 1,1'-Biphenyl	9.44	154	370853	13.731	ng	98
49) Acenaphthylene	9.88	152	111421	3.855	ng	# 93
50) Dimethylphthalate	9.73	163	137715	6.416	ng	99
52) Acenaphthene	10.06	154	876024	47.609	ng	99
55) Dibenzofuran	10.23	168	1085530	41.367	ng	97
58) Fluorene	10.57	166	956524	50.073	ng	99
71) Phenanthrene	11.54	178	2871297	114.371	ng	95
72) Anthracene	11.59	178	562085	22.270	ng	99
73) Carbazole	11.74	167	115425	4.653	ng	99
75) Fluoranthene	12.73	202	2143153	85.454	ng	96
78) Pyrene	12.96	202	1793132	44.912	ng	99
81) Benzo(a)anthracene	14.15	228	777249	24.275	ng	95
83) Chrysene	14.19	228	704289m	23.109	ng	
86) Indeno(1,2,3-cd)pyrene	17.24	276	280138	11.416	ng	97
88) Benzo(b)fluoranthene	15.24	252	1261138m	39.020	ng	
89) Benzo(k)fluoranthene	15.26	252	288656m	9.061	ng	
90) Benzo(a)pyrene	15.63	252	695550	23.399	ng	99
91) Dibenzo(a,h)anthracene	17.26	278	68624	2.605	ng	# 91
92) Benzo(g,h,i)perylene	17.72	276	228386	8.581	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF101618\
Data File : BF109910.D
Acq On : 16 Oct 2018 17:27
Operator : JU/SJ
Sample : J5516-06
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM

Quant Time: Oct 17 10:46:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 14:58:24 2018
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.98 min Scan# 832
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.2	122.7	184.1
115	61.3	49.1	73.7

Manual Integrations
APPROVED




Sohil
10/17/2018 11:15:35 AM




#5

2-Fluorophenol
Concen: 84.231 ng
RT: 5.60 min Scan# 597
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
112	100		
64	62.3	44.1	66.1
63	32.4	23.7	35.5





#7

Phenol-d6

Concen: 82.300 ng

RT: 6.60 min Scan# 767

Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Instrument :

BNA_F

ClientSampleId :

SSSI-1011-S-DH-30

Tot Ion: 99 Resp: 1445152

Ion Ratio Lower Upper

99 100

42 18.1 15.8 23.6


71 31.5 28.0 42.0

Manual Integrations




APPROVED

Sohil

10/17/2018 11:15:35 AM



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.26 min Scan# 1050

Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Tgt Ion: 136 Resp: 791663

Ion Ratio Lower Upper




136 100


137 11.0 8.8 13.2

54 8.5 5.4 8.2#

68 6.5 4.2 6.2#

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
Nitrobenzene-d5
Concen: 72.614 ng
RT: 7.54 min Scan# 927
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
82	100		
128	45.8	34.2	51.2
54	54.3	38.6	58.0




Manual Integrations
APPROVED


Sohil
10/17/2018 11:15:35 AM



#31
Naphthalene
Concen: 3.624 ng
RT: 8.29 min Scan# 1054
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.4	8.8	13.2
127	13.0	10.8	16.2





#37

2-Methylnaphthalene

Concen: 40.650 ng

RT: 8.97 min Scan# 1171

Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Instrument :

BNA_F





ClientSampleId :

SSSI-1011-S-DH-30

Tot Ion:	142	Resp:	955572
Ion Ratio	Lower	Upper	
142	100		
141	88.4	71.4	107.2
115	33.1	28.7	43.1

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM



#38

1-Methylnaphthalene

Concen: 23.390 ng




RT: 9.07 min Scan# 1188


Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Tgt Ion:	142	Resp:	532793
Ion Ratio	Lower	Upper	
142	100		
141	90.6	74.2	111.4
116	3.6	2.6	4.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.02 min Scan# 1349
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
164	100		
162	103.0	83.0	124.6
160	46.4	37.0	55.6




Manual Integrations
APPROVED


Sohil
10/17/2018 11:15:35 AM



#42
2,4,6-Tribromophenol
Concen: 85.300 ng
RT: 10.81 min Scan# 1483
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.6	77.1	115.7
141	36.8	27.0	40.4






#45
2-Fluorobiphenyl
Concen: 73.489 ng
RT: 9.33 min Scan# 1232
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30




Tot Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.6	43.0
170	24.0	19.7	29.5

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM




Abundance Ion 172.00 (171.70 to 172.70): E
2000000
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 13.731 ng
RT: 9.44 min Scan# 1250
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
154	100		
153	39.1	20.4	60.4
76	12.1	0.0	31.9



Abundance Ion 154.00 (153.70 to 154.70): E
600000
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1





#49

Acenaphthylene

Concen: 3.855 ng

RT: 9.88 min Scan# 1325

Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Instrument :

BNA_F


ClientSampleId :

SSSI-1011-S-DH-30




Tot Ion	Ratio	Resp	Lower	Upper
152	100	111421		
151	20.9	15.8	23.8	
153	18.2	10.4	15.6	

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 6.416 ng


RT: 9.73 min Scan# 1299

Delta R.T. -0.01 min



Lab File: BF109910.D


Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	137715		
194	3.6	3.2	4.8	
164	9.8	8.1	12.1	



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#52

Acenaphthene

Concen: 47.609 ng

RT: 10.06 min Scan# 1355

Delta R.T. -0.01 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Instrument :

BNA_F


ClientSampleId :

SSSI-1011-S-DH-30




Tot Ion:	154	Resp:	876024
Ion Ratio	Lower	Upper	
154	100		
153	112.3	90.1	135.1
152	52.3	43.4	65.2

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#55

Dibenzofuran

Concen: 41.367 ng


RT: 10.23 min Scan# 1384

Delta R.T. -0.01 min



Lab File: BF109910.D


Acq: 16 Oct 2018 17:27

Tgt Ion:	168	Resp:	1085530
Ion Ratio	Lower	Upper	
168	100		
139	38.7	33.0	49.6
169	14.3	11.3	16.9



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#58
 Fluorene
 Concen: 50.073 ng
 RT: 10.57 min Scan# 1443
 Delta R.T. -0.00 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Instrument :
 BNA_F
 ClientSampleId :
 SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
166	100		
165	98.9	78.6	117.8
167	13.9	10.8	16.2




Manual Integrations
 APPROVED


Sohil
 10/17/2018 11:15:35 AM



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.52 min Scan# 1603
 Delta R.T. -0.00 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
188	100		
94	10.6	7.2	10.8
80	10.7	7.9	11.9






#71
Phenanthrene
Concen: 114.371 ng
RT: 11.54 min Scan# 1608
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
178	100		
176	21.8	15.5	23.3
179	17.3	12.5	18.7

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72
Anthracene
Concen: 22.270 ng
RT: 11.59 min Scan# 1616
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27




Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.4	23.2
179	15.6	12.6	18.8


Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 4.653 ng
 RT: 11.74 min Scan# 1641
 Delta R.T. -0.01 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Instrument :
 BNA_F
 ClientSampleId :
 SSSI-1011-S-DH-30

Tot Ion	Ratio	Resp	Lower	Upper
167	100			
166	21.9	17.4	26.0	
139	13.0	10.9	16.3	




Manual Integrations
 APPROVED


Sohil
 10/17/2018 11:15:35 AM



#75
 Fluoranthene
 Concen: 85.454 ng
 RT: 12.73 min Scan# 1810
 Delta R.T. -0.00 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Resp	Lower	Upper
202	100			
101	13.7	0.0	31.4	
203	18.6	0.0	37.7	









#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.16 min Scan# 2053
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
240	100		
120	11.4	8.3	12.5
236	25.6	21.2	31.8




Manual Integrations
APPROVED


Sohil
10/17/2018 11:15:35 AM



#78
Pyrene
Concen: 44.912 ng
RT: 12.96 min Scan# 1849
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.8	26.6
203	18.9	14.2	21.4









#79
 Terphenyl-d14
 Concen: 48.803 ng
 RT: 13.10 min Scan# 1872
 Delta R.T. -0.01 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Instrument :
 BNA_F
 ClientSampleId :
 SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.8	8.8
122	13.0	8.7	13.1




Manual Integrations
 APPROVED


Sohil
 10/17/2018 11:15:35 AM



#81
 Benzo(a)anthracene
 Concen: 24.275 ng
 RT: 14.15 min Scan# 2051
 Delta R.T. -0.00 min
 Lab File: BF109910.D
 Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.2	22.1	33.1
229	21.0	15.6	23.4









#83
Chrysene
Concen: 23.109 ng m
RT: 14.19 min Scan# 2057
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.7	37.1
229	24.1	15.9	23.9




Manual Integrations
APPROVED


Sohil
10/17/2018 11:15:35 AM



#86
Indeno(1,2,3-cd)pyrene
Concen: 11.416 ng
RT: 17.24 min Scan# 2577
Delta R.T. -0.02 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
276	100		
138	25.2	18.5	27.7
277	24.1	20.0	30.0






#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.69 min Scan# 2313
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.7	28.1
265	22.6	16.7	25.1

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88
Benzo(b)fluoranthene
Concen: 39.020 ng m
RT: 15.24 min Scan# 2236
Delta R.T. 0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27




Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.4	17.2	25.8
125	12.6	8.2	12.4


Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 9.061 ng m
RT: 15.26 min Scan# 2240
Delta R.T. -0.01 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Instrument :
BNA_F
ClientSampleId :
SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.1	17.2	25.8
125	12.4	7.4	11.0




Manual Integrations
APPROVED


Sohil
10/17/2018 11:15:35 AM



#90
Benzo(a)pyrene
Concen: 23.399 ng
RT: 15.63 min Scan# 2302
Delta R.T. -0.00 min
Lab File: BF109910.D
Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.1	18.2	27.4
125	12.4	9.0	13.6





#91

Dibenzo(a,h)anthracene

Concen: 2.605 ng

RT: 17.26 min Scan# 2579

Delta R.T. -0.02 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Instrument :

BNA_F





ClientSampleId :

SSSI-1011-S-DH-30

Tot Ion	Ratio	Lower	Upper
278	100		
139	21.1	12.7	19.1#
279	27.3	19.0	28.4

Manual Integrations
APPROVED

Sohil
10/17/2018 11:15:35 AM



#92

Benzo(g,h,i)perylene

Concen: 8.581 ng

RT: 17.72 min Scan# 2658

Delta R.T. -0.02 min

Lab File: BF109910.D

Acq: 16 Oct 2018 17:27

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.5	18.8	28.2
138	25.6	16.0	24.0#

