

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101723\
 Data File : BF135817.D
 Acq On : 17 Oct 2023 17:54
 Operator : CG\JU
 Sample : 04723-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-6A

Quant Time: Oct 18 01:59:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 17 01:22:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.739	152	85192	20.000	ng	0.00
21) Naphthalene-d8	8.022	136	312617	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	168626	20.000	ng	-0.01
64) Phenanthrene-d10	11.269	188	300159	20.000	ng	-0.01
76) Chrysene-d12	13.939	240	175672	20.000	ng	-0.01
86) Perylene-d12	15.410	264	183034	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.398	112	130904m	24.704	ng	0.02
7) Phenol-d6	6.404	99	216058	32.679	ng	0.00
23) Nitrobenzene-d5	7.310	82	553273	97.338	ng	0.00
42) 2,4,6-Tribromophenol	10.575	330	229652	141.927	ng	-0.01
45) 2-Fluorobiphenyl	9.098	172	1014734	94.087	ng	0.00
79) Terphenyl-d14	12.868	244	990439	104.411	ng	-0.01
Target Compounds						
17) 2-Methylphenol	7.016	107	66681	13.480	ng	Qvalue 93
20) 3+4-Methylphenols	7.169	107	112876	18.304	ng	# 82
27) 2,4-Dimethylphenol	7.692	122	34168	7.455	ng	98
31) Naphthalene	8.051	128	3049614	199.999	ng	97
37) 2-Methylnaphthalene	8.733	142	365870	36.419	ng	100
38) 1-Methylnaphthalene	8.828	142	199105	21.488	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101723\
Data File : BF135817.D
Acq On : 17 Oct 2023 17:54
Operator : CG\JU
Sample : 04723-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-6A

Quant Time: Oct 18 01:59:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 17 01:22:00 2023
Response via : Initial Calibration

