

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101823\
 Data File : BF135868.D
 Acq On : 19 Oct 2023 01:59
 Operator : CG\JU
 Sample : 04767-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 Q-3(10-15)

Quant Time: Oct 19 04:43:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 17 01:22:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.740	152	80403	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	290758	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	133483	20.000	ng	-0.01
64) Phenanthrene-d10	11.275	188	204470	20.000	ng	0.00
76) Chrysene-d12	13.945	240	157825	20.000	ng	0.00
86) Perylene-d12	15.422	264	147773	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	461724	92.327	ng	0.00
7) Phenol-d6	6.393	99	594223	95.231	ng	0.00
23) Nitrobenzene-d5	7.305	82	371405	70.254	ng	-0.01
42) 2,4,6-Tribromophenol	10.575	330	116008	90.569	ng	-0.01
45) 2-Fluorobiphenyl	9.093	172	639281	74.880	ng	-0.01
79) Terphenyl-d14	12.869	244	528205	61.979	ng	-0.01
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.305	70	49368	12.796	ng	# 77
25) Isophorone	7.305	82	371402	37.239	ng	# 66
31) Naphthalene	8.040	128	171507	12.093	ng	99
33) 4-Chloroaniline	8.040	127	22018	3.554	ng	# 32
37) 2-Methylnaphthalene	8.734	142	36065	3.860	ng	94
38) 1-Methylnaphthalene	8.828	142	58454	6.783	ng	97
49) Acenaphthylene	9.640	152	231809	19.008	ng	99
52) Acenaphthene	9.804	154	55352	7.573	ng	100
55) Dibenzofuran	9.981	168	139446	12.644	ng	96
56) 4-Nitrophenol	9.981	139	49982	48.263	ng	# 8
58) Fluorene	10.328	166	100938	11.690	ng	98
67) 4-Bromophenyl-phenylether	10.575	248	6753	3.290	ng	# 1
71) Phenanthrene	11.304	178	1239612	126.555	ng	99
72) Anthracene	11.351	178	496739	48.940	ng	100
73) Carbazole	11.516	167	62256	6.577	ng	99
75) Fluoranthene	12.504	202	1443817	132.810	ng	99
78) Pyrene	12.734	202	1359202	108.375	ng	99
81) Benzo(a)anthracene	13.934	228	864019	83.347	ng	96
83) Chrysene	13.975	228	669631	66.691	ng	97
87) Indeno(1,2,3-cd)pyrene	16.927	276	196200	24.367	ng	# 93
88) Benzo(b)fluoranthene	14.998	252	894950	108.212	ng	99
89) Benzo(k)fluoranthene	14.998	252	894950	111.461	ng	98
90) Benzo(a)pyrene	15.363	252	491476	63.197	ng	98
91) Dibenzo(a,h)anthracene	16.927	278	56094	8.498	ng	# 87
92) Benzo(g,h,i)perylene	17.386	276	174863	25.790	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101823\
Data File : BF135868.D
Acq On : 19 Oct 2023 01:59
Operator : CG\JU
Sample : 04767-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)

Quant Time: Oct 19 04:43:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 17 01:22:00 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.740 min Scan# 791
 Delta R.T. -0.006 min
 Lab File: BF135868.D
 Acq: 19 Oct 2023 01:59




Instrument :
 BNA_F
 ClientSampleId :
 Q-3(10-15)


Tgt Ion:152 Resp: 80403
 Ion Ratio Lower Upper
 152 100
 150 154.9 124.5 186.7
 115 56.0 49.0 73.4



#5
 2-Fluorophenol
 Concen: 92.327 ng
 RT: 5.375 min Scan# 559
 Delta R.T. 0.000 min
 Lab File: BF135868.D
 Acq: 19 Oct 2023 01:59

Tgt Ion:112 Resp: 461724
 Ion Ratio Lower Upper
 112 100
 64 67.9 51.2 76.8
 63 35.6 26.1 39.1









#7
Phenol-d6
Concen: 95.231 ng
RT: 6.393 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion: 99 Resp: 594223
Ion Ratio Lower Upper
99 100
42 23.3 18.4 27.6
71 34.1 27.2 40.8



#19
n-Nitroso-di-n-propylamine
Concen: 12.796 ng
RT: 7.305 min Scan# 887
Delta R.T. 0.141 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion: 70 Resp: 49368
Ion Ratio Lower Upper
70 100
42 59.1 59.8 89.6#
101 0.0 7.4 11.2#
130 1.8 16.6 24.8#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:136 Resp: 290758
Ion Ratio Lower Upper
136 100
137 10.6 8.6 12.8
54 9.8 8.5 12.7
68 5.5 4.8 7.2



#23
Nitrobenzene-d5
Concen: 70.254 ng
RT: 7.305 min Scan# 887
Delta R.T. -0.012 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion: 82 Resp: 371405
Ion Ratio Lower Upper
82 100
128 43.2 37.0 55.6
54 64.4 52.8 79.2









#25
Isophorone
Concen: 37.239 ng
RT: 7.305 min Scan# 81
Delta R.T. -0.264 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion: 82 Resp: 371402
Ion Ratio Lower Upper
82 100
95 0.0 5.5 8.3#
138 0.0 13.9 20.9#



#31
Naphthalene
Concen: 12.093 ng
RT: 8.040 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:128 Resp: 171507
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.1 10.2 15.4





#33

4-Chloroaniline

Concen: 3.554 ng

RT: 8.040 min Scan# 1130

Delta R.T. -0.064 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :

Q-3(10-15)

Tgt Ion:127 Resp: 22018





Ion Ratio Lower Upper

127 100

129 84.9 26.5 39.7#

65 0.0 27.9 41.9#

92 0.0 15.8 23.8#



#37

2-Methylnaphthalene

Concen: 3.860 ng

RT: 8.734 min Scan# 1130

Delta R.T. -0.006 min

Lab File: BF135868.D




Acq: 19 Oct 2023 01:59


Tgt Ion:142 Resp: 36065

Ion Ratio Lower Upper

142 100

141 93.3 70.6 105.8





#38

1-Methylnaphthalene

Concen: 6.783 ng

RT: 8.828 min Scan# 1146

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :





Q-3(10-15)

Tgt Ion:142 Resp: 58454

Ion Ratio Lower Upper

142 100

141 93.9 72.7 109.1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.775 min Scan# 1307

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59




Tgt Ion:164 Resp: 133483


Ion Ratio Lower Upper

164 100

162 100.5 80.8 121.2

160 45.7 36.6 54.8





#42

2,4,6-Tribromophenol

Concen: 90.569 ng

RT: 10.575 min Scan# 1443

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :

Q-3(10-15)





Tgt Ion:330 Resp: 116008

Ion Ratio Lower Upper

330 100

332 95.6 75.0 112.4

141 32.1 25.8 38.6



#45

2-Fluorobiphenyl

Concen: 74.880 ng

RT: 9.093 min Scan# 1191

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59




Tgt Ion:172 Resp: 639281


Ion Ratio Lower Upper

172 100

171 34.4 28.4 42.6

170 23.4 18.8 28.2









#49
Acenaphthylene
Concen: 19.008 ng
RT: 9.640 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:152 Resp: 231809
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 13.6 10.2 15.4



#52
Acenaphthene
Concen: 7.573 ng
RT: 9.804 min Scan# 1312
Delta R.T. -0.012 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:154 Resp: 55352
Ion Ratio Lower Upper
154 100
153 112.6 90.1 135.1
152 53.2 43.1 64.7









#55
Dibenzenofuran
Concen: 12.644 ng
RT: 9.981 min Scan# 1342
Delta R.T. -0.012 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:168 Resp: 139446
Ion Ratio Lower Upper
168 100
139 36.8 31.0 46.4
169 15.5 10.7 16.1



#56
4-Nitrophenol
Concen: 48.263 ng
RT: 9.981 min Scan# 1342
Delta R.T. 0.036 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:139 Resp: 49982
Ion Ratio Lower Upper
139 100
109 2.5 45.6 85.6#
65 2.8 87.5 127.5#









#58
Fluorene
Concen: 11.690 ng
RT: 10.328 min Scan# 1402
Delta R.T. -0.012 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:166 Resp: 100938
Ion Ratio Lower Upper
166 100
165 89.9 73.4 110.0
167 13.6 11.0 16.6



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.275 min Scan# 1562
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:188 Resp: 204470
Ion Ratio Lower Upper
188 100
94 8.7 8.0 12.0
80 8.4 8.3 12.5









#67
4-Bromophenyl-phenylether
Concen: 3.290 ng
RT: 10.575 min Scan# 1484
Delta R.T. -0.247 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:248 Resp: 6753
Ion Ratio Lower Upper
248 100
250 198.4 78.0 117.0#
141 449.7 59.0 88.6#



#71
Phenanthrene
Concen: 126.555 ng
RT: 11.304 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:178 Resp: 1239612
Ion Ratio Lower Upper
178 100
176 20.1 16.0 24.0
179 15.7 12.2 18.4









#72
 Anthracene
 Concen: 48.940 ng
 RT: 11.351 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BF135868.D
 Acq: 19 Oct 2023 01:59




Instrument :
 BNA_F
 ClientSampleId :
 Q-3(10-15)


Tgt Ion:178 Resp: 496739
 Ion Ratio Lower Upper
 178 100
 176 18.6 15.1 22.7
 179 15.3 12.3 18.5



#73
 Carbazole
 Concen: 6.577 ng
 RT: 11.516 min Scan# 1603
 Delta R.T. -0.006 min
 Lab File: BF135868.D
 Acq: 19 Oct 2023 01:59

Tgt Ion:167 Resp: 62256
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.1 25.7
 139 12.7 10.2 15.4









#75
Fluoranthene
Concen: 132.810 ng
RT: 12.504 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:202 Resp: 1443817
Ion Ratio Lower Upper
202 100
101 11.0 0.0 31.5
203 18.2 0.0 38.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.945 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:240 Resp: 157825
Ion Ratio Lower Upper
240 100
120 10.0 8.2 12.2
236 25.9 21.2 31.8





#78

Pyrene

Concen: 108.375 ng

RT: 12.734 min Scan# 1810

Delta R.T. 0.000 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :

Q-3(10-15)





Tgt Ion:202 Resp: 1359202

Ion Ratio Lower Upper

202 100

200 20.6 16.1 24.1

203 18.4 14.2 21.2



#79

Terphenyl-d14

Concen: 61.979 ng

RT: 12.869 min Scan# 1833

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59




Tgt Ion:244 Resp: 528205


Ion Ratio Lower Upper

244 100

212 6.6 5.6 8.4

122 10.7 9.0 13.6









#81
Benzo(a)anthracene
Concen: 83.347 ng
RT: 13.934 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:228 Resp: 864019
Ion Ratio Lower Upper
228 100
226 30.8 22.2 33.2
229 20.6 16.1 24.1



#83
Chrysene
Concen: 66.691 ng
RT: 13.975 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:228 Resp: 669631
Ion Ratio Lower Upper
228 100
226 31.2 24.4 36.6
229 21.9 15.7 23.5





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.422 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :

Q-3(10-15)





Tgt Ion:264 Resp: 147773

Ion Ratio Lower Upper

264 100

260 23.5 19.2 28.8

265 24.8 17.4 26.2



#87

Indeno(1,2,3-cd)pyrene

Concen: 24.367 ng

RT: 16.927 min Scan# 2523

Delta R.T. -0.006 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59




Tgt Ion:276 Resp: 196200


Ion Ratio Lower Upper

276 100

138 19.3 19.8 29.6#

277 23.5 20.2 30.4









#88
Benzo(b)fluoranthene
Concen: 108.212 ng
RT: 14.998 min Scan# 2195
Delta R.T. 0.006 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59




Instrument :
BNA_F
ClientSampleId :
Q-3(10-15)


Tgt Ion:252 Resp: 894950
Ion Ratio Lower Upper
252 100
253 22.6 18.2 27.2
125 10.7 8.2 12.2



#89
Benzo(k)fluoranthene
Concen: 111.461 ng
RT: 14.998 min Scan# 2195
Delta R.T. -0.023 min
Lab File: BF135868.D
Acq: 19 Oct 2023 01:59

Tgt Ion:252 Resp: 894950
Ion Ratio Lower Upper
252 100
253 22.6 17.4 26.0
125 10.7 8.3 12.5





#90

Benzo(a)pyrene

Concen: 63.197 ng

RT: 15.363 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59

Instrument :

BNA_F

ClientSampleId :

Q-3(10-15)





Tgt Ion:252 Resp: 491476

Ion Ratio Lower Upper

252 100

253 22.9 17.4 26.2

125 11.3 8.8 13.2



#91

Dibenzo(a,h)anthracene

Concen: 8.498 ng

RT: 16.927 min Scan# 2523

Delta R.T. -0.012 min

Lab File: BF135868.D

Acq: 19 Oct 2023 01:59




Tgt Ion:278 Resp: 56094


Ion Ratio Lower Upper

278 100

139 19.3 12.5 18.7#


279 30.8 18.6 27.8#





#92
 Benzo(g,h,i)perylene
 Concen: 25.790 ng
 RT: 17.386 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: BF135868.D
 Acq: 19 Oct 2023 01:59

Instrument :
 BNA_F
 ClientSampleId :
 Q-3(10-15)



Tgt Ion:276 Resp: 174863
 Ion Ratio Lower Upper
 276 100
 277 25.7 19.4 29.2
 138 20.6 17.3 25.9

