

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101823\
 Data File : BF135850.D
 Acq On : 18 Oct 2023 16:38
 Operator : CG\JU
 Sample : 04767-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FB-2

Quant Time: Oct 19 01:20:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 17 01:22:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.739	152	82140	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	318519	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	171849	20.000	ng	-0.01
64) Phenanthrene-d10	11.274	188	347295	20.000	ng	0.00
76) Chrysene-d12	13.939	240	168892	20.000	ng	-0.01
86) Perylene-d12	15.415	264	173658	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	257581	50.417	ng	0.00
7) Phenol-d6	6.398	99	191289	30.008	ng	0.00
23) Nitrobenzene-d5	7.310	82	454979	78.562	ng	0.00
42) 2,4,6-Tribromophenol	10.574	330	219597	133.167	ng	-0.01
45) 2-Fluorobiphenyl	9.098	172	858464	78.104	ng	0.00
79) Terphenyl-d14	12.868	244	990713	108.632	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101823\
Data File : BF135850.D
Acq On : 18 Oct 2023 16:38
Operator : CG\JU
Sample : 04767-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FB-2

Quant Time: Oct 19 01:20:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 17 01:22:00 2023
Response via : Initial Calibration

