

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF101918\
 Data File : BF110029.D
 Acq On : 20 Oct 2018 4:14
 Operator : JU/SJ
 Sample : J5593-14
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 101718-JETT-E-D-S

Quant Time: Oct 20 07:02:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 20 01:53:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	200685	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	788051	20.00	ng	0.00
39) Acenaphthene-d10	10.01	164	335377	20.00	ng	0.00
64) Phenanthrene-d10	11.50	188	520164	20.00	ng	0.00
76) Chrysene-d12	14.14	240	373375	20.00	ng	0.00
87) Perylene-d12	15.67	264	311946	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	450807	35.61	ng	0.00
7) Phenol-d6	6.58	99	332038	21.12	ng	-0.01
23) Nitrobenzene-d5	7.53	82	1178311	100.72	ng	0.00
42) 2,4,6-Tribromophenol	10.80	330	227867	78.48	ng	0.00
45) 2-Fluorobiphenyl	9.33	172	1954884	93.01	ng	0.00
79) Terphenyl-d14	13.09	244	1324810	74.51	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.72	163	49047	2.065	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF101918\
Data File : BF110029.D
Acq On : 20 Oct 2018 4:14
Operator : JU/SJ
Sample : J5593-14
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
101718-JETT-E-D-S

Quant Time: Oct 20 07:02:03 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 20 01:53:35 2018
Response via : Initial Calibration

