

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101923\
 Data File : BF135886.D
 Acq On : 19 Oct 2023 15:04
 Operator : CG\JU
 Sample : 04788-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PMW-2

Quant Time: Oct 20 01:45:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 17 01:22:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.734	152	71380	20.000	ng	-0.01
21) Naphthalene-d8	8.016	136	294010	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	115134	20.000	ng	-0.01
64) Phenanthrene-d10	11.269	188	204644	20.000	ng	-0.01
76) Chrysene-d12	13.933	240	138387	20.000	ng	-0.02
86) Perylene-d12	15.415	264	171419	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	241460	54.386	ng	0.00
7) Phenol-d6	6.392	99	198727	35.874	ng	0.00
23) Nitrobenzene-d5	7.304	82	426274	79.741	ng	-0.01
42) 2,4,6-Tribromophenol	10.575	330	158971	143.891	ng	-0.01
45) 2-Fluorobiphenyl	9.092	172	794291	107.864	ng	-0.01
79) Terphenyl-d14	12.868	244	596571	79.834	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101923\
Data File : BF135886.D
Acq On : 19 Oct 2023 15:04
Operator : CG\JU
Sample : 04788-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PMW-2

Quant Time: Oct 20 01:45:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 17 01:22:00 2023
Response via : Initial Calibration

