

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101923\
 Data File : BF135887.D
 Acq On : 19 Oct 2023 15:35
 Operator : CG\JU
 Sample : 04788-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PMW-1

Quant Time: Oct 20 01:46:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 17 01:22:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.734	152	62014	20.000	ng	-0.01
21) Naphthalene-d8	8.016	136	231453	20.000	ng	# 0.00
39) Acenaphthene-d10	9.775	164	140966	20.000	ng	-0.01
64) Phenanthrene-d10	11.269	188	271061	20.000	ng	-0.01
76) Chrysene-d12	13.939	240	135352	20.000	ng	-0.01
86) Perylene-d12	15.415	264	148268	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	245180	63.564	ng	0.00
7) Phenol-d6	6.393	99	200677	41.697	ng	0.00
23) Nitrobenzene-d5	7.304	82	413773	98.323	ng	-0.01
42) 2,4,6-Tribromophenol	10.575	330	206264	152.485	ng	-0.01
45) 2-Fluorobiphenyl	9.098	172	822309	91.206	ng	0.00
79) Terphenyl-d14	12.869	244	717404	98.157	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101923\
Data File : BF135887.D
Acq On : 19 Oct 2023 15:35
Operator : CG\JU
Sample : 04788-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PMW-1

Quant Time: Oct 20 01:46:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 17 01:22:00 2023
Response via : Initial Calibration

