

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102218\  
 Data File : BF110043.D  
 Acq On : 22 Oct 2018 12:25  
 Operator : JU/SJ  
 Sample : PB114001BL  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :

Quant Time: Oct 23 02:29:07 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Oct 16 14:58:24 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.97  | 152  | 206031   | 20.00  | ng    | -0.01    |
| 21) Naphthalene-d8          | 8.25  | 136  | 797168   | 20.00  | ng    | -0.02    |
| 39) Acenaphthene-d10        | 10.01 | 164  | 379076   | 20.00  | ng    | -0.02    |
| 64) Phenanthrene-d10        | 11.50 | 188  | 745479   | 20.00  | ng    | -0.01    |
| 76) Chrysene-d12            | 14.14 | 240  | 698665   | 20.00  | ng    | -0.02    |
| 87) Perylene-d12            | 15.67 | 264  | 594223   | 20.00  | ng    | -0.02    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.60  | 112  | 1498949  | 115.32 | ng    | 0.00     |
| 7) Phenol-d6                | 6.59  | 99   | 1897924  | 117.59 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.53  | 82   | 1176898  | 99.45  | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 10.80 | 330  | 477120   | 145.38 | ng    | -0.01    |
| 45) 2-Fluorobiphenyl        | 9.33  | 172  | 2087655  | 87.88  | ng    | -0.01    |
| 79) Terphenyl-d14           | 13.09 | 244  | 2369231  | 71.21  | ng    | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102218\  
Data File : BF110043.D  
Acq On : 22 Oct 2018 12:25  
Operator : JU/SJ  
Sample : PB114001BL  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :

Quant Time: Oct 23 02:29:07 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 16 14:58:24 2018  
Response via : Initial Calibration

