

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102218\  
 Data File : BF110048.D  
 Acq On : 22 Oct 2018 14:57  
 Operator : JU/SJ  
 Sample : J5590-01  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :

Quant Time: Oct 23 02:56:07 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Oct 16 14:58:24 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)    |
|-----------------------------|-------|------|----------|-------|-------|-------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.97  | 152  | 232757   | 20.00 | ng    | -0.01       |
| 21) Naphthalene-d8          | 8.25  | 136  | 927093   | 20.00 | ng    | -0.02       |
| 39) Acenaphthene-d10        | 10.01 | 164  | 409540   | 20.00 | ng    | -0.02       |
| 64) Phenanthrene-d10        | 11.50 | 188  | 650800   | 20.00 | ng    | -0.01       |
| 76) Chrysene-d12            | 14.15 | 240  | 460677   | 20.00 | ng    | -0.02       |
| 87) Perylene-d12            | 15.67 | 264  | 377497   | 20.00 | ng    | -0.02       |
| System Monitoring Compounds |       |      |          |       |       |             |
| 5) 2-Fluorophenol           | 5.59  | 112  | 1217163  | 82.89 | ng    | 0.00        |
| 7) Phenol-d6                | 6.59  | 99   | 1577263  | 86.50 | ng    | -0.01       |
| 23) Nitrobenzene-d5         | 7.53  | 82   | 961936   | 69.90 | ng    | -0.02       |
| 42) 2,4,6-Tribromophenol    | 10.80 | 330  | 309158   | 87.19 | ng    | -0.01       |
| 45) 2-Fluorobiphenyl        | 9.33  | 172  | 1586345  | 61.81 | ng    | -0.01       |
| 79) Terphenyl-d14           | 13.09 | 244  | 1187581  | 54.13 | ng    | -0.02       |
| Target Compounds            |       |      |          |       |       |             |
| 10) Phenol                  | 6.60  | 94   | 136349   | 6.926 | ng    | Qvalue # 63 |
| 50) Dimethylphthalate       | 9.72  | 163  | 256444   | 8.843 | ng    | 99          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102218\  
Data File : BF110048.D  
Acq On : 22 Oct 2018 14:57  
Operator : JU/SJ  
Sample : J5590-01  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :

Quant Time: Oct 23 02:56:07 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101518.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 16 14:58:24 2018  
Response via : Initial Calibration

