


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF102318\
 Data File : BF110079.D
 Acq On : 23 Oct 2018 16:03
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 23 15:06:18 2018



AutoFind: Scans 1640, 1641, 1642; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	52231	PASS
68	69	0.00	2	1.6	753	PASS
69	198	0.00	100	34.7	48092	PASS
70	69	0.00	2	0.6	266	PASS
127	198	10	80	47.4	65808	PASS
197	198	0.00	2	0.5	732	PASS
198	198	100	100	100.0	138728	PASS
199	198	5	9	6.6	9141	PASS
275	198	10	60	25.9	35984	PASS
365	198	1	100	3.1	4324	PASS
441	198	0.01	100	15.6	21686	PASS
442	198	50	100	97.1	134717	PASS
443	442	15	24	19.4	26168	PASS