

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102319\
 Data File : BF117492.D
 Acq On : 23 Oct 2019 19:24
 Operator : JU
 Sample : K5139-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BU-03-102119

Quant Time: Oct 24 08:52:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 18:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	42974	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	175881	20.00	ng	0.00
39) Acenaphthene-d10	9.76	164	93139	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	148447	20.00	ng	0.00
76) Chrysene-d12	13.89	240	88681	20.00	ng	-0.01
87) Perylene-d12	15.32	264	96073	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	283970	98.31	ng	0.00
7) Phenol-d6	6.38	99	391154	100.82	ng	0.00
23) Nitrobenzene-d5	7.30	82	236903	66.50	ng	0.00
42) 2,4,6-Tribromophenol	10.56	330	76872	95.39	ng	0.00
45) 2-Fluorobiphenyl	9.09	172	391953	59.29	ng	0.00
79) Terphenyl-d14	12.83	244	354342	65.17	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.48	163	49245	7.417	ng	Qvalue # 99
75) Fluoranthene	12.46	202	19394	2.248	ng	97
78) Pyrene	12.69	202	19813	2.601	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102319\
Data File : BF117492.D
Acq On : 23 Oct 2019 19:24
Operator : JU
Sample : K5139-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BU-03-102119

Quant Time: Oct 24 08:52:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration

