


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF102319\
 Data File : BF117482.D
 Acq On : 23 Oct 2019 14:44
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101819.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 18 18:53:31 2019



AutoFind: Scans 1600, 1601, 1602; Background Corrected with Scan 1596

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	9491	PASS
68	69	0.00	2	1.4	170	PASS
69	198	0.00	100	47.3	12213	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.9	13904	PASS
197	198	0.00	2	0.5	133	PASS
198	198	100	100	100.0	25803	PASS
199	198	5	9	6.6	1693	PASS
275	198	10	60	22.4	5788	PASS
365	198	1	100	2.6	672	PASS
441	198	0.01	100	10.4	2681	PASS
442	198	50	100	61.2	15799	PASS
443	442	15	24	19.7	3116	PASS