

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102418\  
 Data File : BF110143.D  
 Acq On : 25 Oct 2018 5:04  
 Operator : JU/SJ  
 Sample : J5640-11  
 Misc :  
 ALS Vial : 40 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 102218-DBRI-F-D-B

Quant Time: Oct 25 09:30:20 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Oct 25 02:28:50 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)    |
|-----------------------------|-------|------|----------|--------|-------|-------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.97  | 152  | 127410   | 20.00  | ng    | 0.00        |
| 21) Naphthalene-d8          | 8.25  | 136  | 439465   | 20.00  | ng    | 0.00        |
| 39) Acenaphthene-d10        | 10.01 | 164  | 164630   | 20.00  | ng    | 0.00        |
| 64) Phenanthrene-d10        | 11.50 | 188  | 271918   | 20.00  | ng    | 0.00        |
| 76) Chrysene-d12            | 14.14 | 240  | 245791   | 20.00  | ng    | 0.00        |
| 87) Perylene-d12            | 15.67 | 264  | 167868   | 20.00  | ng    | 0.00        |
| System Monitoring Compounds |       |      |          |        |       |             |
| 5) 2-Fluorophenol           | 5.59  | 112  | 339966   | 43.45  | ng    | 0.00        |
| 7) Phenol-d6                | 6.58  | 99   | 291876   | 29.17  | ng    | 0.00        |
| 23) Nitrobenzene-d5         | 7.53  | 82   | 756887   | 102.15 | ng    | 0.00        |
| 42) 2,4,6-Tribromophenol    | 10.80 | 330  | 133086   | 81.61  | ng    | 0.00        |
| 45) 2-Fluorobiphenyl        | 9.33  | 172  | 1246831  | 118.92 | ng    | 0.00        |
| 79) Terphenyl-d14           | 13.09 | 244  | 1033355  | 87.44  | ng    | 0.00        |
| Target Compounds            |       |      |          |        |       |             |
| 10) Phenol                  | 6.59  | 94   | 46800    | 4.375  | ng    | Qvalue # 62 |
| 50) Dimethylphthalate       | 9.72  | 163  | 37296    | 3.069  | ng    | 99          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102418\  
Data File : BF110143.D  
Acq On : 25 Oct 2018 5:04  
Operator : JU/SJ  
Sample : J5640-11  
Misc :  
ALS Vial : 40 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
102218-DBRI-F-D-B

Quant Time: Oct 25 09:30:20 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Oct 25 02:28:50 2018  
Response via : Initial Calibration

