


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF102418\
 Data File : BF110102.D
 Acq On : 24 Oct 2018 9:22
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 23 15:06:18 2018



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1634

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.7	45101	PASS
68	69	0.00	2	1.7	708	PASS
69	198	0.00	100	36.7	40650	PASS
70	69	0.00	2	0.6	257	PASS
127	198	10	80	50.6	56144	PASS
197	198	0.00	2	0.9	984	PASS
198	198	100	100	100.0	110912	PASS
199	198	5	9	6.8	7564	PASS
275	198	10	60	25.2	27961	PASS
365	198	1	100	2.8	3115	PASS
441	198	0.01	100	13.7	15141	PASS
442	198	50	100	87.1	96600	PASS
443	442	15	24	19.6	18922	PASS