

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102523\
 Data File : BF135944.D
 Acq On : 25 Oct 2023 16:49
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102523

Quant Time: Oct 26 03:58:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.776	0.6	100	0.00
3	Pyridine	40.000	40.156	-0.4	101	0.00
4	n-Nitrosodimethylamine	40.000	40.036	-0.1	101	0.00
5 S	2-Fluorophenol	80.000	80.256	-0.3	102	0.00
6	Aniline	40.000	39.446	1.4	100	0.00
7 S	Phenol-d6	80.000	79.836	0.2	101	0.00
8	2-Chlorophenol	40.000	40.016	-0.0	101	0.00
9	Benzaldehyde	40.000	40.222	-0.6	102	0.00
10 C	Phenol	40.000	39.816	0.5	102	0.00
11	bis(2-Chloroethyl)ether	40.000	39.266	1.8	100	0.00
12	1,3-Dichlorobenzene	40.000	39.281	1.8	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.280	1.8	100	0.00
14	1,2-Dichlorobenzene	40.000	39.482	1.3	101	0.00
15	Benzyl Alcohol	40.000	40.546	-1.4	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.933	2.7	100	0.00
17	2-Methylphenol	40.000	39.511	1.2	100	0.00
18	Hexachloroethane	40.000	40.420	-1.1	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.517	1.2	100	0.00
20	3+4-Methylphenols	40.000	40.048	-0.1	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	40.304	-0.8	100	0.00
23 S	Nitrobenzene-d5	80.000	86.425	-8.0	110	0.00
24	Nitrobenzene	40.000	42.398	-6.0	105	0.00
25	Isophorone	40.000	40.249	-0.6	99	0.00
26 C	2-Nitrophenol	40.000	47.175	-17.9	128	0.00
27	2,4-Dimethylphenol	40.000	41.080	-2.7	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.629	-1.6	100	0.00
29 C	2,4-Dichlorophenol	40.000	43.092	-7.7	101	0.00
30	1,2,4-Trichlorobenzene	40.000	40.890	-2.2	101	0.00
31	Naphthalene	40.000	40.787	-2.0	100	0.00
32	Benzoic acid	40.000	47.600	-19.0	134	0.00
33	4-Chloroaniline	40.000	39.986	0.0	97	0.00
34 C	Hexachlorobutadiene	40.000	40.983	-2.5	102	0.00
35	Caprolactam	40.000	42.877	-7.2	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.006	-5.0	99	0.00
37	2-Methylnaphthalene	40.000	40.255	-0.6	98	0.00
38	1-Methylnaphthalene	40.000	40.634	-1.6	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.048	-2.6	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.505	-13.8	103	0.00
42 S	2,4,6-Tribromophenol	80.000	91.864	-14.8	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.859	-12.1	103	0.00
44	2,4,5-Trichlorophenol	40.000	44.277	-10.7	101	0.00
45 S	2-Fluorobiphenyl	80.000	78.709	1.6	97	0.00
46	1,1'-Biphenyl	40.000	40.981	-2.5	99	0.00
47	2-Chloronaphthalene	40.000	40.904	-2.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102523\
 Data File : BF135944.D
 Acq On : 25 Oct 2023 16:49
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102523

Quant Time: Oct 26 03:58:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.531	-13.8	113	0.00
49	Acenaphthylene	40.000	41.052	-2.6	98	0.00
50	Dimethylphthalate	40.000	40.791	-2.0	98	0.00
51	2,6-Dinitrotoluene	40.000	42.533	-6.3	104	0.00
52 C	Acenaphthene	40.000	41.420	-3.6	99	0.00
53	3-Nitroaniline	40.000	42.699	-6.7	106	0.00
54 P	2,4-Dinitrophenol	40.000	47.214	-18.0	132	0.00
55	Dibenzofuran	40.000	40.459	-1.1	96	0.00
56 P	4-Nitrophenol	40.000	43.018	-7.5	107	0.00
57	2,4-Dinitrotoluene	40.000	43.871	-9.7	112	0.00
58	Fluorene	40.000	40.219	-0.5	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	45.364	-13.4	102	0.00
60	Diethylphthalate	40.000	40.737	-1.8	96	0.00
61	4-Chlorophenyl-phenylether	40.000	40.344	-0.9	98	0.00
62	4-Nitroaniline	40.000	42.172	-5.4	105	0.00
63	Azobenzene	40.000	40.507	-1.3	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.111	-15.3	120	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.669	-4.2	98	0.00
67	4-Bromophenyl-phenylether	40.000	41.570	-3.9	96	0.00
68	Hexachlorobenzene	40.000	41.455	-3.6	97	0.00
69	Atrazine	40.000	40.626	-1.6	95	0.00
70 C	Pentachlorophenol	40.000	43.464	-8.7	104	0.00
71	Phenanthrene	40.000	41.083	-2.7	95	0.00
72	Anthracene	40.000	42.002	-5.0	97	0.00
73	Carbazole	40.000	41.147	-2.9	95	0.00
74	Di-n-butylphthalate	40.000	42.121	-5.3	95	0.00
75 C	Fluoranthene	40.000	41.779	-4.4	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	41.915	-4.8	85	0.00
78	Pyrene	40.000	40.189	-0.5	97	0.00
79 S	Terphenyl-d14	80.000	80.161	-0.2	97	0.00
80	Butylbenzylphthalate	40.000	40.821	-2.1	103	0.00
81	Benzo(a)anthracene	40.000	41.747	-4.4	102	0.00
82	3,3'-Dichlorobenzidine	40.000	42.234	-5.6	102	0.00
83	Chrysene	40.000	41.518	-3.8	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.717	-11.8	104	0.00
85 c	Di-n-octyl phthalate	40.000	46.178	-15.4	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.989	-10.0	99	0.00
88	Benzo(b)fluoranthene	40.000	41.539	-3.8	104	0.00
89	Benzo(k)fluoranthene	40.000	40.581	-1.5	101	0.00
90 C	Benzo(a)pyrene	40.000	41.264	-3.2	101	0.00
91	Dibenzo(a,h)anthracene	40.000	42.494	-6.2	99	0.00
92	Benzo(g,h,i)perylene	40.000	39.354	1.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102523\
Data File : BF135944.D
Acq On : 25 Oct 2023 16:49
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102523

Quant Time: Oct 26 03:58:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 26 03:58:01 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0