

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102618\
 Data File : BF110225.D
 Acq On : 26 Oct 2018 22:19
 Operator : JU/SJ
 Sample : J5205-05 5X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR-05-102518

Manual Integrations
 APPROVED

Sohil
 10/29/2018 11:56:47 AM

Quant Time: Oct 27 00:42:28 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 23 15:06:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	135394	20.00	ng	-0.01
21) Naphthalene-d8	8.25	136	520506	20.00	ng	-0.02
39) Acenaphthene-d10	10.01	164	201748	20.00	ng	-0.02
64) Phenanthrene-d10	11.50	188	313435	20.00	ng	-0.01
76) Chrysene-d12	14.15	240	227629	20.00	ng	-0.01
87) Perylene-d12	15.68	264	157327	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	257445	30.96	ng	-0.02
7) Phenol-d6	6.58	99	337735	31.76	ng	-0.02
23) Nitrobenzene-d5	7.53	82	185656	21.16	ng	-0.02
42) 2,4,6-Tribromophenol	10.80	330	50910	25.47	ng	-0.01
45) 2-Fluorobiphenyl	9.32	172	359793	28.00	ng	-0.02
79) Terphenyl-d14	13.09	244	271835	24.84	ng	-0.01
Target Compounds						Qvalue
71) Phenanthrene	11.53	178	36740	2.240	ng	97
75) Fluoranthene	12.72	202	68317	4.104	ng	98
78) Pyrene	12.96	202	56932	3.489	ng	97
81) Benzo(a)anthracene	14.14	228	27985	2.073	ng	96
88) Benzo(b)fluoranthene	15.22	252	26349m	2.900	ng	
90) Benzo(a)pyrene	15.61	252	17608	2.106	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102618\
Data File : BF110225.D
Acq On : 26 Oct 2018 22:19
Operator : JU/SJ
Sample : J5205-05 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR-05-102518

Manual Integrations
APPROVED

Sohil
10/29/2018 11:56:47 AM

Quant Time: Oct 27 00:42:28 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 23 15:06:18 2018
Response via : Initial Calibration

