


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF102618\
 Data File : BF110206.D
 Acq On : 26 Oct 2018 13:36
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 23 15:06:18 2018



AutoFind: Scans 1640, 1641, 1642; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.6	35759	PASS
68	69	0.00	2	1.8	602	PASS
69	198	0.00	100	35.4	32794	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.1	45520	PASS
197	198	0.00	2	0.4	358	PASS
198	198	100	100	100.0	92648	PASS
199	198	5	9	6.7	6229	PASS
275	198	10	60	26.1	24170	PASS
365	198	1	100	3.0	2764	PASS
441	198	0.01	100	14.9	13776	PASS
442	198	50	100	95.0	88034	PASS
443	442	15	24	19.1	16852	PASS