

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102623\
 Data File : BF135958.D
 Acq On : 26 Oct 2023 10:24
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 12:01:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.602	0.591	1.8	102	0.00
3	Pyridine	1.661	1.622	2.3	101	0.00
4	n-Nitrosodimethylamine	0.724	0.705	2.6	101	0.00
5 S	2-Fluorophenol	1.295	1.272	1.8	103	0.00
6	Aniline	2.171	2.105	3.0	102	0.00
7 S	Phenol-d6	1.612	1.589	1.4	103	0.00
8	2-Chlorophenol	1.341	1.342	-0.1	104	0.00
9	Benzaldehyde	0.872	0.851	2.4	102	0.00
10 C	Phenol	1.686	1.624	3.7	102	0.00
11	bis(2-Chloroethyl)ether	1.312	1.291	1.6	104	0.00
12	1,3-Dichlorobenzene	1.424	1.386	2.7	103	0.00
13 C	1,4-Dichlorobenzene	1.430	1.379	3.6	101	0.00
14	1,2-Dichlorobenzene	1.328	1.281	3.5	102	0.00
15	Benzyl Alcohol	1.140	1.035	9.2	93	0.00
16	2,2'-oxybis(1-Chloropropane	1.932	1.819	5.8	99	0.00
17	2-Methylphenol	1.182	1.169	1.1	104	0.00
18	Hexachloroethane	0.486	0.484	0.4	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.917	0.879	4.1	100	0.00
20	3+4-Methylphenols	1.426	1.390	2.5	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.452	0.449	0.7	101	0.00
23 S	Nitrobenzene-d5	0.290	0.337	-16.2	114	0.00
24	Nitrobenzene	0.300	0.350	-16.7	108	0.00
25	Isophorone	0.632	0.632	0.0	100	0.00
26 C	2-Nitrophenol	0.085	0.147	-72.9#	144	0.00
27	2,4-Dimethylphenol	0.288	0.305	-5.9	105	0.00
28	bis(2-Chloroethoxy)methane	0.388	0.385	0.8	100	0.00
29 C	2,4-Dichlorophenol	0.258	0.278	-7.8	104	0.00
30	1,2,4-Trichlorobenzene	0.292	0.297	-1.7	103	0.00
31	Naphthalene	0.967	0.971	-0.4	101	0.00
32	Benzoic acid	0.123	0.188	-52.8#	159#	0.00
33	4-Chloroaniline	0.427	0.432	-1.2	101	0.00
34 C	Hexachlorobutadiene	0.165	0.166	-0.6	103	0.00
35	Caprolactam	0.086	0.094	-9.3	102	0.00
36 C	4-Chloro-3-methylphenol	0.274	0.294	-7.3	103	0.00
37	2-Methylnaphthalene	0.640	0.649	-1.4	101	0.00
38	1-Methylnaphthalene	0.598	0.603	-0.8	101	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.565	0.577	-2.1	103	0.00
41 P	Hexachlorocyclopentadiene	0.279	0.305	-9.3	103	0.00
42 S	2,4,6-Tribromophenol	0.175	0.190	-8.6	99	0.00
43 C	2,4,6-Trichlorophenol	0.331	0.354	-6.9	102	0.00
44	2,4,5-Trichlorophenol	0.393	0.434	-10.4	104	0.00
45 S	2-Fluorobiphenyl	1.249	1.233	1.3	101	0.00
46	1,1'-Biphenyl	1.515	1.526	-0.7	101	0.00
47	2-Chloronaphthalene	1.121	1.144	-2.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102623\
 Data File : BF135958.D
 Acq On : 26 Oct 2023 10:24
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 12:01:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.251	0.353	-40.6#	120	0.00
49	Acenaphthylene	1.768	1.814	-2.6	101	0.00
50	Dimethylphthalate	1.297	1.324	-2.1	101	0.00
51	2,6-Dinitrotoluene	0.220	0.295	-34.1#	113	0.00
52 C	Acenaphthene	1.133	1.169	-3.2	102	0.00
53	3-Nitroaniline	0.268	0.348	-29.9#	112	0.00
54 P	2,4-Dinitrophenol	0.060	0.092	-53.3#	163#	0.00
55	Dibenzofuran	1.637	1.658	-1.3	100	0.00
56 P	4-Nitrophenol	0.220	0.258	-17.3	109	0.00
57	2,4-Dinitrotoluene	0.262	0.369	-40.8#	118	0.00
58	Fluorene	1.225	1.213	1.0	98	0.00
59	2,3,4,6-Tetrachlorophenol	0.295	0.344	-16.6	108	0.00
60	Diethylphthalate	1.237	1.261	-1.9	99	0.00
61	4-Chlorophenyl-phenylether	0.591	0.584	1.2	99	0.00
62	4-Nitroaniline	0.266	0.337	-26.7#	110	0.00
63	Azobenzene	1.255	1.256	-0.1	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.057	0.085	-49.1#	145	0.00
66 c	n-Nitrosodiphenylamine	0.604	0.626	-3.6	100	0.00
67	4-Bromophenyl-phenylether	0.205	0.190	7.3	88	0.00
68	Hexachlorobenzene	0.219	0.227	-3.7	100	0.00
69	Atrazine	0.165	0.166	-0.6	97	0.00
70 C	Pentachlorophenol	0.124	0.146	-17.7	109	0.00
71	Phenanthrene	1.012	1.031	-1.9	97	0.00
72	Anthracene	1.036	1.078	-4.1	99	0.00
73	Carbazole	0.924	0.943	-2.1	97	0.00
74	Di-n-butylphthalate	0.983	1.037	-5.5	98	0.00
75 C	Fluoranthene	1.065	1.103	-3.6	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzydine	0.386	0.532	-37.8#	118	0.00
78	Pyrene	1.679	1.635	2.6	99	0.00
79 S	Terphenyl-d14	1.197	1.161	3.0	99	0.00
80	Butylbenzylphthalate	0.548	0.617	-12.6	105	0.00
81	Benzo(a)anthracene	1.360	1.384	-1.8	105	0.00
82	3,3'-Dichlorobenzidine	0.414	0.444	-7.2	109	0.00
83	Chrysene	1.324	1.357	-2.5	105	0.00
84	Bis(2-ethylhexyl)phthalate	0.691	0.760	-10.0	108	0.00
85 c	Di-n-octyl phthalate	1.088	1.267	-16.5	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.134	1.233	-8.7	103	0.00
88	Benzo(b)fluoranthene	1.190	1.198	-0.7	107	0.00
89	Benzo(k)fluoranthene	1.169	1.226	-4.9	110	0.00
90 C	Benzo(a)pyrene	1.097	1.134	-3.4	107	0.00
91	Dibenzo(a,h)anthracene	0.970	1.020	-5.2	104	0.00
92	Benzo(g,h,i)perylene	0.983	1.010	-2.7	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102623\
Data File : BF135958.D
Acq On : 26 Oct 2023 10:24
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 26 12:01:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 26 03:58:01 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1