

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102723\
 Data File : BF135984.D
 Acq On : 27 Oct 2023 14:56
 Operator : CG\JU
 Sample : 04882-04RE
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-228RE

Quant Time: Oct 27 22:30:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	171667	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	669469	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	326192	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	510390	20.000	ng	0.00
76) Chrysene-d12	14.010	240	346402	20.000	ng	0.00
86) Perylene-d12	15.498	264	404198	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	1065662	95.858	ng	0.00
7) Phenol-d6	6.457	99	1299350	93.881	ng	0.00
23) Nitrobenzene-d5	7.387	82	744910	73.368	ng	0.00
42) 2,4,6-Tribromophenol	10.651	330	303038	106.137	ng	0.00
45) 2-Fluorobiphenyl	9.187	172	1340882	65.823	ng	0.00
79) Terphenyl-d14	12.951	244	989767	47.739	ng	0.00
Target Compounds						
						Qvalue
19) n-Nitroso-di-n-propyla...	7.387	70	99693	12.667	ng	# 76
24) Nitrobenzene	7.387	77	2487	2.794	ng	# 29
25) Isophorone	7.387	82	744901	35.201	ng	# 64
32) Benzoic acid	7.881	122	65	5.733	ng	# 1
48) 2-Nitroaniline	9.298	65	895	4.400	ng	# 4
50) Dimethylphthalate	9.863	163	67811	3.206	ng	# 1
51) 2,6-Dinitrotoluene	9.863	165	42839	13.641	ng	# 21
53) 3-Nitroaniline	9.969	138	68	4.574	ng	# 1
56) 4-Nitrophenol	10.063	139	8094	7.934	ng	# 15
57) 2,4-Dinitrotoluene	9.863	165	42839	12.762	ng	# 26
62) 4-Nitroaniline	10.410	138	308	4.703	ng	# 18
65) 4,6-Dinitro-2-methylph...	10.445	198	264	7.120	ng	# 61
67) 4-Bromophenyl-phenylether	10.651	248	18613	3.561	ng	# 1
70) Pentachlorophenol	11.151	266	2492	4.891	ng	99
71) Phenanthrene	11.375	178	265119	10.267	ng	99
75) Fluoranthene	12.575	202	481045	17.708	ng	99
78) Pyrene	12.804	202	382399	13.147	ng	97
80) Butylbenzylphthalate	13.439	149	1813	3.979	ng	# 94
81) Benzo(a)anthracene	14.033	228	274017	11.634	ng	95
83) Chrysene	14.033	228	274017	11.945	ng	98
87) Indeno(1,2,3-cd)pyrene	16.992	276	71182	3.106	ng	98
88) Benzo(b)fluoranthene	15.057	252	537378	22.340	ng	98
89) Benzo(k)fluoranthene	15.057	252	537378	22.746	ng	97
90) Benzo(a)pyrene	15.427	252	188079	8.484	ng	99
92) Benzo(g,h,i)perylene	17.451	276	54457	6.010	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102723\
Data File : BF135984.D
Acq On : 27 Oct 2023 14:56
Operator : CG\JU
Sample : 04882-04RE
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-228RE

Quant Time: Oct 27 22:30:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 26 03:58:01 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.828 min Scan# 806
Delta R.T. -0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:152 Resp: 171667
Ion Ratio Lower Upper
152 100
150 153.2 124.6 186.8
115 58.9 46.6 69.8



#5
2-Fluorophenol
Concen: 95.858 ng
RT: 5.457 min Scan# 573
Delta R.T. 0.006 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:112 Resp: 1065662
Ion Ratio Lower Upper
112 100
64 58.3 45.1 67.7
63 29.5 23.0 34.4









#7
Phenol-d6
Concen: 93.881 ng
RT: 6.457 min Scan# 743
Delta R.T. 0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion: 99 Resp: 1299350
Ion Ratio Lower Upper
99 100
42 16.9 13.8 20.8
71 31.7 25.1 37.7



#19
n-Nitroso-di-n-propylamine
Concen: 12.667 ng
RT: 7.387 min Scan# 901
Delta R.T. 0.147 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion: 70 Resp: 99693
Ion Ratio Lower Upper
70 100
42 45.1 47.3 70.9#
101 0.0 7.8 11.8#
130 2.5 16.9 25.3#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.104 min Scan# 1023
Delta R.T. 0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:136 Resp: 669469
Ion Ratio Lower Upper
136 100
137 10.9 9.0 13.6
54 8.2 7.1 10.7
68 5.8 5.2 7.8



#23
Nitrobenzene-d5
Concen: 73.368 ng
RT: 7.387 min Scan# 901
Delta R.T. -0.005 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion: 82 Resp: 744910
Ion Ratio Lower Upper
82 100
128 46.9 36.8 55.2
54 52.1 43.4 65.0









#24
Nitrobenzene
Concen: 2.794 ng
RT: 7.387 min Scan# 901
Delta R.T. -0.023 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion: 77 Resp: 2487
Ion Ratio Lower Upper
77 100
123 0.0 40.2 60.2#
65 40.6 11.0 16.4#



#25
Isophorone
Concen: 35.201 ng
RT: 7.387 min Scan# 901
Delta R.T. -0.259 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion: 82 Resp: 744901
Ion Ratio Lower Upper
82 100
95 0.0 5.6 8.4#
138 0.0 14.8 22.2#









#32
Benzoic acid
Concen: 5.733 ng
RT: 7.881 min Scan# 91
Delta R.T. 0.012 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:122 Resp: 65
Ion Ratio Lower Upper
122 100
105 0.0 113.1 153.1#
77 0.0 75.2 115.2#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.863 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:164 Resp: 326192
Ion Ratio Lower Upper
164 100
162 102.6 82.0 123.0
160 45.6 35.6 53.4









#42
2,4,6-Tribromophenol
Concen: 106.137 ng
RT: 10.651 min Scan# 1456
Delta R.T. -0.006 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:330 Resp: 303038
Ion Ratio Lower Upper
330 100
332 97.5 77.4 116.2
141 29.6 24.9 37.3



#45
2-Fluorobiphenyl
Concen: 65.823 ng
RT: 9.187 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:172 Resp: 1340882
Ion Ratio Lower Upper
172 100
171 35.8 28.6 43.0
170 23.6 18.6 28.0









#48
2-Nitroaniline
Concen: 4.400 ng
RT: 9.298 min Scan# 11
Delta R.T. -0.106 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion: 65 Resp: 895
Ion Ratio Lower Upper
65 100
92 0.0 54.9 82.3#
138 0.0 85.2 127.8#



#50
Dimethylphthalate
Concen: 3.206 ng
RT: 9.863 min Scan# 1322
Delta R.T. 0.271 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion: 163 Resp: 67811
Ion Ratio Lower Upper
163 100
194 0.0 3.4 5.0#
164 485.4 8.3 12.5#









#51
2,6-Dinitrotoluene
Concen: 13.641 ng
RT: 9.863 min Scan# 11
Delta R.T. 0.218 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:165 Resp: 42839
Ion Ratio Lower Upper
165 100
63 1.3 51.7 77.5#
89 1.0 48.3 72.5#



#53
3-Nitroaniline
Concen: 4.574 ng
RT: 9.969 min Scan# 1340
Delta R.T. 0.153 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:138 Resp: 68
Ion Ratio Lower Upper
138 100
108 0.0 11.3 16.9#
92 0.0 95.0 142.6#









#56
4-Nitrophenol
Concen: 7.934 ng
RT: 10.063 min Scan# 11
Delta R.T. 0.094 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:139 Resp: 8094
Ion Ratio Lower Upper
139 100
109 2.9 45.2 85.2#
65 3.1 65.1 105.1#



#57
2,4-Dinitrotoluene
Concen: 12.762 ng
RT: 9.863 min Scan# 1322
Delta R.T. -0.188 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:165 Resp: 42839
Ion Ratio Lower Upper
165 100
63 1.3 33.7 50.5#
89 1.0 54.6 81.8#
182 0.0 2.1 3.1#









#62
4-Nitroaniline
Concen: 4.703 ng
RT: 10.410 min Scan# 1415
Delta R.T. -0.023 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:138 Resp: 308
Ion Ratio Lower Upper
138 100
92 0.0 32.9 72.9#
108 0.0 53.5 93.5#



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:188 Resp: 510390
Ion Ratio Lower Upper
188 100
94 10.7 8.6 13.0
80 11.4 9.0 13.6









#65
4,6-Dinitro-2-methylphenol
Concen: 7.120 ng
RT: 10.445 min Scan# 1423
Delta R.T. -0.012 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56




Instrument :
BNA_F
ClientSampleId :
VNJ-228RE


Tgt Ion:198 Resp: 264
Ion Ratio Lower Upper
198 100
51 0.0 28.7 68.7#
105 43.6 27.2 67.2



#67
4-Bromophenyl-phenylether
Concen: 3.561 ng
RT: 10.651 min Scan# 1456
Delta R.T. -0.247 min
Lab File: BF135984.D
Acq: 27 Oct 2023 14:56

Tgt Ion:248 Resp: 18613
Ion Ratio Lower Upper
248 100
250 198.0 77.9 116.9#
141 472.5 60.8 91.2#





#70

Pentachlorophenol

Concen: 4.891 ng

RT: 11.151 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:266 Resp: 2492

Ion Ratio Lower Upper

266 100

268 61.8 50.8 76.2

264 61.8 49.8 74.6



#71

Phenanthrene

Concen: 10.267 ng

RT: 11.375 min Scan# 1579

Delta R.T. -0.006 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:178 Resp: 265119


Ion Ratio Lower Upper

178 100

176 18.6 15.4 23.2

179 15.6 12.5 18.7





#75

Fluoranthene

Concen: 17.708 ng

RT: 12.575 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:202 Resp: 481045

Ion Ratio Lower Upper

202 100

101 11.1 0.0 30.6

203 17.3 0.0 37.5



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.010 min Scan# 2027

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:240 Resp: 346402


Ion Ratio Lower Upper

240 100

120 10.7 8.6 13.0

236 24.9 20.6 30.8





#78

Pyrene

Concen: 13.147 ng

RT: 12.804 min Scan# 1822

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:202 Resp: 382399

Ion Ratio Lower Upper

202 100

200 20.3 17.0 25.6

203 16.7 14.5 21.7



#79

Terphenyl-d14

Concen: 47.739 ng

RT: 12.951 min Scan# 1847

Delta R.T. -0.006 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:244 Resp: 989767


Ion Ratio Lower Upper

244 100

212 6.4 5.0 7.6

122 12.3 8.0 12.0#





#80

Butylbenzylphthalate

Concen: 3.979 ng

RT: 13.439 min Scan# 1930

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:149 Resp: 1813

Ion Ratio Lower Upper

149 100

91 63.6 53.4 80.0

206 28.6 17.5 26.3#



#81

Benzo(a)anthracene

Concen: 11.634 ng

RT: 14.033 min Scan# 2031

Delta R.T. 0.036 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:228 Resp: 274017


Ion Ratio Lower Upper

228 100

226 29.7 21.8 32.6

229 22.2 16.2 24.4





#83

Chrysene

Concen: 11.945 ng

RT: 14.033 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:228 Resp: 274017

Ion Ratio Lower Upper

228 100

226 29.7 23.9 35.9

229 22.2 16.2 24.2



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.498 min Scan# 2280

Delta R.T. 0.006 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:264 Resp: 404198


Ion Ratio Lower Upper

264 100

260 22.8 18.6 28.0

265 23.0 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 3.106 ng

RT: 16.992 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:276 Resp: 71182

Ion Ratio Lower Upper

276 100

138 25.3 21.8 32.6

277 24.8 20.0 30.0



#88

Benzo(b)fluoranthene

Concen: 22.340 ng

RT: 15.057 min Scan# 2205

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:252 Resp: 537378


Ion Ratio Lower Upper

252 100

253 23.3 18.0 27.0

125 11.9 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 22.746 ng

RT: 15.057 min Scan# 2210

Delta R.T. -0.029 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56

Instrument :

BNA_F

ClientSampleId :

VNJ-228RE





Tgt Ion:252 Resp: 537378

Ion Ratio Lower Upper

252 100

253 23.3 17.7 26.5

125 11.9 8.5 12.7



#90

Benzo(a)pyrene

Concen: 8.484 ng

RT: 15.427 min Scan# 2268

Delta R.T. 0.000 min

Lab File: BF135984.D

Acq: 27 Oct 2023 14:56




Tgt Ion:252 Resp: 188079


Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.2


125 11.6 9.6 14.4





#92
 Benzo(g,h,i)perylene
 Concen: 6.010 ng
 RT: 17.451 min Scan# 20
 Delta R.T. -0.005 min
 Lab File: BF135984.D
 Acq: 27 Oct 2023 14:56

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-228RE



Tgt Ion:	276	Resp:	54457
Ion Ratio	Lower	Upper	
276	100		
277	25.0	18.6	28.0
138	25.4	18.4	27.6

