

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102823\
 Data File : BF136010.D
 Acq On : 28 Oct 2023 17:39
 Operator : CG\JU
 Sample : 05051-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUCT-BANK-1A

Quant Time: Oct 30 00:27:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 26 03:58:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	152716	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	596647	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	298937	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	544893	20.000	ng	0.00
76) Chrysene-d12	14.004	240	333100	20.000	ng	0.00
86) Perylene-d12	15.492	264	356951	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	751463	75.983	ng	0.00
7) Phenol-d6	6.451	99	928368	75.400	ng	0.00
23) Nitrobenzene-d5	7.387	82	532271	60.763	ng	0.00
42) 2,4,6-Tribromophenol	10.651	330	249682	95.423	ng	0.00
45) 2-Fluorobiphenyl	9.181	172	970082	51.962	ng	0.00
79) Terphenyl-d14	12.951	244	1047985	52.565	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102823\
Data File : BF136010.D
Acq On : 28 Oct 2023 17:39
Operator : CG\JU
Sample : 05051-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DUCT-BANK-1A

Quant Time: Oct 30 00:27:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 26 03:58:01 2023
Response via : Initial Calibration

