




Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF102919\
Data File : BF117578.D
Acq On : 29 Oct 2019 9:29
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 29 19:17:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration



TIC: BF117578.D

(77) Benzidine




12.586min (+0.000) 0.00ng

response 191211

Ion	Exp%	Act%
184.00	100	100
185.00	14.20	13.89
183.00	12.00	10.49
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF102919\
Data File : BF117578.D
Acq On : 29 Oct 2019 9:29
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 29 19:17:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration



TIC: BF117578.D

(70) Pentachlorophenol (C)

11.075min (+0.006) 0.00ng

response 23640

Ion	Exp%	Act%
265.70	100	100
268.00	64.50	58.25
264.00	64.30	63.78
0.00	0.00	0.00