

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
 Data File : BF144143.D
 Acq On : 29 Oct 2025 22:27
 Operator : RC/JU
 Sample : Q3472-14 50X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 155

Quant Time: Oct 30 03:49:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 17:30:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	53936	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	198447	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	105399	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	164955	20.000	ng	# 0.00
76) Chrysene-d12	14.063	240	127321	20.000	ng	0.00
86) Perylene-d12	15.551	264	128263	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1221	0.358	ng	0.04
7) Phenol-d6	6.528	99	1847	0.491	ng	0.02
23) Nitrobenzene-d5	7.445	82	2068	0.571	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	1015	0.756	ng	0.01
45) 2-Fluorobiphenyl	9.234	172	8880	1.186	ng	0.00
79) Terphenyl-d14	12.998	244	7392	1.015	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	11.969	149	35956	4.070	ng	# 96
84) Bis(2-ethylhexyl)phtha...	14.045	149	1629220	344.627	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144143.D
Acq On : 29 Oct 2025 22:27
Operator : RC/JU
Sample : Q3472-14 50X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
155

Quant Time: Oct 30 03:49:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

