

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/30/2023	BNA_F	<u>BF136022.D</u>
Compound Name	Response	Retention Time
DDT	571640	13.533
DDD	9565	13.263
DDE	214	12.892
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9779	581419	1.68