


Benzoic acid

Response Ratio




$$\text{Response} = 2.594\text{e-}001 * \text{Amt} - 6.634\text{e-}002$$

Coef of Det (r^2) = 0.997858 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF103023.M


Calibration Table Last Updated: Tue Oct 31 01:13:02 2023

2,4-Dinitrophenol



Response = 1.699e-001 * Amt - 6.775e-002
 Coef of Det (r^2) = 0.993305 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF103023.M
 Calibration Table Last Updated: Tue Oct 31 01:13:02 2023

4,6-Dinitro-2-methylphenol



$$\text{Response} = 1.250\text{e-}001 * \text{Amt} - 3.618\text{e-}002$$


Coef of Det (r^2) = 0.996843 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF103023.M

Calibration Table Last Updated: Tue Oct 31 01:13:02 2023

Benzo(g,h,i)perylene

Response Ratio



$$\text{Response} = 1.297\text{e}+000 * \text{Amt} - 1.878\text{e}-001$$

Coef of Det (r^2) = 0.996651 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF103023.M

Calibration Table Last Updated: Tue Oct 31 01:13:02 2023