

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103025\
 Data File : BF144153.D
 Acq On : 30 Oct 2025 21:36
 Operator : RC/JU
 Sample : Q3488-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MOO-25-0294-0298

Quant Time: Oct 31 01:02:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 17:30:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	59484	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	221189	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	111825	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	156904	20.000	ng	0.00
76) Chrysene-d12	14.051	240	166602	20.000	ng	0.00
86) Perylene-d12	15.539	264	209021	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	150165	39.940	ng	0.00
7) Phenol-d6	6.498	99	257094	61.916	ng	0.00
23) Nitrobenzene-d5	7.434	82	187891	46.539	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	37736	26.492	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	415562	52.295	ng	0.00
79) Terphenyl-d14	12.992	244	407224	42.732	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103025\
Data File : BF144153.D
Acq On : 30 Oct 2025 21:36
Operator : RC/JU
Sample : Q3488-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298


Quant Time: Oct 31 01:02:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 797
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298







Tgt Ion:152 Resp: 59484
Ion Ratio Lower Upper
152 100
150 163.1 124.4 186.6
115 63.1 49.4 74.0



#5
2-Fluorophenol
Concen: 39.940 ng
RT: 5.493 min Scan# 561
Delta R.T. 0.000 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Tgt Ion:112 Resp: 150165
Ion Ratio Lower Upper
112 100
64 57.2 49.8 74.8
63 28.8 27.4 41.2









#7
Phenol-d6
Concen: 61.916 ng
RT: 6.498 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36




Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298


Tgt Ion: 99 Resp: 257094
Ion Ratio Lower Upper
99 100
42 25.0 23.9 35.9
71 38.3 34.4 51.6



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Tgt Ion: 136 Resp: 221189
Ion Ratio Lower Upper
136 100
137 10.5 8.6 13.0
54 8.7 6.6 9.8
68 5.4 4.6 6.8









#23
Nitrobenzene-d5
Concen: 46.539 ng
RT: 7.434 min Scan# 891
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36




Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298


Tgt Ion: 82 Resp: 187891
Ion Ratio Lower Upper
82 100
128 42.8 32.4 48.6
54 63.0 54.0 81.0



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Tgt Ion:164 Resp: 111825
Ion Ratio Lower Upper
164 100
162 98.4 79.4 119.2
160 43.4 36.7 55.1





#42

2,4,6-Tribromophenol

Concen: 26.492 ng

RT: 10.704 min Scan# 1447

Delta R.T. -0.006 min

Lab File: BF144153.D


Acq: 30 Oct 2025 21:36

Instrument :

BNA_F

ClientSampleId :

MOO-25-0294-0298






Tgt Ion:330 Resp: 37736

Ion Ratio Lower Upper

330 100

332 89.8 77.0 115.6

141 22.0 19.4 29.0



#45

2-Fluorobiphenyl

Concen: 52.295 ng

RT: 9.234 min Scan# 1197

Delta R.T. -0.006 min

Lab File: BF144153.D

Acq: 30 Oct 2025 21:36




Tgt Ion:172 Resp: 415562


Ion Ratio Lower Upper

172 100

171 34.4 27.5 41.3

170 22.6 19.0 28.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1566
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36




Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298


Tgt Ion:188 Resp: 156904
Ion Ratio Lower Upper
188 100
94 7.3 5.4 8.0
80 8.1 6.0 9.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36


Tgt Ion:240 Resp: 166602
Ion Ratio Lower Upper
240 100
120 8.6 6.5 9.7
236 26.5 20.1 30.1








#79
Terphenyl-d14
Concen: 42.732 ng
RT: 12.992 min Scan# 1836
Delta R.T. -0.006 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Instrument :
BNA_F
ClientSampleId :
MOO-25-0294-0298



Tgt Ion:244 Resp: 407224
Ion Ratio Lower Upper
244 100
212 6.2 5.8 8.6
122 7.2 4.9 7.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144153.D
Acq: 30 Oct 2025 21:36

Tgt Ion:264 Resp: 209021
Ion Ratio Lower Upper
264 100
260 21.3 17.0 25.4
265 21.5 16.6 24.8

