

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF103119\
 Data File : BF117657.D
 Acq On : 31 Oct 2019 14:51
 Operator : JU
 Sample : PB124409BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB124409BL

Quant Time: Nov 01 01:38:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 18:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	39032	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	159194	20.00	ng	0.00
39) Acenaphthene-d10	9.76	164	82815	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	139813	20.00	ng	0.00
76) Chrysene-d12	13.90	240	97170	20.00	ng	0.00
87) Perylene-d12	15.34	264	82661	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	186233	70.98	ng	0.00
7) Phenol-d6	6.39	99	268916	76.31	ng	0.00
23) Nitrobenzene-d5	7.29	82	244035	75.68	ng	-0.01
42) 2,4,6-Tribromophenol	10.56	330	49391	68.93	ng	0.00
45) 2-Fluorobiphenyl	9.08	172	422328	71.85	ng	-0.01
79) Terphenyl-d14	12.83	244	432612	72.62	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF103119\
Data File : BF117657.D
Acq On : 31 Oct 2019 14:51
Operator : JU
Sample : PB124409BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB124409BL

Quant Time: Nov 01 01:38:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration

