

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110123\
 Data File : BF136085.D
 Acq On : 01 Nov 2023 19:33
 Operator : CG\JU
 Sample : 05142-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PSC-124037

Quant Time: Nov 01 22:42:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF103023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 31 01:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	148402	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	574586	20.000	ng	0.00
39) Acenaphthene-d10	9.863	164	279799	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	436782	20.000	ng	0.00
76) Chrysene-d12	14.010	240	312754	20.000	ng	0.00
86) Perylene-d12	15.504	264	290774	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	1208189	127.934	ng	0.01
7) Phenol-d6	6.463	99	1461412	124.201	ng	0.00
23) Nitrobenzene-d5	7.387	82	903301	94.770	ng	0.00
42) 2,4,6-Tribromophenol	10.657	330	340018	113.851	ng	0.00
45) 2-Fluorobiphenyl	9.186	172	1520682	86.638	ng	0.00
79) Terphenyl-d14	12.957	244	1240985	61.663	ng	0.00
Target Compounds						
92) Benzo(g,h,i)perylene	17.468	276	6231	3.225	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110123\
Data File : BF136085.D
Acq On : 01 Nov 2023 19:33
Operator : CG\JU
Sample : 05142-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PSC-124037

Quant Time: Nov 01 22:42:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF103023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 31 01:13:02 2023
Response via : Initial Calibration

