


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110218\
 Data File : BF110438.D
 Acq On : 3 Nov 2018 3:50
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 01 16:06:12 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.0	14713	PASS
68	69	0.00	2	1.7	233	PASS
69	198	0.00	100	34.9	13873	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.3	19604	PASS
197	198	0.00	2	0.3	135	PASS
198	198	100	100	100.0	39778	PASS
199	198	5	9	7.0	2785	PASS
275	198	10	60	26.4	10487	PASS
365	198	1	100	2.6	1050	PASS
441	198	0.01	100	15.9	6306	PASS
442	198	50	100	96.0	38178	PASS
443	442	15	24	19.8	7561	PASS