

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110323\
 Data File : BF136114.D
 Acq On : 03 Nov 2023 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 04 00:04:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF103023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 31 01:13:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.554	0.532	4.0	65	0.00
3	Pyridine	1.513	1.420	6.1	64	-0.01
4	n-Nitrosodimethylamine	0.632	0.637	-0.8	69	0.00
5 S	2-Fluorophenol	1.273	1.231	3.3	67	0.00
6	Aniline	2.053	1.990	3.1	67	0.00
7 S	Phenol-d6	1.586	1.535	3.2	66	0.00
8	2-Chlorophenol	1.361	1.346	1.1	67	0.00
9	Benzaldehyde	0.884	0.861	2.6	68	0.00
10 C	Phenol	1.640	1.583	3.5	65	0.00
11	bis(2-Chloroethyl)ether	1.284	1.223	4.8	66	0.00
12	1,3-Dichlorobenzene	1.416	1.373	3.0	67	0.00
13 C	1,4-Dichlorobenzene	1.419	1.403	1.1	68	0.00
14	1,2-Dichlorobenzene	1.327	1.285	3.2	66	0.00
15	Benzyl Alcohol	1.124	1.154	-2.7	70	0.00
16	2,2'-oxybis(1-Chloropropane	1.767	1.646	6.8	64	0.00
17	2-Methylphenol	1.154	1.123	2.7	67	0.00
18	Hexachloroethane	0.499	0.507	-1.6	69	0.00
19 P	n-Nitroso-di-n-propylamine	0.906	0.866	4.4	67	0.00
20	3+4-Methylphenols	1.408	1.373	2.5	67	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
22	Acetophenone	0.443	0.448	-1.1	68	0.00
23 S	Nitrobenzene-d5	0.332	0.359	-8.1	71	0.00
24	Nitrobenzene	0.336	0.352	-4.8	69	0.00
25	Isophorone	0.626	0.633	-1.1	68	0.00
26 C	2-Nitrophenol	0.153	0.175	-14.4	71	0.00
27	2,4-Dimethylphenol	0.289	0.289	0.0	66	0.00
28	bis(2-Chloroethoxy)methane	0.384	0.380	1.0	66	0.00
29 C	2,4-Dichlorophenol	0.273	0.284	-4.0	68	0.00
30	1,2,4-Trichlorobenzene	0.298	0.302	-1.3	68	0.00
31	Naphthalene	0.964	0.985	-2.2	69	0.00
32	Benzoic acid	0.213	0.223	-4.7	67	0.00
33	4-Chloroaniline	0.422	0.424	-0.5	67	0.00
34 C	Hexachlorobutadiene	0.171	0.182	-6.4	71	0.00
35	Caprolactam	0.089	0.090	-1.1	67	0.00
36 C	4-Chloro-3-methylphenol	0.286	0.300	-4.9	69	0.00
37	2-Methylnaphthalene	0.647	0.655	-1.2	68	0.00
38	1-Methylnaphthalene	0.602	0.618	-2.7	68	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
40	1,2,4,5-Tetrachlorobenzene	0.570	0.594	-4.2	69	0.00
41 P	Hexachlorocyclopentadiene	0.305	0.318	-4.3	67	0.00
42 S	2,4,6-Tribromophenol	0.213	0.221	-3.8	67	0.00
43 C	2,4,6-Trichlorophenol	0.372	0.382	-2.7	66	0.00
44	2,4,5-Trichlorophenol	0.431	0.453	-5.1	69	0.00
45 S	2-Fluorobiphenyl	1.255	1.309	-4.3	72	0.00
46	1,1'-Biphenyl	1.525	1.573	-3.1	69	0.00
47	2-Chloronaphthalene	1.124	1.161	-3.3	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110323\
 Data File : BF136114.D
 Acq On : 03 Nov 2023 09:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 04 00:04:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF103023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 31 01:13:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.333	0.373	-12.0	70	0.00
49	Acenaphthylene	1.753	1.827	-4.2	69	0.00
50	Dimethylphthalate	1.319	1.372	-4.0	69	0.00
51	2,6-Dinitrotoluene	0.282	0.311	-10.3	70	0.00
52 C	Acenaphthene	1.115	1.131	-1.4	67	0.00
53	3-Nitroaniline	0.329	0.345	-4.9	67	0.00
54 P	2,4-Dinitrophenol	0.123	0.138	-12.2	72	0.00
55	Dibenzofuran	1.625	1.668	-2.6	68	0.00
56 P	4-Nitrophenol	0.246	0.251	-2.0	64	0.00
57	2,4-Dinitrotoluene	0.357	0.404	-13.2	70	0.00
58	Fluorene	1.217	1.252	-2.9	69	0.00
59	2,3,4,6-Tetrachlorophenol	0.342	0.350	-2.3	67	0.00
60	Diethylphthalate	1.260	1.313	-4.2	68	0.00
61	4-Chlorophenyl-phenylether	0.607	0.629	-3.6	70	0.00
62	4-Nitroaniline	0.317	0.331	-4.4	66	0.00
63	Azobenzene	1.214	1.245	-2.6	68	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
65	4,6-Dinitro-2-methylphenol	0.099	0.117	-18.2	70	0.00
66 c	n-Nitrosodiphenylamine	0.603	0.644	-6.8	67	0.00
67	4-Bromophenyl-phenylether	0.210	0.227	-8.1	67	0.00
68	Hexachlorobenzene	0.224	0.239	-6.7	67	0.00
69	Atrazine	0.164	0.163	0.6	65	0.00
70 C	Pentachlorophenol	0.144	0.156	-8.3	66	0.00
71	Phenanthrene	1.007	1.062	-5.5	67	0.00
72	Anthracene	1.032	1.074	-4.1	66	0.00
73	Carbazole	0.897	0.916	-2.1	64	0.00
74	Di-n-butylphthalate	1.025	1.052	-2.6	65	0.00
75 C	Fluoranthene	1.035	1.023	1.2	62	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
77	Benzidine	0.390	0.474	-21.5	57	0.00
78	Pyrene	1.763	2.053	-16.4	61	0.00
79 S	Terphenyl-d14	1.287	1.479	-14.9	60	0.00
80	Butylbenzylphthalate	0.631	0.683	-8.2	56	0.00
81	Benzo(a)anthracene	1.324	1.370	-3.5	55	0.00
82	3,3'-Dichlorobenzidine	0.423	0.438	-3.5	54	0.00
83	Chrysene	1.304	1.320	-1.2	54	0.00
84	Bis(2-ethylhexyl)phthalate	0.735	0.752	-2.3	54	0.00
85 c	Di-n-octyl phthalate	1.102	1.064	3.4	51	0.00
86 I	Perylene-d12	1.000	1.000	0.0	63	0.00
87	Indeno(1,2,3-cd)pyrene	1.307	1.561	-19.4	68	0.00
88	Benzo(b)fluoranthene	1.183	1.152	2.6	62	0.00
89	Benzo(k)fluoranthene	1.171	1.092	6.7	58	0.00
90 C	Benzo(a)pyrene	1.100	1.127	-2.5	63	0.00
91	Dibenzo(a,h)anthracene	1.071	1.275	-19.0	68	0.00
92	Benzo(g,h,i)perylene	1.105	1.316	-19.1	67	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110323\
Data File : BF136114.D
Acq On : 03 Nov 2023 09:59
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 00:04:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF103023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 31 01:13:02 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0