

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110518\
 Data File : BF110469.D
 Acq On : 5 Nov 2018 17:10
 Operator : JU/SJ
 Sample : J5812-05
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP1-B

Manual Integrations
 APPROVED

Sohil
 11/6/2018 9:00:58 AM

Quant Time: Nov 06 04:09:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	97666	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	406512	20.00	ng	0.00
39) Acenaphthene-d10	10.00	164	201605	20.00	ng	0.00
64) Phenanthrene-d10	11.49	188	375715	20.00	ng	0.00
76) Chrysene-d12	14.13	240	211742	20.00	ng	0.00
87) Perylene-d12	15.66	264	168171	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	631023	105.21	ng	0.00
7) Phenol-d6	6.58	99	825403	107.60	ng	0.00
23) Nitrobenzene-d5	7.52	82	523744	76.42	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	209266	104.79	ng	0.00
45) 2-Fluorobiphenyl	9.31	172	938720	73.11	ng	0.00
79) Terphenyl-d14	13.07	244	839820	82.49	ng	0.00
Target Compounds						
10) Phenol	6.59	94	51380	6.266	ng	Qvalue # 64
50) Dimethylphthalate	9.70	163	162612	10.927	ng	99
75) Fluoranthene	12.70	202	100849	5.055	ng	98
78) Pyrene	12.93	202	78939	5.200	ng	97
88) Benzo(b)fluoranthene	15.20	252	19704m	2.029	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110518\
Data File : BF110469.D
Acq On : 5 Nov 2018 17:10
Operator : JU/SJ
Sample : J5812-05
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP1-B

Manual Integrations
APPROVED

Sohil
11/6/2018 9:00:58 AM

Quant Time: Nov 06 04:09:20 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

