

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110518\
 Data File : BF110471.D
 Acq On : 5 Nov 2018 18:03
 Operator : JU/SJ
 Sample : J5812-25
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP4-A

Manual Integrations
 APPROVED

Sohil
 11/6/2018 1:19:17 PM

Quant Time: Nov 06 10:28:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	94700	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	386564	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	193319	20.00	ng	0.00
64) Phenanthrene-d10	11.49	188	342830	20.00	ng	0.00
76) Chrysene-d12	14.13	240	190894	20.00	ng	0.00
87) Perylene-d12	15.66	264	160434	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	704988	121.23	ng	0.00
7) Phenol-d6	6.58	99	917421	123.34	ng	0.00
23) Nitrobenzene-d5	7.52	82	591958	90.83	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	226413	118.23	ng	0.00
45) 2-Fluorobiphenyl	9.31	172	1057030	85.86	ng	0.00
79) Terphenyl-d14	13.07	244	898146	97.85	ng	0.00
Target Compounds						
10) Phenol	6.59	94	57372	7.216	ng	Qvalue # 67
50) Dimethylphthalate	9.70	163	144422	10.120	ng	99
71) Phenanthrene	11.51	178	54109	3.016	ng	100
75) Fluoranthene	12.70	202	121037	6.648	ng	98
78) Pyrene	12.93	202	97006	7.088	ng	98
81) Benzo(a)anthracene	14.12	228	26653m	2.354	ng	
83) Chrysene	14.16	228	30411	2.828	ng	96
88) Benzo(b)fluoranthene	15.20	252	22019m	2.377	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110518\
Data File : BF110471.D
Acq On : 5 Nov 2018 18:03
Operator : JU/SJ
Sample : J5812-25
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP4-A

Manual Integrations
APPROVED

Sohil
11/6/2018 1:19:17 PM

Quant Time: Nov 06 10:28:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

