

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF110518\
 Data File : BF110455.D
 Acq On : 5 Nov 2018 10:44
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 11:22:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	51	0.00
2	1,4-Dioxane	0.643	0.612	4.8	49#	0.00
3	Pyridine	1.433	1.375	4.0	47#	0.00
4	n-Nitrosodimethylamine	0.600	0.592	1.3	49#	0.00
5 S	2-Fluorophenol	1.228	1.248	-1.6	52	0.00
6	Aniline	2.046	1.983	3.1	49#	0.00
7 S	Phenol-d6	1.571	1.559	0.8	51	0.00
8	2-Chlorophenol	1.416	1.449	-2.3	52	0.00
9	Benzaldehyde	0.954	0.842	11.7	46#	0.00
10 C	Phenol	1.679	1.807	-7.6	55	0.00
11	bis(2-Chloroethyl)ether	1.381	1.342	2.8	50#	0.00
12	1,3-Dichlorobenzene	1.558	1.565	-0.4	51	0.00
13 C	1,4-Dichlorobenzene	1.576	1.594	-1.1	52	0.00
14	1,2-Dichlorobenzene	1.463	1.479	-1.1	52	0.00
15	Benzyl Alcohol	1.208	1.206	0.2	50#	0.00
16	2,2'-oxybis(1-Chloropropane	2.209	2.157	2.4	50#	0.00
17	2-Methylphenol	1.128	1.115	1.2	50	0.00
18	Hexachloroethane	0.577	0.593	-2.8	53	0.00
19 P	n-Nitroso-di-n-propylamine	1.008	0.988	2.0	50	0.00
20	3+4-Methylphenols	1.459	1.417	2.9	49#	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	51	0.00
22	Acetophenone	0.461	0.466	-1.1	52	0.00
23 S	Nitrobenzene-d5	0.337	0.350	-3.9	52	0.00
24	Nitrobenzene	0.348	0.354	-1.7	51	0.00
25	Isophorone	0.575	0.561	2.4	49#	0.00
26 C	2-Nitrophenol	0.166	0.179	-7.8	52	0.00
27	2,4-Dimethylphenol	0.275	0.267	2.9	49#	0.00
28	bis(2-Chloroethoxy)methane	0.390	0.384	1.5	50	0.00
29 C	2,4-Dichlorophenol	0.277	0.281	-1.4	51	0.00
30	1,2,4-Trichlorobenzene	0.311	0.314	-1.0	51	0.00
31	Naphthalene	0.922	0.924	-0.2	51	0.00
32	Benzoic acid	0.212	0.191	9.9	44#	0.00
33	4-Chloroaniline	0.415	0.412	0.7	51	0.00
34 C	Hexachlorobutadiene	0.173	0.180	-4.0	54	0.00
35	Caprolactam	0.097	0.091	6.2	46#	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.302	-0.7	50	0.00
37	2-Methylnaphthalene	0.610	0.611	-0.2	51	0.00
38	1-Methylnaphthalene	0.589	0.588	0.2	51	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	49#	0.00
40	1,2,4,5-Tetrachlorobenzene	0.623	0.650	-4.3	51	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.328	-2.8	47#	0.00
42 S	2,4,6-Tribromophenol	0.198	0.203	-2.5	50#	0.00
43 C	2,4,6-Trichlorophenol	0.402	0.412	-2.5	49#	0.00
44	2,4,5-Trichlorophenol	0.425	0.430	-1.2	49#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF110518\
 Data File : BF110455.D
 Acq On : 5 Nov 2018 10:44
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 05 11:22:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.274	1.398	-9.7	54	0.00
46	1,1'-Biphenyl	1.809	1.894	-4.7	52	0.00
47	2-Chloronaphthalene	1.327	1.368	-3.1	51	0.00
48	2-Nitroaniline	0.402	0.422	-5.0	50	0.00
49	Acenaphthylene	1.916	1.951	-1.8	50	0.00
50	Dimethylphthalate	1.476	1.485	-0.6	50	0.00
51	2,6-Dinitrotoluene	0.318	0.338	-6.3	50#	0.00
52 C	Acenaphthene	1.268	1.219	3.9	48#	0.00
53	3-Nitroaniline	0.378	0.379	-0.3	47#	0.00
54 P	2,4-Dinitrophenol	0.112	0.116	-3.6	51	0.00
55	Dibenzofuran	1.775	1.805	-1.7	50	0.00
56 P	4-Nitrophenol	0.280	0.257	8.2	42#	0.00
57	2,4-Dinitrotoluene	0.404	0.435	-7.7	50	0.00
58	Fluorene	1.321	1.363	-3.2	51	0.00
59	2,3,4,6-Tetrachlorophenol	0.346	0.339	2.0	47#	0.00
60	Diethylphthalate	1.441	1.483	-2.9	51	0.00
61	4-Chlorophenyl-phenylether	0.697	0.715	-2.6	51	0.00
62	4-Nitroaniline	0.376	0.379	-0.8	49#	0.00
63	Azobenzene	1.431	1.441	-0.7	50#	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	48#	0.00
65	4,6-Dinitro-2-methylphenol	0.091	0.099	-8.8	50	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.685	-4.4	51	0.00
67	4-Bromophenyl-phenylether	0.217	0.220	-1.4	49#	0.00
68	Hexachlorobenzene	0.230	0.235	-2.2	49#	0.00
69	Atrazine	0.207	0.205	1.0	47#	0.00
70 C	Pentachlorophenol	0.134	0.127	5.2	41#	0.00
71	Phenanthrene	1.046	1.072	-2.5	50	0.00
72	Anthracene	1.049	1.085	-3.4	51	0.00
73	Carbazole	1.024	1.050	-2.5	50	0.00
74	Di-n-butylphthalate	1.157	1.224	-5.8	51	0.00
75 C	Fluoranthene	1.062	1.056	0.6	49#	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	46#	0.00
77	Benzidine	0.681	0.571	16.2	43#	0.00
78	Pyrene	1.434	1.543	-7.6	50#	0.00
79 S	Terphenyl-d14	0.962	1.052	-9.4	52	0.00
80	Butylbenzylphthalate	0.622	0.667	-7.2	49#	0.00
81	Benzo(a)anthracene	1.186	1.199	-1.1	46#	0.00
82	3,3'-Dichlorobenzidine	0.413	0.393	4.8	43#	0.00
83	Chrysene	1.127	1.134	-0.6	47#	0.00
84	Bis(2-ethylhexyl)phthalate	0.841	0.847	-0.7	46#	0.00
85 c	Di-n-octyl phthalate	1.308	1.292	1.2	45#	0.00
86	Indeno(1,2,3-cd)pyrene	0.935	0.833	10.9	45#	0.00
87 I	Perylene-d12	1.000	1.000	0.0	43#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF110518\
Data File : BF110455.D
Acq On : 5 Nov 2018 10:44
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 05 11:22:46 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.155	1.171	-1.4	43#	0.00
89	Benzo(k)fluoranthene	1.135	1.173	-3.3	44#	0.00
90 C	Benzo(a)pyrene	1.063	1.075	-1.1	43#	0.00
91	Dibenzo(a,h)anthracene	0.917	0.938	-2.3	45#	0.00
92	Benzo(q,h,i)perylene	0.904	0.929	-2.8	46#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0