


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110518\
 Data File : BF110480.D
 Acq On : 5 Nov 2018 23:00
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 01 16:06:12 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	18885	PASS
68	69	0.00	2	0.9	159	PASS
69	198	0.00	100	36.8	17442	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.4	22933	PASS
197	198	0.00	2	0.8	375	PASS
198	198	100	100	100.0	47354	PASS
199	198	5	9	6.7	3151	PASS
275	198	10	60	24.2	11465	PASS
365	198	1	100	2.1	1016	PASS
441	198	0.01	100	13.2	6241	PASS
442	198	50	100	81.0	38366	PASS
443	442	15	24	19.0	7280	PASS