

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110525\
 Data File : BF144171.D
 Acq On : 05 Nov 2025 13:50
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Nov 05 17:23:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 17:20:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	112609	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	438512	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	261107	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	469428	20.000	ng	0.00
76) Chrysene-d12	14.057	240	296506	20.000	ng	0.00
86) Perylene-d12	15.539	264	311371	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	146731	20.391	ng	0.00
7) Phenol-d6	6.498	99	170929	20.963	ng	-0.01
23) Nitrobenzene-d5	7.434	82	154488	20.347	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	66981	20.442	ng	-0.01
45) 2-Fluorobiphenyl	9.228	172	372845	21.090	ng	-0.01
79) Terphenyl-d14	12.992	244	419548	22.476	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.616	88	27411	9.214	ng	97
3) Pyridine	3.393	79	79148	9.536	ng	99
4) n-Nitrosodimethylamine	3.322	42	40456	9.331	ng	95
6) Aniline	6.534	93	122118	10.512	ng	98
8) 2-Chlorophenol	6.657	128	73729	10.448	ng	99
9) Benzaldehyde	6.422	77	58146	12.629	ng	98
10) Phenol	6.510	94	106765	10.717	ng	96
11) bis(2-Chloroethyl)ether	6.604	93	75075	10.342	ng	98
12) 1,3-Dichlorobenzene	6.816	146	88594	10.174	ng	98
13) 1,4-Dichlorobenzene	6.893	146	87984	10.085	ng	96
14) 1,2-Dichlorobenzene	7.045	146	84997	10.165	ng	98
15) Benzyl Alcohol	7.010	79	64160	9.937	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.140	45	139062	10.403	ng	97
17) 2-Methylphenol	7.122	107	59368	10.575	ng	98
18) Hexachloroethane	7.387	117	28876	9.963	ng	97
19) n-Nitroso-di-n-propyla...	7.275	70	57621	10.735	ng	98
20) 3+4-Methylphenols	7.275	107	74874	11.314	ng	96
22) Acetophenone	7.281	105	101777	10.819	ng	97
24) Nitrobenzene	7.451	77	85567	10.380	ng	98
25) Isophorone	7.692	82	146541	10.204	ng	98
26) 2-Nitrophenol	7.769	139	41083	10.067	ng	97
27) 2,4-Dimethylphenol	7.804	122	60896	9.956	ng	99
28) bis(2-Chloroethoxy)met...	7.898	93	91497	10.202	ng	95
29) 2,4-Dichlorophenol	8.016	162	75032	10.643	ng	97
30) 1,2,4-Trichlorobenzene	8.098	180	76623	10.585	ng	94
31) Naphthalene	8.181	128	220903	10.590	ng	99
32) Benzoic acid	7.910	122	58822	13.120	ng	92
33) 4-Chloroaniline	8.228	127	79703	10.476	ng	97
34) Hexachlorobutadiene	8.292	225	54849	10.054	ng	96
35) Caprolactam	8.575	113	20560	10.645	ng	99
36) 4-Chloro-3-methylphenol	8.704	107	67228	10.641	ng	96
37) 2-Methylnaphthalene	8.869	142	147427	10.571	ng	99
38) 1-Methylnaphthalene	8.969	142	149280	11.093	ng	97
40) 1,2,4,5-Tetrachloroben...	9.034	216	86408	10.098	ng	99
41) Hexachlorocyclopentadiene	9.022	237	15765	10.609	ng	93
43) 2,4,6-Trichlorophenol	9.145	196	58560	10.054	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110525\
 Data File : BF144171.D
 Acq On : 05 Nov 2025 13:50
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Quant Time: Nov 05 17:23:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 17:20:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	63779	10.160	ng	98
46) 1,1'-Biphenyl	9.334	154	188739	10.355	ng	99
47) 2-Chloronaphthalene	9.357	162	158745	10.211	ng	100
48) 2-Nitroaniline	9.457	65	45323	10.104	ng	97
49) Acenaphthylene	9.775	152	220118	10.390	ng	99
50) Dimethylphthalate	9.628	163	175284	10.133	ng	99
51) 2,6-Dinitrotoluene	9.692	165	35433	10.119	ng	98
52) Acenaphthene	9.945	154	151661	10.705	ng	99
53) 3-Nitroaniline	9.863	138	39270	10.259	ng	99
54) 2,4-Dinitrophenol	9.981	184	26319	12.447	ng	97
55) Dibenzofuran	10.122	168	208943	10.530	ng	99
56) 4-Nitrophenol	10.028	139	35627	12.867	ng	98
57) 2,4-Dinitrotoluene	10.098	165	49257	10.508	ng	# 98
58) Fluorene	10.463	166	159623	10.581	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.239	232	45333	10.207	ng	97
60) Diethylphthalate	10.328	149	166246	10.426	ng	98
61) 4-Chlorophenyl-phenyle...	10.457	204	91478	10.645	ng	99
62) 4-Nitroaniline	10.475	138	37821	10.375	ng	98
63) Azobenzene	10.616	77	157098	10.586	ng	96
65) 4,6-Dinitro-2-methylph...	10.510	198	34516	10.823	ng	99
66) n-Nitrosodiphenylamine	10.569	169	158661	10.509	ng	98
67) 4-Bromophenyl-phenylether	10.945	248	60929	10.169	ng	96
68) Hexachlorobenzene	11.016	284	71759	10.168	ng	94
69) Atrazine	11.098	200	49466	10.302	ng	97
70) Pentachlorophenol	11.210	266	31673	9.012	ng	97
71) Phenanthrene	11.428	178	244478	10.460	ng	99
72) Anthracene	11.480	178	248511	10.457	ng	99
73) Carbazole	11.639	167	215269	10.652	ng	98
74) Di-n-butylphthalate	11.963	149	258464	10.580	ng	99
75) Fluoranthene	12.622	202	250112	10.359	ng	99
77) Benzidine	12.745	184	104876	11.958	ng	99
78) Pyrene	12.851	202	258743	10.823	ng	99
80) Butylbenzylphthalate	13.469	149	87462	10.556	ng	96
81) Benzo(a)anthracene	14.045	228	198590	9.664	ng	100
82) 3,3'-Dichlorobenzidine	14.010	252	67166	9.535	ng	97
83) Chrysene	14.080	228	184604	9.731	ng	98
84) Bis(2-ethylhexyl)phtha...	14.027	149	109593	9.662	ng	98
85) Di-n-octyl phthalate	14.645	149	172428	8.811	ng	99
87) Indeno(1,2,3-cd)pyrene	17.039	276	219736	9.831	ng	99
88) Benzo(b)fluoranthene	15.104	252	176369	9.963	ng	99
89) Benzo(k)fluoranthene	15.133	252	170386	10.283	ng	98
90) Benzo(a)pyrene	15.474	252	163529	10.013	ng	97
91) Dibenzo(a,h)anthracene	17.051	278	179793	10.016	ng	98
92) Benzo(g,h,i)perylene	17.492	276	175090	9.877	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 05 17:23:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 05 17:20:41 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.875 min Scan# 79
 Delta R.T. -0.000 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:152 Resp: 112609
 Ion Ratio Lower Upper
 152 100
 150 159.0 127.4 191.0
 115 62.3 49.5 74.3



#2
 1,4-Dioxane
 Concen: 9.214 ng
 RT: 2.616 min Scan# 72
 Delta R.T. 0.012 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50

Tgt Ion: 88 Resp: 27411
 Ion Ratio Lower Upper
 88 100
 58 87.1 73.1 109.7
 43 38.9 31.0 46.6









#3
Pyridine
Concen: 9.536 ng
RT: 3.393 min Scan# 201
Delta R.T. 0.018 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 79 Resp: 79148
Ion Ratio Lower Upper
79 100
52 81.3 64.5 96.7
51 37.8 31.7 47.5



#4
n-Nitrosodimethylamine
Concen: 9.331 ng
RT: 3.322 min Scan# 192
Delta R.T. -0.023 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion: 42 Resp: 40456
Ion Ratio Lower Upper
42 100
74 105.1 88.2 132.4
44 6.8 5.7 8.5









#5
2-Fluorophenol
Concen: 20.391 ng
RT: 5.487 min Scan# 51
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 146731
Ion Ratio Lower Upper
112 100
64 58.0 47.2 70.8
63 29.3 24.4 36.6



#6
Aniline
Concen: 10.512 ng
RT: 6.534 min Scan# 738
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion: 93 Resp: 122118
Ion Ratio Lower Upper
93 100
66 42.0 34.3 51.5
65 20.1 16.9 25.3









#7
Phenol-d6
Concen: 20.963 ng
RT: 6.498 min Scan# 711
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 99 Resp: 170929
Ion Ratio Lower Upper
99 100
42 24.9 20.9 31.3
71 35.1 31.4 47.2



#8
2-Chlorophenol
Concen: 10.448 ng
RT: 6.657 min Scan# 759
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion: 128 Resp: 73729
Ion Ratio Lower Upper
128 100
130 31.3 10.5 50.5
64 53.3 33.2 73.2





#9

Benzaldehyde

Concen: 12.629 ng

RT: 6.422 min Scan# 719

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 77 Resp: 58146

Ion Ratio Lower Upper

77 100

105 96.7 77.2 117.2

106 95.1 72.2 112.2



#10

Phenol

Concen: 10.717 ng

RT: 6.510 min Scan# 734

Delta R.T. -0.018 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion: 94 Resp: 106765


Ion Ratio Lower Upper

94 100

65 28.3 10.8 50.8

66 41.8 24.3 64.3









#11
bis(2-Chloroethyl)ether
Concen: 10.342 ng
RT: 6.604 min Scan# 751
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 93 Resp: 75075
Ion Ratio Lower Upper
93 100
63 75.9 55.2 95.2
95 33.7 11.2 51.2



#12
1,3-Dichlorobenzene
Concen: 10.174 ng
RT: 6.816 min Scan# 786
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:146 Resp: 88594
Ion Ratio Lower Upper
146 100
148 66.3 51.8 77.8
75 30.8 25.0 37.4









#13
1,4-Dichlorobenzene
Concen: 10.085 ng
RT: 6.893 min Scan# 799
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:146 Resp: 87984
Ion Ratio Lower Upper
146 100
148 67.1 50.2 75.2
111 39.9 32.6 48.8



#14
1,2-Dichlorobenzene
Concen: 10.165 ng
RT: 7.045 min Scan# 825
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:146 Resp: 84997
Ion Ratio Lower Upper
146 100
148 64.6 50.7 76.1
111 40.3 34.2 51.4









#15
Benzyl Alcohol
Concen: 9.937 ng
RT: 7.010 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 64160
Ion Ratio Lower Upper
79 100
108 75.8 58.2 87.4
77 63.8 49.9 74.9



#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.403 ng
RT: 7.140 min Scan# 841
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion: 45 Resp: 139062
Ion Ratio Lower Upper
45 100
77 13.3 0.0 31.8
79 9.9 0.0 29.1









#17
2-Methylphenol
Concen: 10.575 ng
RT: 7.122 min Scan# 811
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:	107	Resp:	59368
Ion	Ratio	Lower	Upper
107	100		
108	115.6	91.0	136.4
77	51.0	42.7	64.1
79	50.0	41.4	62.2



#18
Hexachloroethane
Concen: 9.963 ng
RT: 7.387 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:	117	Resp:	28876
Ion	Ratio	Lower	Upper
117	100		
119	96.2	75.8	113.8
201	129.1	98.9	148.3









#19
n-Nitroso-di-n-propylamine
Concen: 10.735 ng
RT: 7.275 min Scan# 864
Delta R.T. -0.023 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:	70	Resp:	57621
Ion	Ratio	Lower	Upper
70	100		
42	75.0	58.3	87.5
101	9.2	7.0	10.6
130	16.9	14.4	21.6



#20
3+4-Methylphenols
Concen: 11.314 ng
RT: 7.275 min Scan# 864
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:	107	Resp:	74874
Ion	Ratio	Lower	Upper
107	100		
108	90.0	70.2	110.2
77	150.7	123.9	163.9
79	31.2	14.2	54.2









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:136 Resp: 438512
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 8.5 7.0 10.6
68 5.5 4.3 6.5



#22
Acetophenone
Concen: 10.819 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:105 Resp: 101777
Ion Ratio Lower Upper
105 100
71 7.4 7.2 10.8
51 39.9 34.0 51.0
120 21.7 17.8 26.6





#23

Nitrobenzene-d5

Concen: 20.347 ng

RT: 7.434 min Scan# 891

Delta R.T. -0.012 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 82 Resp: 154488

Ion Ratio Lower Upper

82 100

128 42.0 34.7 52.1

54 62.0 49.5 74.3



#24

Nitrobenzene

Concen: 10.380 ng

RT: 7.451 min Scan# 894

Delta R.T. -0.012 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion: 77 Resp: 85567


Ion Ratio Lower Upper

77 100

123 46.4 35.8 53.8

65 14.3 12.2 18.2









#25
Isophorone
Concen: 10.204 ng
RT: 7.692 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 82 Resp: 146541
Ion Ratio Lower Upper
82 100
95 7.5 5.5 8.3
138 17.9 13.7 20.5



#26
2-Nitrophenol
Concen: 10.067 ng
RT: 7.769 min Scan# 948
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:139 Resp: 41083
Ion Ratio Lower Upper
139 100
109 25.4 22.2 33.2
65 52.6 40.7 61.1









#27
2,4-Dimethylphenol
Concen: 9.956 ng
RT: 7.804 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:122 Resp: 60896
Ion Ratio Lower Upper
122 100
107 109.2 86.2 129.4
121 56.0 44.6 66.8



#28
bis(2-Chloroethoxy)methane
Concen: 10.202 ng
RT: 7.898 min Scan# 970
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion: 93 Resp: 91497
Ion Ratio Lower Upper
93 100
95 33.1 24.2 36.4
123 13.2 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 10.643 ng

RT: 8.016 min Scan# 990

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 75032

Ion Ratio Lower Upper

162 100

164 60.2 42.9 82.9

98 32.7 13.3 53.3



#30

1,2,4-Trichlorobenzene

Concen: 10.585 ng

RT: 8.098 min Scan# 1004

Delta R.T. 0.000 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:180 Resp: 76623


Ion Ratio Lower Upper

180 100

182 102.7 87.7 131.5

145 29.8 25.4 38.0









#31
Naphthalene
Concen: 10.590 ng
RT: 8.181 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:128 Resp: 220903
Ion Ratio Lower Upper
128 100
129 10.3 8.9 13.3
127 12.6 10.3 15.5



#32
Benzoic acid
Concen: 13.120 ng
RT: 7.910 min Scan# 972
Delta R.T. -0.041 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:122 Resp: 58822
Ion Ratio Lower Upper
122 100
105 112.1 99.9 139.9
77 88.0 77.5 117.5









#33
4-Chloroaniline
Concen: 10.476 ng
RT: 8.228 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:127	Resp:	79703
Ion	Ratio	Lower Upper
127	100	
129	30.6	25.3 37.9
65	31.7	27.6 41.4
92	22.2	18.4 27.6



#34
Hexachlorobutadiene
Concen: 10.054 ng
RT: 8.292 min Scan# 1037
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:225	Resp:	54849
Ion	Ratio	Lower Upper
225	100	
223	67.2	50.2 75.4
227	61.5	50.1 75.1









#35
 Caprolactam
 Concen: 10.645 ng
 RT: 8.575 min Scan# 1107
 Delta R.T. -0.047 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:113 Resp: 20560
 Ion Ratio Lower Upper
 113 100
 55 184.5 166.8 206.8
 56 146.2 125.9 165.9



#36
 4-Chloro-3-methylphenol
 Concen: 10.641 ng
 RT: 8.704 min Scan# 1107
 Delta R.T. -0.012 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50

Tgt Ion:107 Resp: 67228
 Ion Ratio Lower Upper
 107 100
 144 26.7 23.3 34.9
 142 76.8 63.8 95.6









#37
2-Methylnaphthalene
Concen: 10.571 ng
RT: 8.869 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:142 Resp: 147427
Ion Ratio Lower Upper
142 100
141 88.7 69.9 104.9



#38
1-Methylnaphthalene
Concen: 11.093 ng
RT: 8.969 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:142 Resp: 149280
Ion Ratio Lower Upper
142 100
141 87.6 72.1 108.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:164 Resp: 261107
Ion Ratio Lower Upper
164 100
162 97.4 81.1 121.7
160 43.9 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.098 ng
RT: 9.034 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:216 Resp: 86408
Ion Ratio Lower Upper
216 100
214 79.8 63.4 95.2
179 19.0 15.0 22.4
108 12.7 10.5 15.7









#41
Hexachlorocyclopentadiene
Concen: 10.609 ng
RT: 9.022 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:237 Resp: 15765
Ion Ratio Lower Upper
237 100
235 59.3 45.5 85.5
272 13.9 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 20.442 ng
RT: 10.704 min Scan# 1447
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:330 Resp: 66981
Ion Ratio Lower Upper
330 100
332 93.9 76.7 115.1
141 22.6 17.8 26.8





#43

2,4,6-Trichlorophenol

Concen: 10.054 ng

RT: 9.145 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 58560

Ion Ratio Lower Upper

196 100

198 94.0 75.3 112.9

200 29.8 23.4 35.2



#44

2,4,5-Trichlorophenol

Concen: 10.160 ng

RT: 9.192 min Scan# 1190

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Tgt Ion:196 Resp: 63779




Ion Ratio Lower Upper


196 100

198 93.9 76.1 114.1

97 38.4 31.9 47.9

132 23.4 19.4 29.2









#45
2-Fluorobiphenyl
Concen: 21.090 ng
RT: 9.228 min Scan# 1196
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:172 Resp: 372845
Ion Ratio Lower Upper
172 100
171 35.0 28.1 42.1
170 23.5 18.6 28.0



#46
1,1'-Biphenyl
Concen: 10.355 ng
RT: 9.334 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:154 Resp: 188739
Ion Ratio Lower Upper
154 100
153 38.4 19.2 59.2
76 11.3 0.0 31.8





#47

2-Chloronaphthalene

Concen: 10.211 ng

RT: 9.357 min Scan# 1218

Delta R.T. -0.012 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 158745

Ion Ratio Lower Upper

162 100

127 36.4 28.8 43.2

164 32.5 25.8 38.8



#48

2-Nitroaniline

Concen: 10.104 ng

RT: 9.457 min Scan# 1235

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion: 65 Resp: 45323


Ion Ratio Lower Upper

65 100

92 65.0 53.0 79.6

138 105.1 81.6 122.4





#49

Acenaphthylene

Concen: 10.390 ng

RT: 9.775 min Scan# 1289

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:152 Resp: 220118

Ion Ratio Lower Upper

152 100

151 19.4 15.9 23.9

153 12.6 10.5 15.7



#50

Dimethylphthalate

Concen: 10.133 ng

RT: 9.628 min Scan# 1264

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:163 Resp: 175284


Ion Ratio Lower Upper

163 100

194 4.4 3.3 4.9

164 10.3 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 10.119 ng
RT: 9.692 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 35433
Ion Ratio Lower Upper
165 100
63 55.4 43.8 65.8
89 49.8 37.5 56.3



#52
Acenaphthene
Concen: 10.705 ng
RT: 9.945 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:154 Resp: 151661
Ion Ratio Lower Upper
154 100
153 109.5 89.0 133.4
152 52.7 42.2 63.4





#53

3-Nitroaniline

Concen: 10.259 ng

RT: 9.863 min Scan# 11

Delta R.T. -0.012 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:138 Resp: 39270

Ion Ratio Lower Upper

138 100

108 10.2 7.5 11.3

92 116.1 91.8 137.6



#54

2,4-Dinitrophenol

Concen: 12.447 ng

RT: 9.981 min Scan# 1324

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:184 Resp: 26319


Ion Ratio Lower Upper

184 100

63 58.3 44.0 66.0

154 51.9 42.4 63.6









#55
Dibenzenofuran
Concen: 10.530 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:168 Resp: 208943
Ion Ratio Lower Upper
168 100
139 37.2 30.2 45.2
169 15.2 12.2 18.4



#56
4-Nitrophenol
Concen: 12.867 ng
RT: 10.028 min Scan# 1332
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:139 Resp: 35627
Ion Ratio Lower Upper
139 100
109 57.6 39.9 79.9
65 98.0 79.5 119.5









#57
2,4-Dinitrotoluene
Concen: 10.508 ng
RT: 10.098 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 49257
Ion Ratio Lower Upper
165 100
63 42.0 35.5 53.3
89 60.1 47.4 71.2
182 3.0 1.9 2.9#



#58
Fluorene
Concen: 10.581 ng
RT: 10.463 min Scan# 1406
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:166 Resp: 159623
Ion Ratio Lower Upper
166 100
165 99.9 78.6 118.0
167 13.6 10.4 15.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 10.207 ng
RT: 10.239 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 45333




Ion	Ratio	Lower	Upper
232	100		
131	39.1	30.3	45.5
130	1.6	1.2	1.8
166	25.0	21.5	32.3




#60
Diethylphthalate
Concen: 10.426 ng
RT: 10.328 min Scan# 1383
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:149 Resp: 166246

Ion	Ratio	Lower	Upper
149	100		
177	23.5	18.3	27.5
150	13.0	9.8	14.6









#61
4-Chlorophenyl-phenylether
Concen: 10.645 ng
RT: 10.457 min Scan# 1405
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 91478
Ion Ratio Lower Upper
204 100
206 34.2 27.4 41.2
141 48.5 40.0 60.0



#62
4-Nitroaniline
Concen: 10.375 ng
RT: 10.475 min Scan# 1408
Delta R.T. -0.023 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:138 Resp: 37821
Ion Ratio Lower Upper
138 100
92 49.8 29.4 69.4
108 63.7 46.4 86.4









#63
Azobenzene
Concen: 10.586 ng
RT: 10.616 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 77	Resp: 157098
Ion Ratio	Lower Upper
77 100	
182 30.8	8.1 48.1
105 22.3	2.0 42.0
51 31.6	13.9 53.9



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1566
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:188	Resp: 469428
Ion Ratio	Lower Upper
188 100	
94 7.3	6.2 9.2
80 8.1	6.9 10.3









#65
4,6-Dinitro-2-methylphenol
Concen: 10.823 ng
RT: 10.510 min Scan# 1415
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:198 Resp: 34516
Ion Ratio Lower Upper
198 100
51 46.1 25.6 65.6
105 33.8 14.6 54.6



#66
n-Nitrosodiphenylamine
Concen: 10.509 ng
RT: 10.569 min Scan# 1424
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:169 Resp: 158661
Ion Ratio Lower Upper
169 100
168 55.2 46.2 69.2
167 32.0 25.7 38.5









#67
4-Bromophenyl-phenylether
Concen: 10.169 ng
RT: 10.945 min Scan# 1488
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:248 Resp: 60929
Ion Ratio Lower Upper
248 100
250 93.2 78.4 117.6
141 58.6 45.1 67.7



#68
Hexachlorobenzene
Concen: 10.168 ng
RT: 11.016 min Scan# 1500
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:284 Resp: 71759
Ion Ratio Lower Upper
284 100
142 21.2 19.2 28.8
249 26.3 23.8 35.8





#69

Atrazine

Concen: 10.302 ng

RT: 11.098 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:200 Resp: 49466

Ion Ratio Lower Upper

200 100

173 28.1 5.0 45.0

215 49.6 28.7 68.7



#70

Pentachlorophenol

Concen: 9.012 ng

RT: 11.210 min Scan# 1533

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:266 Resp: 31673


Ion Ratio Lower Upper

266 100

268 59.6 49.5 74.3

264 64.0 49.4 74.2





#71

Phenanthrene

Concen: 10.460 ng

RT: 11.428 min Scan# 1570

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:178 Resp: 244478

Ion Ratio Lower Upper

178 100

176 19.3 15.5 23.3

179 14.9 12.2 18.4



#72

Anthracene

Concen: 10.457 ng

RT: 11.480 min Scan# 1579

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:178 Resp: 248511


Ion Ratio Lower Upper

178 100

176 18.8 15.3 22.9

179 15.4 12.0 18.0





#73

Carbazole

Concen: 10.652 ng

RT: 11.639 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:167 Resp: 215269

Ion Ratio Lower Upper

167 100

166 20.3 17.2 25.8

139 11.7 9.9 14.9



#74

Di-n-butylphthalate

Concen: 10.580 ng

RT: 11.963 min Scan# 1661

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:149 Resp: 258464


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.0


104 3.9 3.4 5.2








#75
 Fluoranthene
 Concen: 10.359 ng
 RT: 12.622 min Scan# 1774
 Delta R.T. -0.006 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010







Tgt Ion:202 Resp: 250112
 Ion Ratio Lower Upper
 202 100
 101 8.5 0.0 27.9
 203 17.4 0.0 37.5



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.057 min Scan# 2017
 Delta R.T. -0.006 min
 Lab File: BF144171.D
 Acq: 05 Nov 2025 13:50

Tgt Ion:240 Resp: 296506
 Ion Ratio Lower Upper
 240 100
 120 8.1 6.6 10.0
 236 25.5 21.4 32.2





#77

Benzidine

Concen: 11.958 ng

RT: 12.745 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:184 Resp: 104876

Ion Ratio Lower Upper

184 100

185 14.1 11.7 17.5

183 10.7 8.3 12.5



#78

Pyrene

Concen: 10.823 ng

RT: 12.851 min Scan# 1812

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:202 Resp: 258743


Ion Ratio Lower Upper

202 100

200 21.3 17.2 25.8

203 17.9 13.7 20.5





#79

Terphenyl-d14

Concen: 22.476 ng

RT: 12.992 min Scan# 1836

Delta R.T. -0.006 min

Lab File: BF144171.D


Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010






Tgt Ion:244 Resp: 419548

Ion Ratio Lower Upper

244 100

212 5.9 5.3 7.9

122 6.8 5.6 8.4



#80

Butylbenzylphthalate

Concen: 10.556 ng

RT: 13.469 min Scan# 1917

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:149 Resp: 87462


Ion Ratio Lower Upper

149 100

91 66.1 56.1 84.1

206 23.6 18.8 28.2









#81
Benzo(a)anthracene
Concen: 9.664 ng
RT: 14.045 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 198590
Ion Ratio Lower Upper
228 100
226 27.3 21.8 32.8
229 19.4 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 9.535 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:252 Resp: 67166
Ion Ratio Lower Upper
252 100
254 64.2 53.5 80.3
126 6.7 6.4 9.6





#83

Chrysene

Concen: 9.731 ng

RT: 14.080 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 184604

Ion Ratio Lower Upper

228 100

226 30.0 22.9 34.3

229 19.7 15.0 22.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 9.662 ng

RT: 14.027 min Scan# 2012

Delta R.T. -0.006 min

Lab File: BF144171.D

Acq: 05 Nov 2025 13:50




Tgt Ion:149 Resp: 109593


Ion Ratio Lower Upper

149 100

167 28.2 21.8 32.8

279 4.3 3.7 5.5









#85
Di-n-octyl phthalate
Concen: 8.811 ng
RT: 14.645 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:149 Resp: 172428
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 11.7 9.0 13.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:264 Resp: 311371
Ion Ratio Lower Upper
264 100
260 21.9 17.1 25.7
265 22.2 17.4 26.0









#87
Indeno(1,2,3-cd)pyrene
Concen: 9.831 ng
RT: 17.039 min Scan# 219736
Delta R.T. -0.023 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:276 Resp: 219736
Ion Ratio Lower Upper
276 100
138 17.7 13.9 20.9
277 23.8 19.8 29.6



#88
Benzo(b)fluoranthene
Concen: 9.963 ng
RT: 15.104 min Scan# 2195
Delta R.T. -0.006 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:252 Resp: 176369
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.2
125 6.9 5.3 7.9









#89
Benzo(k)fluoranthene
Concen: 10.283 ng
RT: 15.133 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:252 Resp: 170386
Ion Ratio Lower Upper
252 100
253 21.4 18.0 27.0
125 6.7 5.4 8.2



#90
Benzo(a)pyrene
Concen: 10.013 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.012 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:252 Resp: 163529
Ion Ratio Lower Upper
252 100
253 22.3 16.7 25.1
125 7.8 5.7 8.5









#91
Dibenzo(a,h)anthracene
Concen: 10.016 ng
RT: 17.051 min Scan# 21
Delta R.T. -0.029 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:278 Resp: 179793
Ion Ratio Lower Upper
278 100
139 12.5 9.1 13.7
279 23.7 18.2 27.4



#92
Benzo(g,h,i)perylene
Concen: 9.877 ng
RT: 17.492 min Scan# 2601
Delta R.T. -0.029 min
Lab File: BF144171.D
Acq: 05 Nov 2025 13:50

Tgt Ion:276 Resp: 175090
Ion Ratio Lower Upper
276 100
277 23.6 19.0 28.4
138 15.9 12.9 19.3

