

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
 Data File : BF110523.D
 Acq On : 6 Nov 2018 18:58
 Operator : JU/SJ
 Sample : J5826-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-1

Manual Integrations
 APPROVED

Sohil
 11/7/2018 2:54:30 PM

Quant Time: Nov 07 09:57:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	83153	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	351032	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	177632	20.00	ng	0.00
64) Phenanthrene-d10	11.49	188	340924	20.00	ng	0.00
76) Chrysene-d12	14.13	240	174785	20.00	ng	0.00
87) Perylene-d12	15.66	264	159143	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	465291	91.12	ng	0.00
7) Phenol-d6	6.57	99	608165	93.12	ng	0.00
23) Nitrobenzene-d5	7.52	82	386466	65.30	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	159966	90.91	ng	0.00
45) 2-Fluorobiphenyl	9.31	172	746712	66.01	ng	0.00
79) Terphenyl-d14	13.07	244	663688	78.97	ng	0.00
Target Compounds						
10) Phenol	6.59	94	38818	5.560	ng	Qvalue # 73
50) Dimethylphthalate	9.70	163	102078	7.785	ng	99
71) Phenanthrene	11.51	178	46394	2.601	ng	98
75) Fluoranthene	12.70	202	60047	3.317	ng	99
78) Pyrene	12.94	202	51176	4.084	ng	99
84) Bis(2-ethylhexyl)phthalate	14.08	149	44192	6.011	ng	92
85) Di-n-octyl phthalate	14.73	149	29432	2.575	ng	# 44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
Data File : BF110523.D
Acq On : 6 Nov 2018 18:58
Operator : JU/SJ
Sample : J5826-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-1

Manual Integrations
APPROVED

Sohil
11/7/2018 2:54:30 PM

Quant Time: Nov 07 09:57:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

