

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
 Data File : BF110528.D
 Acq On : 6 Nov 2018 21:13
 Operator : JU/SJ
 Sample : J5764-01 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-110218-A

Manual Integrations
 APPROVED

Sohil
 11/7/2018 2:54:37 PM

Quant Time: Nov 07 12:33:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	93849	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	369479	20.00	ng	0.00
39) Acenaphthene-d10	10.00	164	141213	20.00	ng	0.00
64) Phenanthrene-d10	11.49	188	208598	20.00	ng	0.00
76) Chrysene-d12	14.14	240	172708	20.00	ng	0.00
87) Perylene-d12	15.66	264	118038	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	279711	48.53	ng	0.00
7) Phenol-d6	6.57	99	358888	48.69	ng	0.00
23) Nitrobenzene-d5	7.52	82	217128	34.86	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	55120	39.40	ng	0.00
45) 2-Fluorobiphenyl	9.31	172	390908	43.47	ng	0.00
79) Terphenyl-d14	13.07	244	263910	31.78	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.70	163	54100	5.190	ng	Qvalue 100
71) Phenanthrene	11.51	178	63809	5.846	ng	99
75) Fluoranthene	12.71	202	101683	9.179	ng	99
78) Pyrene	12.94	202	95039	7.676	ng	99
81) Benzo(a)anthracene	14.13	228	46953	4.584	ng	99
83) Chrysene	14.16	228	45178m	4.644	ng	
84) Bis(2-ethylhexyl)phthalate	14.09	149	40206	5.534	ng	# 96
88) Benzo(b)fluoranthene	15.21	252	41924m	6.151	ng	
89) Benzo(k)fluoranthene	15.24	252	16461m	2.456	ng	
90) Benzo(a)pyrene	15.60	252	27737	4.422	ng	# 86
92) Benzo(g,h,i)perylene	17.72	276	17608	3.302	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
Data File : BF110528.D
Acq On : 6 Nov 2018 21:13
Operator : JU/SJ
Sample : J5764-01 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-110218-A

Manual Integrations
APPROVED

Sohil
11/7/2018 2:54:37 PM

Quant Time: Nov 07 12:33:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

